


2,4,6-Tribromophenol

Response Ratio



$$\text{Response} = 1.903\text{e-}001 * \text{Amt} - 7.513\text{e-}002$$


Coef of Det (r^2) = 0.999381 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN111121.M

Calibration Table Last Updated: Thu Nov 11 08:04:11 2021

4,6-Dinitro-2-methylphenol

Response Ratio



$$\text{Response} = 7.975\text{e-}002 * \text{Amt} - 4.085\text{e-}002$$


Coef of Det (r^2) = 0.990412 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN111121.M

Calibration Table Last Updated: Thu Nov 11 08:04:11 2021

Atrazine

Response Ratio



$$\text{Response} = 2.121\text{e-}001 * \text{Amt} - 7.916\text{e-}002$$


Coef of Det (r^2) = 0.997629 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN111121.M

Calibration Table Last Updated: Thu Nov 11 08:04:11 2021

Pentachlorophenol

Response Ratio



$$\text{Response} = 1.335\text{e-}001 * \text{Amt} - 9.391\text{e-}002$$


Coef of Det (r^2) = 0.996098 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN111121.M

Calibration Table Last Updated: Thu Nov 11 08:04:11 2021

Bis(2-ethylhexyl)phthalate

Response Ratio



$$\text{Response} = 8.182\text{e-}001 * \text{Amt} - 3.830\text{e-}001$$

Coef of Det (r^2) = 0.996373 Curve Fit: Linear

Method Name: Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN111121.M

Calibration Table Last Updated: Thu Nov 11 08:04:11 2021