

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111121\
 Data File : BN017392.D
 Acq On : 10 Nov 2021 22:04
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN111121

Quant Time: Nov 11 08:07:08 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN111121.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 11 08:04:11 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	110	0.00
2	1,4-Dioxane	0.178	0.178	0.0	107	0.00
3	n-Nitrosodimethylamine	0.396	0.387	2.3	108	0.00
4 S	2-Fluorophenol	0.920	0.968	-5.2	114	0.00
5 S	Phenol-d6	1.018	1.076	-5.7	117	0.00
6	bis(2-Chloroethyl)ether	1.076	1.148	-6.7	111	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	110	0.00
8 S	Nitrobenzene-d5	0.265	0.270	-1.9	113	0.00
9	Naphthalene	1.076	1.108	-3.0	110	0.00
10	Hexachlorobutadiene	0.204	0.215	-5.4	113	0.00
11 SURR	2-Methylnaphthalene-d10	0.565	0.591	-4.6	112	0.00
12	2-Methylnaphthalene	0.672	0.712	-6.0	112	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	110	0.00
14 S	2,4,6-Tribromophenol	0.149	0.127	14.8	118	0.00
15 S	2-Fluorobiphenyl	1.535	1.675	-9.1	112	0.00
16	Acenaphthylene	1.680	1.770	-5.4	116	-0.01
17	Acenaphthene	1.122	1.170	-4.3	111	0.00
18	Fluorene	1.353	1.401	-3.5	113	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	110	0.00
20	4,6-Dinitro-2-methylphenol	0.045	0.027	40.0#	84	0.00
21	4-Bromophenyl-phenylether	0.234	0.239	-2.1	112	0.00
22	Hexachlorobenzene	0.262	0.279	-6.5	114	0.00
23	Atrazine	0.157	0.142	9.6	116	0.00
24	Pentachlorophenol	0.087	0.062	28.7#	113	0.00
25	Phenanthrene	1.151	1.232	-7.0	112	0.00
26	Anthracene	0.963	0.996	-3.4	116	0.00
27 SURR	Fluoranthene-d10	1.032	1.053	-2.0	108	0.00
28	Fluoranthene	1.273	1.350	-6.0	107	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	110	0.00
30	Pyrene	1.249	1.302	-4.2	109	0.00
31 S	Terphenyl-d14	0.849	0.887	-4.5	109	0.00
32	Benzo(a)anthracene	1.201	1.239	-3.2	112	-0.01
33	Chrysene	1.321	1.348	-2.0	106	0.00
34	Bis(2-ethylhexyl)phthalate	0.571	0.498	12.8	132	0.00
35 I	Perylene-d12	1.000	1.000	0.0	109	0.00
36	Indeno(1,2,3-cd)pyrene	1.522	1.682	-10.5	113	0.00
37	Benzo(b)fluoranthene	1.226	1.297	-5.8	109	0.00
38	Benzo(k)fluoranthene	1.308	1.358	-3.8	110	0.00
39 C	Benzo(a)pyrene	1.128	1.187	-5.2	113	0.00
40	Dibenzo(a,h)anthracene	1.231	1.352	-9.8	113	0.00
41	Benzo(g,h,i)perylene	1.336	1.383	-3.5	107	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN111121