

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN111318\
 Data File : BN003515.D
 Acq On : 13 Nov 2018 17:11
 Operator : MJ/SJ
 Sample : J5925-07
 Misc :
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 TP-1D

Quant Time: Nov 14 01:35:02 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-BN111218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 12 15:59:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.85	152	95462	20.00	ng	0.00
21) Naphthalene-d8	10.65	136	372528	20.00	ng	0.00
39) Acenaphthene-d10	14.49	164	212058	20.00	ng	0.00
64) Phenanthrene-d10	17.23	188	438349	20.00	ng	0.00
76) Chrysene-d12	21.40	240	405706	20.00	ng	0.00
87) Perylene-d12	23.72	264	399214	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.42	112	818408	154.62	ng	0.00
7) Phenol-d6	7.01	99	1086795	148.05	ng	0.00
23) Nitrobenzene-d5	9.02	82	632385	106.86	ng	0.00
42) 2,4,6-Tribromophenol	15.99	330	518680	164.07	ng	0.01
45) 2-Fluorobiphenyl	13.11	172	1517667	95.53	ng	0.00
79) Terphenyl-d14	19.84	244	2143887	104.03	ng	0.00
Target Compounds						
10) Phenol	7.03	94	44408	5.607	ng	97
37) 2-Methylnaphthalene	12.32	142	1506381	116.315	ng	99
38) 1-Methylnaphthalene	12.53	142	1213634	96.749	ng	100
50) Dimethylphthalate	13.94	163	239394	14.410	ng	# 96
52) Acenaphthene	14.56	154	107942	8.538	ng	# 83
58) Fluorene	15.54	166	297196	19.311	ng	# 93
71) Phenanthrene	17.28	178	747035	32.622	ng	96
78) Pyrene	19.65	202	97527	3.561	ng	# 77

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN111318\
Data File : BN003515.D
Acq On : 13 Nov 2018 17:11
Operator : MJ/SJ
Sample : J5925-07
Misc :
ALS Vial : 3 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
TP-1D

Quant Time: Nov 14 01:35:02 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-BN111218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 12 15:59:25 2018
Response via : Initial Calibration

