

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN111318\
 Data File : BN003523.D
 Acq On : 13 Nov 2018 21:52
 Operator : MJ/SJ
 Sample : J5908-08
 Misc :
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SR-3-S

Quant Time: Nov 14 02:46:50 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-BN111218.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 12 15:59:25 2018
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.85	152	101459	20.00	ng	0.00
21) Naphthalene-d8	10.65	136	424279	20.00	ng	0.00
39) Acenaphthene-d10	14.48	164	250288	20.00	ng	0.00
64) Phenanthrene-d10	17.22	188	558032	20.00	ng	0.00
76) Chrysene-d12	21.40	240	455598	20.00	ng	0.00
87) Perylene-d12	23.72	264	453119	20.00	ng	0.00
System Monitoring Compounds						
5) 2-Fluorophenol	5.42	112	616144	109.53	ng	0.00
7) Phenol-d6	7.00	99	728097	93.32	ng	0.00
23) Nitrobenzene-d5	9.02	82	546569	81.09	ng	0.00
42) 2,4,6-Tribromophenol	15.97	330	472196	126.55	ng	0.00
45) 2-Fluorobiphenyl	13.10	172	1515005	80.79	ng	0.00
79) Terphenyl-d14	19.85	244	2079336	89.85	ng	0.00

Target Compounds Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN111318\
Data File : BN003523.D
Acq On : 13 Nov 2018 21:52
Operator : MJ/SJ
Sample : J5908-08
Misc :
ALS Vial : 11 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SR-3-S

Quant Time: Nov 14 02:46:50 2018
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-BN111218.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 12 15:59:25 2018
Response via : Initial Calibration

