


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111320\
 Data File : BN012570.D
 Acq On : 13 Nov 2020 08:33
 Operator : CG/JU
 Sample : DFTPP15
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP151

Integration File: rteint.p

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN111020.M
 Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 Last Update : Fri Nov 13 11:54:51 2020



AutoFind: Scans 2436, 2437, 2438; Background Corrected with Scan 2429

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	51.0	115074	PASS
68	69	0.00	2	2.0	2164	PASS
69	198	0.00	100	47.9	108181	PASS
70	69	0.00	2	0.3	375	PASS
127	198	10	80	50.2	113336	PASS
197	198	0.00	2	0.4	885	PASS
198	198	100	100	100.0	225749	PASS
199	198	5	9	6.8	15360	PASS
275	198	10	60	29.9	67429	PASS
365	198	1	100	4.8	10742	PASS
441	198	0.01	100	15.5	34898	PASS
442	198	50	100	98.8	222976	PASS
443	442	15	24	20.4	45424	PASS

Scan 2429 (17.275 min): BN012570.D
DFTPP15

m/z	Abundance
39.90	503.0
44.10	1023.0
49.00	567.0
73.00	881.0
191.00	356.0
207.00	1260.0
208.00	353.0

Instrument :
BNA_N
ClientSampleId :
DFTPP151

m/z	Abundance
37.00	1088.0
38.10	2257.0
39.10	12817.0
40.00	830.0
40.90	356.0
44.00	1422.0
44.90	591.0
49.20	1905.0
50.00	44728.0
51.00	150464.0
52.00	8537.0
55.10	776.0
56.00	4394.0
57.00	10234.0
58.00	655.0
61.00	2268.0
62.00	2505.0
63.00	5011.0
64.10	1114.0
65.10	3005.0
68.10	2875.0
69.00	140928.0
69.90	1127.0
73.00	2398.0
74.10	18008.0
75.00	23304.0
76.00	8437.0
77.00	150720.0
78.00	11157.0
78.90	12641.0
80.00	9293.0
81.00	13458.0
82.00	2564.0
83.00	2277.0
83.90	360.0
84.90	1816.0
86.00	5194.0
86.90	1946.0
88.00	445.0
89.00	548.0
91.00	2624.0
92.00	3505.0
93.00	24704.0
93.90	1474.0
95.00	362.0
96.10	789.0
97.10	549.0
98.00	18136.0
99.00	12923.0
100.00	1053.0
101.00	5957.0
101.90	388.0
103.00	2192.0
104.00	3901.0
105.00	4492.0
106.00	1973.0
107.00	50248.0
108.00	7031.0
109.00	1945.0
110.00	75648.0
111.00	12786.0
112.00	1802.0
113.00	819.0
115.00	336.0
116.00	2842.0
117.00	59840.0
118.00	4523.0
119.00	566.0
120.00	690.0
122.00	4496.0
123.00	5292.0
124.00	2389.0
124.90	1922.0
127.00	137280.0
128.00	11708.0
129.00	60656.0
130.00	4955.0
131.10	1279.0
132.10	652.0
133.10	457.0
134.00	1922.0
135.00	5612.0
135.90	1973.0
136.90	2738.0
138.00	615.0
139.90	845.0
141.00	9308.0
142.00	2593.0
143.00	1520.0
144.00	353.0
145.00	373.0
146.00	1479.0

Instrument :
BNA_N
ClientSampleId :
DFTPP151

147.00	4743.0
148.00	13215.0
149.00	2636.0
150.00	746.0
151.00	1177.0
151.90	747.0
153.00	2975.0
154.00	2103.0
155.00	5335.0
156.00	6329.0
157.00	1015.0
158.00	1381.0
158.90	1462.0
160.00	3209.0
161.00	3937.0
161.90	1155.0
163.10	426.0
164.00	518.0
164.90	4131.0
166.10	2136.0
167.00	18584.0
168.00	13013.0
169.00	1921.0
170.00	375.0
171.00	1014.0
171.90	1744.0
173.00	1491.0
174.00	3680.0
175.00	6225.0
176.00	1389.0
176.90	3368.0
178.00	1396.0
179.00	12992.0
180.00	8930.0
180.90	3879.0
182.00	486.0
183.90	1278.0
185.00	5221.0
186.00	44584.0
187.00	11936.0
188.00	1424.0
189.10	2987.0
190.00	633.0
191.00	1469.0
191.90	3937.0
193.00	4995.0
193.90	1079.0
194.80	326.0
196.00	6619.0
197.20	2656.0
197.90	262656.0
199.00	17688.0
200.00	1671.0
201.40	866.0
201.80	790.0
203.00	2541.0
204.00	11411.0
205.00	17648.0
206.00	77464.0
207.00	10668.0
208.00	2988.0
208.90	1713.0
210.20	1584.0
211.00	3174.0
211.90	510.0
215.00	945.0
216.00	1898.0
217.00	21344.0
218.00	3222.0
221.10	8824.0
222.00	2799.0
223.00	5216.0
224.00	38096.0
225.00	9747.0
226.00	1532.0
227.00	20848.0
228.00	2774.0
229.00	3852.0
229.90	454.0
231.10	1904.0
232.00	401.0
233.10	445.0
234.10	1141.0
234.90	1482.0
235.90	1413.0
236.90	2256.0
239.00	772.0
240.00	677.0
240.90	1333.0
242.00	2173.0
243.00	1979.0
244.00	27800.0
245.00	3871.0
246.00	7368.0
247.00	1324.0
249.10	845.0

Instrument :
BNA_N
ClientSampleId :
DFTPP151

251.00	795.0
252.90	858.0
255.00	156352.0
256.00	26104.0
256.90	2102.0
257.90	11562.0
258.90	1987.0
259.90	849.0
265.00	4575.0
265.80	551.0
271.00	363.0
272.00	712.0
272.90	4816.0
274.00	11514.0
275.00	69344.0
276.00	10364.0
277.00	7348.0
278.00	1715.0
278.80	444.0
280.90	526.0
282.90	643.0
284.00	421.0
285.00	1287.0
289.00	411.0
289.90	421.0
292.90	1825.0
294.90	593.0
296.00	28456.0
297.00	3211.0
297.90	318.0
301.10	534.0
301.90	460.0
303.00	2382.0
304.00	676.0
308.10	373.0
313.90	1466.0
314.90	3419.0
316.00	1797.0
317.20	307.0
321.10	624.0
322.00	461.0
322.90	5780.0
323.90	1009.0
327.00	1411.0
328.00	630.0
332.10	617.0
333.00	935.0
334.00	4484.0
335.00	984.0
340.90	572.0
346.00	1737.0
347.00	434.0
351.90	1682.0
352.90	883.0
354.10	2123.0
355.00	490.0
364.90	10092.0
365.90	1127.0
370.10	312.0
371.10	490.0
372.00	3929.0
373.00	748.0
383.00	589.0
384.00	402.0
390.00	376.0
391.00	453.0
392.00	383.0
402.00	1287.0
402.90	1310.0
404.00	713.0
414.80	456.0
420.90	1551.0
422.00	1820.0
423.00	9880.0
424.00	3777.0
425.10	430.0
441.10	28872.0
442.00	184512.0
443.10	35568.0
444.00	3913.0

Instrument :
BNA_N
ClientSampleId :
DFTPP151

m/z	Abundance
37.00	987.0
38.00	2103.0
39.10	11061.0
40.00	941.0
43.00	309.0
44.00	1734.0
45.00	576.0
49.00	1398.0
50.00	39464.0
51.00	136192.0
52.00	5889.0
52.90	324.0
55.10	615.0
56.00	3316.0
57.00	8385.0
57.90	526.0
61.00	1487.0
61.90	2328.0
63.00	5371.0
64.00	691.0
65.00	2273.0
68.00	2525.0
69.00	129400.0
70.10	692.0
73.10	1943.0
74.00	16278.0
75.00	22736.0
76.10	8140.0
77.10	137792.0
78.00	8661.0
79.00	11863.0
80.00	8016.0
81.00	10692.0
82.00	2250.0
83.00	2261.0
84.10	433.0
85.00	2230.0
86.00	3540.0
86.90	1903.0
88.00	623.0
88.90	328.0
91.10	2587.0
92.00	3397.0
93.00	20488.0
94.00	1493.0
95.20	321.0
96.00	691.0
98.00	17456.0
99.00	11750.0
100.10	873.0
101.00	5936.0
102.00	362.0
102.90	2449.0
104.00	4657.0
105.00	4015.0
107.00	47648.0
108.10	5748.0
109.10	1478.0
110.00	73672.0
111.00	11627.0
112.00	1950.0
113.00	475.0
115.90	3220.0
117.00	55568.0
118.00	3909.0
119.00	500.0
122.00	2202.0
123.00	5210.0
124.00	2384.0
125.00	1822.0
125.80	570.0
127.00	136640.0
128.00	11150.0
129.00	61760.0
130.00	5744.0
131.00	816.0
132.10	330.0
134.00	2323.0
135.00	5608.0
136.00	2072.0
137.10	2676.0
137.90	369.0
138.80	400.0
140.10	753.0
141.00	8964.0
142.00	3016.0
142.90	2015.0
143.90	535.0
144.90	481.0
146.00	1707.0
147.00	5003.0
148.00	12951.0

Instrument :
BNA_N
ClientSampleId :
DFTPP151

149.00	2602.0
150.00	659.0
151.10	843.0
152.20	600.0
153.00	3356.0
154.00	2274.0
155.00	4926.0
156.00	6868.0
157.10	1221.0
158.00	1847.0
159.00	1204.0
160.00	3156.0
161.00	3499.0
161.90	1064.0
163.10	498.0
164.10	311.0
164.90	3657.0
166.10	2517.0
167.00	19840.0
168.00	9843.0
168.90	2168.0
170.00	539.0
170.90	906.0
171.90	1522.0
173.00	2591.0
174.00	3698.0
175.00	7270.0
175.90	1748.0
177.10	3295.0
177.90	1202.0
179.00	12943.0
180.00	9308.0
181.00	3548.0
182.00	507.0
184.00	1205.0
185.00	6597.0
186.00	43896.0
187.00	13218.0
188.00	2057.0
189.00	3229.0
190.10	573.0
191.00	2024.0
192.00	3767.0
193.00	4468.0
193.90	972.0
195.20	490.0
196.00	7576.0
198.00	268480.0
199.00	18840.0
199.90	1877.0
201.30	1227.0
202.10	411.0
203.00	2798.0
204.00	11577.0
205.00	19688.0
206.00	81552.0
207.00	11220.0
207.90	3611.0
209.00	1113.0
210.10	1244.0
211.10	3153.0
211.90	920.0
214.90	1464.0
216.00	2012.0
217.00	23440.0
218.00	3312.0
221.10	9321.0
221.90	2507.0
223.10	5313.0
224.00	42960.0
225.00	10818.0
225.90	1465.0
227.00	22528.0
228.00	3816.0
229.00	4586.0
229.90	705.0
231.00	1654.0
233.00	494.0
234.00	1415.0
235.00	1488.0
236.00	1339.0
237.00	2055.0
237.90	363.0
238.90	831.0
240.00	599.0
241.00	1611.0
242.00	2154.0
243.00	3025.0
244.00	36128.0
245.00	4334.0
246.00	7965.0
247.00	1832.0
248.10	375.0
248.90	1398.0
250.90	470.0
251.90	631.0

Instrument :
BNA_N
ClientSampleId :
DFTPP151

252.90	855.0
255.00	187264.0
256.00	26480.0
257.00	2674.0
257.90	12283.0
258.90	2030.0
259.90	356.0
264.00	510.0
265.00	4994.0
265.90	673.0
269.90	441.0
270.90	524.0
272.10	526.0
273.00	5141.0
274.00	15010.0
275.00	86840.0
276.00	10437.0
277.00	7854.0
278.00	1487.0
280.90	359.0
282.00	316.0
283.00	761.0
284.00	771.0
285.00	1530.0
285.90	320.0
289.00	325.0
290.00	376.0
290.90	354.0
292.10	336.0
292.90	2166.0
294.00	797.0
294.80	573.0
296.00	25456.0
297.00	3508.0
301.00	421.0
301.80	541.0
303.00	2969.0
304.00	1044.0
308.00	308.0
310.00	496.0
312.90	467.0
313.90	1516.0
314.90	2817.0
315.90	1593.0
316.90	575.0
320.90	1004.0
321.90	380.0
323.00	8231.0
323.90	1139.0
326.90	1858.0
328.00	1006.0
328.80	325.0
332.10	553.0
333.10	653.0
334.00	6398.0
335.00	1517.0
341.10	1128.0
341.90	311.0
345.90	2449.0
351.90	2591.0
353.00	1511.0
354.10	2785.0
355.00	789.0
364.90	13077.0
365.90	2638.0
369.90	370.0
372.00	5030.0
372.90	840.0
382.90	901.0
389.90	706.0
391.90	488.0
400.90	308.0
402.00	1553.0
402.90	1917.0
404.00	662.0
420.90	1980.0
422.00	2177.0
423.00	13753.0
424.00	3850.0
424.90	836.0
441.10	44632.0
442.00	280256.0
443.00	61104.0
444.10	4779.0

Instrument :
BNA_N
ClientSampleId :
DFTPP151

m/z	Abundance
38.10	944.0
39.10	4467.0
39.90	932.0
41.20	330.0
44.00	1487.0
48.90	690.0
50.00	17352.0
51.10	58568.0
52.10	3266.0
55.00	701.0
55.90	1488.0
57.00	3596.0
61.00	487.0
62.00	857.0
63.10	2618.0
64.00	586.0
65.00	1499.0
68.00	1092.0
69.00	54216.0
73.00	1706.0
74.00	8036.0
75.00	10019.0
76.10	4002.0
77.00	63008.0
78.00	4951.0
79.00	4840.0
80.00	3993.0
81.00	5541.0
81.90	1211.0
83.00	875.0
84.90	1145.0
85.90	1352.0
87.10	666.0
91.10	825.0
92.00	1607.0
93.00	10153.0
93.90	1050.0
98.00	8794.0
99.00	6207.0
99.90	690.0
101.00	2461.0
102.90	1013.0
103.90	2139.0
104.90	1804.0
106.10	978.0
107.00	20992.0
108.00	3092.0
109.00	517.0
110.00	33696.0
111.00	6186.0
112.00	652.0
116.00	1411.0
117.00	29440.0
118.00	1862.0
118.90	334.0
120.20	330.0
122.10	1875.0
123.00	2582.0
123.90	1138.0
125.10	833.0
127.00	66088.0
128.00	4780.0
129.00	31568.0
130.00	2337.0
131.10	698.0
134.00	1189.0
135.00	2373.0
136.00	808.0
137.10	1389.0
138.00	461.0
140.00	407.0
140.90	4593.0
142.10	1433.0
143.00	1088.0
144.00	354.0
146.00	1192.0
146.90	2691.0
148.00	6111.0
149.10	1176.0
149.90	557.0
151.00	339.0
151.90	693.0
152.90	1554.0
154.10	806.0
155.00	2315.0
155.90	3625.0
156.90	490.0
157.80	633.0
159.00	805.0
160.00	1051.0
161.00	2251.0
161.80	419.0

Instrument :
BNA_N
ClientSampleId :
DFTPP151

162.20	352.0
164.00	611.0
165.00	1441.0
166.00	747.0
167.00	10613.0
168.00	6009.0
168.90	612.0
171.00	470.0
172.10	916.0
173.00	1035.0
174.00	1873.0
175.00	3188.0
176.10	444.0
177.00	2128.0
178.00	772.0
179.00	7356.0
179.90	4575.0
181.00	2251.0
182.00	344.0
182.70	354.0
183.80	615.0
185.00	3520.0
186.00	23656.0
187.00	6235.0
188.00	654.0
189.00	2086.0
191.10	789.0
192.00	2690.0
193.00	2382.0
194.10	462.0
196.00	4035.0
198.00	146112.0
199.00	9553.0
200.10	841.0
201.50	750.0
203.00	1468.0
204.00	6331.0
205.00	10134.0
206.00	44144.0
207.10	6345.0
208.10	1632.0
208.90	926.0
210.00	1097.0
210.90	1990.0
216.00	1349.0
217.00	11019.0
218.00	1413.0
221.00	4554.0
222.10	1354.0
223.00	2642.0
224.00	24320.0
225.10	5726.0
226.00	744.0
227.00	12728.0
228.00	1368.0
228.90	2388.0
231.10	885.0
234.00	630.0
234.90	688.0
235.90	780.0
236.90	722.0
239.00	386.0
240.10	474.0
240.90	749.0
242.10	1616.0
243.10	1361.0
244.00	17800.0
244.90	1936.0
246.00	4919.0
246.90	1041.0
248.90	1059.0
252.90	574.0
255.00	111112.0
256.00	14121.0
257.10	817.0
257.90	6502.0
258.90	1088.0
263.80	558.0
265.00	3173.0
266.10	506.0
271.00	392.0
272.10	554.0
273.10	2867.0
274.00	8447.0
275.00	46104.0
276.10	6202.0
277.10	4873.0
278.00	1158.0
280.90	578.0
282.90	690.0
284.00	500.0
285.00	675.0
292.10	319.0
292.90	1124.0
294.10	340.0
294.90	365.0

Instrument :
BNA_N
ClientSampleId :
DFTPP151

296.00	15525.0
297.00	1837.0
303.00	1901.0
304.10	662.0
314.00	813.0
314.90	2090.0
316.00	1390.0
320.90	413.0
322.00	415.0
323.00	4501.0
324.00	1086.0
326.90	1192.0
327.90	696.0
331.80	350.0
334.00	2936.0
335.10	820.0
341.00	395.0
345.90	1064.0
352.00	2010.0
352.90	665.0
354.00	1302.0
354.80	331.0
364.90	9058.0
365.90	1370.0
370.90	471.0
371.90	2900.0
373.00	775.0
383.00	855.0
390.10	466.0
402.00	1183.0
402.90	1565.0
404.10	549.0
421.10	1753.0
421.90	1165.0
423.00	9004.0
424.00	2308.0
441.00	31192.0
442.00	204160.0
443.00	39600.0
444.00	3777.0

Instrument :
BNA_N
ClientSampleId :
DFTPP151