




Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111324\
Data File : BN035061.D
Acq On : 13 Nov 2024 12:01
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 14 15:14:42 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Nov 09 04:05:27 2024
Response via : Initial Calibration



TIC: BN035061.D\data.ms

(70) Pentachlorophenol (C)




16.592min (-0.024) 20966.65 ng

response 202879

Ion	Exp%	Act%
265.70	100.00	100.00
268.00	62.20	64.29
264.00	61.60	63.61
0.00	0.00	0.00

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111324\
Data File : BN035061.D
Acq On : 13 Nov 2024 12:01
Operator : RC/JU
Sample : DFTPP
Misc :
ALS Vial : 1 Sample Multiplier: 1

Quant Time: Nov 14 15:14:42 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270E-Tune.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Sat Nov 09 04:05:27 2024
Response via : Initial Calibration



TIC: BN035061.D\data.ms

(77) Benzidine

19.204min (-0.000) 0.00 ng

response 1216684

Ion	Exp%	Act%
184.00	100.00	100.00
185.00	15.50	14.30
183.00	13.20	12.08
0.00	0.00	0.00