

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111324\
 Data File : BN035069.D
 Acq On : 13 Nov 2024 16:51
 Operator : RC/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN111324

Quant Time: Nov 13 17:18:36 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN111324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 13 17:18:14 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	89	0.00
2	1,4-Dioxane	0.400	0.371	7.3	90	0.00
3	n-Nitrosodimethylamine	0.400	0.368	8.0	85	0.00
4 S	2-Fluorophenol	0.400	0.368	8.0	88	0.00
5 S	Phenol-d6	0.400	0.359	10.3	88	0.00
6	bis(2-Chloroethyl)ether	0.400	0.370	7.5	89	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	93	0.00
8 S	Nitrobenzene-d5	0.400	0.358	10.5	90	0.00
9	Naphthalene	0.400	0.366	8.5	91	0.00
10	Hexachlorobutadiene	0.400	0.373	6.8	90	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.367	8.3	91	0.00
12	2-Methylnaphthalene	0.400	0.368	8.0	92	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	92	0.00
14 S	2,4,6-Tribromophenol	0.400	0.322	19.5	89	0.00
15 S	2-Fluorobiphenyl	0.400	0.357	10.8	91	0.00
16	Acenaphthylene	0.400	0.348	13.0	92	0.00
17	Acenaphthene	0.400	0.359	10.3	92	0.00
18	Fluorene	0.400	0.352	12.0	91	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	94	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.330	17.5	89	0.00
21	4-Bromophenyl-phenylether	0.400	0.371	7.3	90	0.00
22	Hexachlorobenzene	0.400	0.379	5.3	93	0.00
23	Atrazine	0.400	0.370	7.5	92	0.00
24	Pentachlorophenol	0.400	0.324	19.0	91	0.00
25	Phenanthrene	0.400	0.375	6.3	92	0.00
26	Anthracene	0.400	0.366	8.5	90	0.00
27 SURR	Fluoranthene-d10	0.400	0.374	6.5	92	0.00
28	Fluoranthene	0.400	0.372	7.0	91	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	94	0.00
30	Pyrene	0.400	0.376	6.0	91	0.00
31 S	Terphenyl-d14	0.400	0.381	4.8	93	0.00
32	Benzo(a)anthracene	0.400	0.376	6.0	93	0.00
33	Chrysene	0.400	0.386	3.5	94	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.338	15.5	90	0.00
35 I	Perylene-d12	0.400	0.400	0.0	92	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.375	6.3	91	0.00
37	Benzo(b)fluoranthene	0.400	0.379	5.3	92	0.00
38	Benzo(k)fluoranthene	0.400	0.380	5.0	92	0.00
39 C	Benzo(a)pyrene	0.400	0.378	5.5	92	0.00
40	Dibenzo(a,h)anthracene	0.400	0.373	6.8	90	0.00
41	Benzo(g,h,i)perylene	0.400	0.372	7.0	89	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN111324