

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111324\
 Data File : BN035080.D
 Acq On : 13 Nov 2024 23:24
 Operator : RC/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 14 02:08:20 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN111324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 13 17:18:14 2024
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	90	0.00
2	1,4-Dioxane	0.400	0.377	5.8	92	0.00
3	n-Nitrosodimethylamine	0.400	0.352	12.0	82	0.00
4 S	2-Fluorophenol	0.400	0.367	8.3	89	0.00
5 S	Phenol-d6	0.400	0.379	5.3	94	0.00
6	bis(2-Chloroethyl)ether	0.400	0.373	6.8	90	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	92	0.00
8 S	Nitrobenzene-d5	0.400	0.362	9.5	91	0.00
9	Naphthalene	0.400	0.373	6.8	92	0.00
10	Hexachlorobutadiene	0.400	0.374	6.5	90	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.374	6.5	93	0.00
12	2-Methylnaphthalene	0.400	0.375	6.3	94	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	94	0.00
14 S	2,4,6-Tribromophenol	0.400	0.340	15.0	96	0.00
15 S	2-Fluorobiphenyl	0.400	0.361	9.8	94	0.00
16	Acenaphthylene	0.400	0.352	12.0	95	0.00
17	Acenaphthene	0.400	0.357	10.8	94	0.00
18	Fluorene	0.400	0.364	9.0	96	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	96	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.331	17.3	91	0.00
21	4-Bromophenyl-phenylether	0.400	0.379	5.3	95	0.00
22	Hexachlorobenzene	0.400	0.375	6.3	95	0.00
23	Atrazine	0.400	0.353	11.8	90	0.00
24	Pentachlorophenol	0.400	0.308	23.0	88	0.00
25	Phenanthrene	0.400	0.387	3.3	98	0.00
26	Anthracene	0.400	0.373	6.8	94	0.00
27 SURR	Fluoranthene-d10	0.400	0.372	7.0	94	0.00
28	Fluoranthene	0.400	0.376	6.0	94	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	92	0.00
30	Pyrene	0.400	0.392	2.0	93	0.00
31 S	Terphenyl-d14	0.400	0.388	3.0	93	0.00
32	Benzo(a)anthracene	0.400	0.375	6.3	91	0.00
33	Chrysene	0.400	0.392	2.0	94	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.324	19.0	84	0.00
35 I	Perylene-d12	0.400	0.400	0.0	90	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.374	6.5	88	0.00
37	Benzo(b)fluoranthene	0.400	0.388	3.0	92	0.00
38	Benzo(k)fluoranthene	0.400	0.391	2.3	92	0.00
39 C	Benzo(a)pyrene	0.400	0.377	5.8	89	0.00
40	Dibenzo(a,h)anthracene	0.400	0.373	6.8	87	0.00
41	Benzo(g,h,i)perylene	0.400	0.363	9.3	85	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4