

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111324\
 Data File : BN035065.D
 Acq On : 13 Nov 2024 14:28
 Operator : RC/JU
 Sample : SSTDICC0.8
 Misc :
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8

Quant Time: Nov 13 16:44:09 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN111324.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 13 16:42:01 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.553	152	3429	0.400	ng	0.00
7) Naphthalene-d8	10.319	136	8612	0.400	ng	0.00
13) Acenaphthene-d10	14.190	164	6454	0.400	ng	0.00
19) Phenanthrene-d10	16.933	188	17302	0.400	ng	# 0.00
29) Chrysene-d12	21.134	240	19278	0.400	ng	# 0.00
35) Perylene-d12	23.300	264	21751	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.177	112	7450	0.856	ng	0.00
5) Phenol-d6	6.737	99	9293	0.851	ng	0.00
8) Nitrobenzene-d5	8.685	82	6336	0.848	ng	-0.01
11) 2-Methylnaphthalene-d10	11.912	152	13119	0.855	ng	0.00
14) 2,4,6-Tribromophenol	15.679	330	3866	0.831	ng	0.00
15) 2-Fluorobiphenyl	12.811	172	22399	0.855	ng	0.00
27) Fluoranthene-d10	18.973	212	44583	0.841	ng	0.00
31) Terphenyl-d14	19.582	244	34460	0.852	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.177	88	2545	0.817	ng	99
3) n-Nitrosodimethylamine	3.473	42	2466	0.853	ng	# 96
6) bis(2-Chloroethyl)ether	6.990	93	7043	0.860	ng	99
9) Naphthalene	10.362	128	19223	0.855	ng	98
10) Hexachlorobutadiene	10.661	225	5612	0.851	ng	# 100
12) 2-Methylnaphthalene	11.988	142	14179	0.855	ng	99
16) Acenaphthylene	13.902	152	23520	0.853	ng	99
17) Acenaphthene	14.254	154	15511	0.858	ng	99
18) Fluorene	15.238	166	22649	0.852	ng	99
20) 4,6-Dinitro-2-methylph...	15.323	198	3013	0.837	ng	# 87
21) 4-Bromophenyl-phenylether	16.138	248	9403	0.853	ng	# 85
22) Hexachlorobenzene	16.250	284	9574	0.837	ng	99
23) Atrazine	16.424	200	8427	0.852	ng	# 92
24) Pentachlorophenol	16.597	266	4494	0.841	ng	98
25) Phenanthrene	16.970	178	38573	0.847	ng	99
26) Anthracene	17.069	178	35437	0.847	ng	100
28) Fluoranthene	19.006	202	53391	0.853	ng	100
30) Pyrene	19.368	202	54892	0.855	ng	100
32) Benzo(a)anthracene	21.116	228	56460	0.841	ng	98
33) Chrysene	21.169	228	57039	0.858	ng	98
34) Bis(2-ethylhexyl)phtha...	21.080	149	29628	0.848	ng	100
36) Indeno(1,2,3-cd)pyrene	25.452	276	75270	0.867	ng	100
37) Benzo(b)fluoranthene	22.657	252	62606	0.855	ng	97
38) Benzo(k)fluoranthene	22.701	252	61702	0.843	ng	96
39) Benzo(a)pyrene	23.206	252	54251	0.842	ng	# 95
40) Dibenzo(a,h)anthracene	25.466	278	59829	0.870	ng	99
41) Benzo(g,h,i)perylene	26.104	276	63167	0.864	ng	98


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111324\
Data File : BN035065.D
Acq On : 13 Nov 2024 14:28
Operator : RC/JU
Sample : SSTDICC0.8
Misc :
ALS Vial : 5 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Quant Time: Nov 13 16:44:09 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\8270-SIM-BN111324.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 13 16:42:01 2024
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.553 min Scan# 64
 Delta R.T. 0.000 min
 Lab File: BN035065.D
 Acq: 13 Nov 2024 14:28




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion:152 Resp: 3429
 Ion Ratio Lower Upper
 152 100
 150 151.6 124.5 186.7
 115 56.8 47.8 71.6



#2
 1,4-Dioxane
 Concen: 0.817 ng
 RT: 3.177 min Scan# 41
 Delta R.T. 0.000 min
 Lab File: BN035065.D
 Acq: 13 Nov 2024 14:28

Tgt Ion: 88 Resp: 2545
 Ion Ratio Lower Upper
 88 100
 43 19.8 16.9 25.3
 58 49.3 39.0 58.4





#3

n-Nitrosodimethylamine

Concen: 0.853 ng

RT: 3.473 min Scan# 81

Delta R.T. -0.007 min

Lab File: BN035065.D

Acq: 13 Nov 2024 14:28

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion: 42 Resp: 2466

Ion Ratio Lower Upper

42 100

74 155.8 120.8 181.2

44 5.4 2.9 4.3



#4

2-Fluorophenol

Concen: 0.856 ng

RT: 5.177 min Scan# 318

Delta R.T. 0.000 min

Lab File: BN035065.D

Acq: 13 Nov 2024 14:28




Tgt Ion: 112 Resp: 7450


Ion Ratio Lower Upper

112 100

64 49.1 39.7 59.5

63 29.6 23.0 34.4









#5
Phenol-d6
Concen: 0.851 ng
RT: 6.737 min Scan# 51
Delta R.T. 0.000 min
Lab File: BN035065.D
Acq: 13 Nov 2024 14:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion: 99 Resp: 9293
Ion Ratio Lower Upper
99 100
42 14.5 11.4 17.0
71 42.9 34.6 51.8



#6
bis(2-Chloroethyl)ether
Concen: 0.860 ng
RT: 6.990 min Scan# 569
Delta R.T. 0.000 min
Lab File: BN035065.D
Acq: 13 Nov 2024 14:28

Tgt Ion: 93 Resp: 7043
Ion Ratio Lower Upper
93 100
63 60.1 47.4 71.2
95 31.8 26.2 39.4









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.319 min Scan# 940
Delta R.T. 0.000 min
Lab File: BN035065.D
Acq: 13 Nov 2024 14:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:	136	Resp:	8612
Ion	Ratio	Lower	Upper
136	100		
137	12.1	9.8	14.8
54	4.6	4.0	6.0
68	4.8	4.2	6.2



#8
Nitrobenzene-d5
Concen: 0.848 ng
RT: 8.685 min Scan# 787
Delta R.T. -0.011 min
Lab File: BN035065.D
Acq: 13 Nov 2024 14:28

Tgt Ion:	82	Resp:	6336
Ion	Ratio	Lower	Upper
82	100		
128	34.6	36.5	54.7#
54	40.8	28.7	43.1









#9
Naphthalene
Concen: 0.855 ng
RT: 10.362 min Scan# 944
Delta R.T. 0.000 min
Lab File: BN035065.D
Acq: 13 Nov 2024 14:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:128 Resp: 19223
Ion Ratio Lower Upper
128 100
129 11.4 9.6 14.4
127 13.8 11.6 17.4



#10
Hexachlorobutadiene
Concen: 0.851 ng
RT: 10.661 min Scan# 972
Delta R.T. 0.000 min
Lab File: BN035065.D
Acq: 13 Nov 2024 14:28

Tgt Ion:225 Resp: 5612
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.5 51.5 77.3









#11
2-Methylnaphthalene-d10
Concen: 0.855 ng
RT: 11.912 min Scan# 1159
Delta R.T. -0.004 min
Lab File: BN035065.D
Acq: 13 Nov 2024 14:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:152 Resp: 13119
Ion Ratio Lower Upper
152 100
151 20.3 16.2 24.4



#12
2-Methylnaphthalene
Concen: 0.855 ng
RT: 11.988 min Scan# 1179
Delta R.T. 0.000 min
Lab File: BN035065.D
Acq: 13 Nov 2024 14:28

Tgt Ion:142 Resp: 14179
Ion Ratio Lower Upper
142 100
141 88.4 70.1 105.1
115 42.4 34.4 51.6









#13
 Acenaphthene-d10
 Concen: 0.400 ng
 RT: 14.190 min Scan# 1491
 Delta R.T. 0.004 min
 Lab File: BN035065.D
 Acq: 13 Nov 2024 14:28




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion:164 Resp: 6454
 Ion Ratio Lower Upper
 164 100
 162 102.9 82.2 123.2
 160 48.3 40.2 60.4



#14
 2,4,6-Tribromophenol
 Concen: 0.831 ng
 RT: 15.679 min Scan# 1628
 Delta R.T. -0.007 min
 Lab File: BN035065.D
 Acq: 13 Nov 2024 14:28

Tgt Ion:330 Resp: 3866
 Ion Ratio Lower Upper
 330 100
 332 95.2 76.9 115.3
 141 36.9 29.6 44.4









#15
2-Fluorobiphenyl
Concen: 0.855 ng
RT: 12.811 min Scan# 1361
Delta R.T. 0.004 min
Lab File: BN035065.D
Acq: 13 Nov 2024 14:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:172 Resp: 22399
Ion Ratio Lower Upper
172 100
171 34.1 28.2 42.4
170 22.9 19.0 28.6



#16
Acenaphthylene
Concen: 0.853 ng
RT: 13.902 min Scan# 1464
Delta R.T. -0.007 min
Lab File: BN035065.D
Acq: 13 Nov 2024 14:28

Tgt Ion:152 Resp: 23520
Ion Ratio Lower Upper
152 100
151 19.6 15.8 23.8
153 13.1 10.2 15.4





#17

Acenaphthene

Concen: 0.858 ng

RT: 14.254 min Scan# 1497

Delta R.T. 0.004 min

Lab File: BN035065.D

Acq: 13 Nov 2024 14:28

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:154 Resp: 15511

Ion Ratio Lower Upper

154 100

153 114.0 91.6 137.4

152 56.6 45.8 68.6



#18

Fluorene

Concen: 0.852 ng

RT: 15.238 min Scan# 1589

Delta R.T. -0.007 min

Lab File: BN035065.D

Acq: 13 Nov 2024 14:28




Tgt Ion:166 Resp: 22649


Ion Ratio Lower Upper

166 100

165 98.3 79.1 118.7

167 13.6 10.6 16.0









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 16.933 min Scan# 1729
Delta R.T. -0.007 min
Lab File: BN035065.D
Acq: 13 Nov 2024 14:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:188 Resp: 17302
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 7.5 4.3 6.5#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.837 ng
RT: 15.323 min Scan# 1597
Delta R.T. -0.007 min
Lab File: BN035065.D
Acq: 13 Nov 2024 14:28

Tgt Ion:198 Resp: 3013
Ion Ratio Lower Upper
198 100
51 24.1 29.9 44.9#
105 30.0 23.7 35.5









#21
4-Bromophenyl-phenylether
Concen: 0.853 ng
RT: 16.138 min Scan# 1665
Delta R.T. -0.007 min
Lab File: BN035065.D
Acq: 13 Nov 2024 14:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:248 Resp: 9403
Ion Ratio Lower Upper
248 100
250 93.3 81.2 121.8
141 55.6 29.1 43.7#



#22
Hexachlorobenzene
Concen: 0.837 ng
RT: 16.250 min Scan# 1674
Delta R.T. 0.006 min
Lab File: BN035065.D
Acq: 13 Nov 2024 14:28

Tgt Ion:284 Resp: 9574
Ion Ratio Lower Upper
284 100
142 34.4 28.2 42.4
249 32.9 26.2 39.2









#23
Atrazine
Concen: 0.852 ng
RT: 16.424 min Scan# 1
Delta R.T. 0.006 min
Lab File: BN035065.D
Acq: 13 Nov 2024 14:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:200 Resp: 8427
Ion Ratio Lower Upper
200 100
173 19.4 22.6 33.8#
215 49.1 37.8 56.6



#24
Pentachlorophenol
Concen: 0.841 ng
RT: 16.597 min Scan# 1702
Delta R.T. 0.006 min
Lab File: BN035065.D
Acq: 13 Nov 2024 14:28

Tgt Ion:266 Resp: 4494
Ion Ratio Lower Upper
266 100
264 62.3 49.4 74.0
268 63.1 52.6 79.0









#25
Phenanthrene
Concen: 0.847 ng
RT: 16.970 min Scan# 1732
Delta R.T. -0.007 min
Lab File: BN035065.D
Acq: 13 Nov 2024 14:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:178 Resp: 38573
Ion Ratio Lower Upper
178 100
176 19.3 15.2 22.8
179 15.3 12.6 18.8



#26
Anthracene
Concen: 0.847 ng
RT: 17.069 min Scan# 1740
Delta R.T. 0.006 min
Lab File: BN035065.D
Acq: 13 Nov 2024 14:28

Tgt Ion:178 Resp: 35437
Ion Ratio Lower Upper
178 100
176 18.6 14.6 22.0
179 15.2 12.2 18.2









#27
Fluoranthene-d10
Concen: 0.841 ng
RT: 18.973 min Scan# 1955
Delta R.T. -0.002 min
Lab File: BN035065.D
Acq: 13 Nov 2024 14:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:212 Resp: 44583
Ion Ratio Lower Upper
212 100
106 7.5 6.0 9.0
104 4.4 3.5 5.3



#28
Fluoranthene
Concen: 0.853 ng
RT: 19.006 min Scan# 1963
Delta R.T. 0.002 min
Lab File: BN035065.D
Acq: 13 Nov 2024 14:28

Tgt Ion:202 Resp: 53391
Ion Ratio Lower Upper
202 100
101 6.3 5.2 7.8
203 17.3 13.8 20.8





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.134 min Scan# 2408

Delta R.T. -0.007 min

Lab File: BN035065.D

Acq: 13 Nov 2024 14:28

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:240 Resp: 19278

Ion Ratio Lower Upper

240 100

120 7.1 3.8 5.6#

236 28.4 22.2 33.2



#30

Pyrene

Concen: 0.855 ng

RT: 19.368 min Scan# 2041

Delta R.T. 0.002 min

Lab File: BN035065.D

Acq: 13 Nov 2024 14:28




Tgt Ion:202 Resp: 54892


Ion Ratio Lower Upper

202 100

200 21.0 16.7 25.1

203 17.8 14.2 21.4









#31
 Terphenyl-d14
 Concen: 0.852 ng
 RT: 19.582 min Scan# 2087
 Delta R.T. -0.007 min
 Lab File: BN035065.D
 Acq: 13 Nov 2024 14:28




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.8


Tgt Ion:244 Resp: 34460
 Ion Ratio Lower Upper
 244 100
 212 9.5 7.6 11.4
 122 8.6 6.1 9.1



#32
 Benzo(a)anthracene
 Concen: 0.841 ng
 RT: 21.116 min Scan# 2406
 Delta R.T. -0.007 min
 Lab File: BN035065.D
 Acq: 13 Nov 2024 14:28

Tgt Ion:228 Resp: 56460
 Ion Ratio Lower Upper
 228 100
 226 27.8 21.4 32.0
 229 19.3 15.7 23.5





#33

Chrysene

Concen: 0.858 ng

RT: 21.169 min Scan# 2411

Delta R.T. -0.007 min

Lab File: BN035065.D

Acq: 13 Nov 2024 14:28

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.8





Tgt Ion:228 Resp: 57039

Ion Ratio Lower Upper

228 100

226 30.5 23.7 35.5

229 19.1 16.2 24.4



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.848 ng

RT: 21.080 min Scan# 2402

Delta R.T. 0.002 min

Lab File: BN035065.D

Acq: 13 Nov 2024 14:28




Tgt Ion:149 Resp: 29628


Ion Ratio Lower Upper

149 100

167 29.1 23.2 34.8

279 4.0 3.2 4.8









#35
Perylene-d12
Concen: 0.400 ng
RT: 23.300 min Scan# 21
Delta R.T. -0.004 min
Lab File: BN035065.D
Acq: 13 Nov 2024 14:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:264 Resp: 21751
Ion Ratio Lower Upper
264 100
260 25.6 20.9 31.3
265 38.8 35.4 53.2



#36
Indeno(1,2,3-cd)pyrene
Concen: 0.867 ng
RT: 25.452 min Scan# 3592
Delta R.T. -0.001 min
Lab File: BN035065.D
Acq: 13 Nov 2024 14:28

Tgt Ion:276 Resp: 75270
Ion Ratio Lower Upper
276 100
138 16.2 13.0 19.4
277 24.7 19.8 29.6









#37
Benzo(b)fluoranthene
Concen: 0.855 ng
RT: 22.657 min Scan# 2636
Delta R.T. -0.001 min
Lab File: BN035065.D
Acq: 13 Nov 2024 14:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:252 Resp: 62606
Ion Ratio Lower Upper
252 100
253 22.4 19.3 28.9
125 6.5 6.2 9.2



#38
Benzo(k)fluoranthene
Concen: 0.843 ng
RT: 22.701 min Scan# 2651
Delta R.T. -0.001 min
Lab File: BN035065.D
Acq: 13 Nov 2024 14:28

Tgt Ion:252 Resp: 61702
Ion Ratio Lower Upper
252 100
253 22.3 19.5 29.3
125 6.7 6.5 9.7









#39
Benzo(a)pyrene
Concen: 0.842 ng
RT: 23.206 min Scan# 21
Delta R.T. -0.001 min
Lab File: BN035065.D
Acq: 13 Nov 2024 14:28




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8


Tgt Ion:252 Resp: 54251
Ion Ratio Lower Upper
252 100
253 22.6 20.1 30.1
125 7.4 7.5 11.3#



#40
Dibenzo(a,h)anthracene
Concen: 0.870 ng
RT: 25.466 min Scan# 3597
Delta R.T. -0.001 min
Lab File: BN035065.D
Acq: 13 Nov 2024 14:28

Tgt Ion:278 Resp: 59829
Ion Ratio Lower Upper
278 100
139 11.1 9.3 13.9
279 24.0 19.7 29.5





#41
Benzo(g,h,i)perylene
Concen: 0.864 ng
RT: 26.104 min Scan# 3812
Delta R.T. -0.001 min
Lab File: BN035065.D
Acq: 13 Nov 2024 14:28

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.8

Tgt Ion:276 Resp: 63167
Ion Ratio Lower Upper
276 100
277 23.7 19.9 29.9
138 14.0 11.6 17.4

