


Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN111518\
 Data File : BN003551.D
 Acq On : 15 Nov 2018 11:53
 Operator : MJ/SJ
 Sample : DFTPP51
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP51

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN111418MA.M
 Title : SVOA CALIBRATION
 Last Update : Thu Nov 15 17:18:23 2018



AutoFind: Scans 2489, 2490, 2491; Background Corrected with Scan 2484

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	25.9	73632	PASS
68	69	0.00	2	1.8	1491	PASS
69	198	0.00	100	29.0	82237	PASS
70	69	0.00	2	0.5	447	PASS
127	198	10	80	41.3	117176	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	283946	PASS
199	198	5	9	6.8	19342	PASS
275	198	10	60	30.1	85453	PASS
365	198	1	100	3.5	9869	PASS
441	198	0.01	100	12.2	34600	PASS
442	198	50	100	78.7	223549	PASS
443	442	15	24	19.3	43093	PASS

m/z	Abundance
38.00	907.0
39.00	4697.0
40.00	453.0
44.00	1357.0
49.10	561.0
50.00	16984.0
51.00	67168.0
52.00	3484.0
55.00	427.0
56.00	2021.0
57.00	4602.0
61.00	946.0
62.00	1012.0
63.00	2850.0
64.00	368.0
65.00	1409.0
68.00	1301.0
69.00	73544.0
70.00	365.0
73.00	641.0
74.00	7569.0
75.00	12110.0
76.00	4025.0
77.00	85880.0
78.00	5974.0
79.00	5359.0
80.00	4388.0
81.00	5715.0
82.00	1592.0
83.00	1479.0
85.00	1165.0
86.00	1587.0
87.00	779.0
91.00	1375.0
92.00	1579.0
93.00	9198.0
94.00	726.0
96.00	435.0
98.00	7269.0
99.00	6806.0
100.00	650.0
101.00	3453.0
103.00	1304.0
104.00	2163.0
105.00	2082.0
106.10	752.0
107.00	24912.0
108.00	4211.0
109.10	914.0
110.00	51728.0
111.00	7446.0
112.00	908.0
116.00	1437.0
117.00	20152.0
118.00	1728.0
120.00	331.0
122.00	1675.0
123.00	2790.0
124.00	1308.0
124.90	988.0
127.00	97872.0
128.00	7268.0
129.00	36464.0
130.00	3373.0
131.00	563.0
132.00	345.0
134.00	962.0
135.00	2876.0
136.00	1127.0
137.00	1734.0
137.90	354.0
140.00	382.0
141.00	4781.0
142.00	1596.0
142.90	1093.0
144.90	358.0
146.00	908.0
147.00	2344.0
148.00	5364.0
149.00	1328.0
149.90	359.0
151.10	553.0
151.70	439.0
153.00	1488.0
154.00	1140.0
155.00	2621.0
156.00	3947.0
157.00	949.0
157.90	740.0
158.90	625.0
160.00	1550.0

Instrument :
BNA_N
ClientSampleId :
DFTPP51

161.00	2173.0
161.90	648.0
165.00	1815.0
166.00	1441.0
167.00	9430.0
168.00	6907.0
169.00	1117.0
170.00	418.0
170.90	454.0
171.90	944.0
173.00	1091.0
174.00	2097.0
175.00	3857.0
176.00	1110.0
176.90	1887.0
178.00	656.0
179.00	7016.0
180.00	5255.0
181.00	2401.0
181.90	397.0
184.00	489.0
185.00	3393.0
186.00	27704.0
187.00	8063.0
188.10	1007.0
189.00	1601.0
190.90	855.0
192.00	2338.0
193.00	2526.0
193.90	513.0
195.00	363.0
196.00	6507.0
198.00	213696.0
199.00	15179.0
200.00	1056.0
201.50	1070.0
203.00	1323.0
204.00	7212.0
205.00	12061.0
206.00	54648.0
207.00	7718.0
208.00	2030.0
208.90	679.0
210.00	865.0
211.00	2050.0
211.90	481.0
214.90	602.0
216.00	1175.0
217.00	13124.0
218.00	1814.0
221.00	10257.0
223.00	3242.0
224.00	29024.0
225.00	7795.0
226.00	951.0
227.00	11201.0
228.00	1735.0
229.00	2447.0
230.00	308.0
231.00	1081.0
234.00	755.0
235.00	765.0
235.90	552.0
237.00	1368.0
239.00	537.0
239.90	373.0
241.00	748.0
242.00	1476.0
243.10	1743.0
244.00	23960.0
245.00	3339.0
246.00	4319.0
246.90	921.0
248.90	905.0
252.90	464.0
255.00	121384.0
256.00	18776.0
257.00	1542.0
258.00	6341.0
259.00	1145.0
265.00	2407.0
266.00	343.0
273.00	3745.0
274.00	9762.0
275.00	58568.0
276.00	8145.0
277.00	4537.0
278.00	842.0
283.00	538.0
284.00	375.0
285.00	858.0
293.00	1066.0
296.00	14468.0
297.00	1861.0
302.00	303.0

Instrument :

BNA_N

ClientSampleId :

DFTPP51

303.00	1643.0
304.00	504.0
314.00	608.0
315.00	1646.0
316.00	1042.0
320.90	488.0
323.00	4604.0
324.00	948.0
326.90	1057.0
328.00	462.0
332.00	370.0
332.90	491.0
334.00	3100.0
335.00	918.0
341.00	595.0
345.90	970.0
352.00	1481.0
352.90	1082.0
354.00	1500.0
354.90	356.0
365.00	6320.0
366.00	929.0
371.00	363.0
372.00	2381.0
372.90	606.0
383.00	624.0
390.00	324.0
401.90	962.0
403.00	1296.0
404.00	468.0
420.90	993.0
422.00	1071.0
423.00	7894.0
424.00	2591.0
425.10	350.0
441.10	18376.0
442.00	120376.0
443.10	24248.0
444.00	2410.0

Instrument :
BNA_N
ClientSampleId :
DFTPP51

m/z	Abundance
37.00	399.0
38.00	966.0
39.00	6211.0
40.00	491.0
44.00	1314.0
49.00	713.0
50.00	21552.0
51.00	86184.0
52.00	4464.0
54.90	440.0
56.00	2611.0
57.00	6071.0
61.00	1268.0
62.00	1328.0
63.00	3606.0
64.00	556.0
65.00	1888.0
68.10	1831.0
69.00	97008.0
70.00	477.0
73.00	860.0
74.00	9978.0
75.00	15974.0
76.00	5284.0
77.00	113736.0
78.00	7527.0
79.00	7131.0
80.00	5791.0
81.00	8240.0
82.00	2144.0
83.00	1875.0
85.00	1662.0
86.00	2231.0
87.00	1010.0
88.00	390.0
91.00	1883.0
92.00	1981.0
93.00	12654.0
94.00	913.0
95.90	623.0
98.00	9832.0
99.00	8838.0
100.00	827.0
101.00	4662.0
103.00	1429.0
104.00	2881.0
105.00	2812.0
106.10	953.0
107.00	35792.0
108.00	5565.0
109.10	1098.0
110.00	71528.0
111.00	10677.0
112.00	1244.0
112.90	436.0
116.00	1954.0
117.00	28320.0
118.00	2202.0
119.90	515.0
122.00	2515.0
123.00	3758.0
124.00	1690.0
125.00	1571.0
127.00	139328.0
128.00	10513.0
129.00	51792.0
130.00	4567.0
131.00	831.0
132.00	511.0
133.90	1485.0
135.00	4407.0
136.00	1676.0
137.00	2349.0
137.90	500.0
139.00	303.0
140.00	738.0
141.00	7129.0
142.00	2421.0
143.00	1665.0
144.00	333.0
145.00	414.0
146.00	1357.0
147.00	3297.0
148.00	7733.0
149.00	1637.0
149.90	502.0
151.00	940.0
151.80	742.0
153.00	2107.0
154.00	1824.0
155.00	3675.0

Instrument :
BNA_N
ClientSampleId :
DFTPP51

156.00	6231.0
157.10	1208.0
158.00	1292.0
159.00	1007.0
160.00	2191.0
161.00	3381.0
162.00	1040.0
164.00	439.0
165.00	2799.0
166.00	2373.0
167.00	14035.0
168.00	8271.0
169.00	1238.0
170.00	511.0
170.90	626.0
172.00	1242.0
173.00	1756.0
174.00	3076.0
175.00	5522.0
176.00	1791.0
176.90	2593.0
178.00	916.0
179.00	10942.0
180.00	7715.0
181.00	3585.0
182.00	644.0
182.90	340.0
184.00	878.0
185.00	5325.0
186.00	43096.0
187.00	12099.0
188.00	1422.0
189.00	2625.0
189.90	390.0
191.10	1172.0
192.00	3632.0
193.00	4027.0
194.00	759.0
195.20	568.0
196.00	10552.0
198.00	338240.0
199.00	23016.0
200.00	1756.0
201.40	1887.0
203.00	2014.0
204.00	10765.0
205.00	19240.0
206.00	86352.0
207.00	11690.0
208.00	2727.0
209.00	885.0
211.00	3172.0
214.90	851.0
216.00	1562.0
217.00	21200.0
218.00	2708.0
221.00	16480.0
223.00	4916.0
224.00	46704.0
225.00	12222.0
226.00	1310.0
227.00	18008.0
228.00	2698.0
229.00	4254.0
229.90	520.0
231.00	1898.0
233.10	316.0
234.00	1206.0
235.00	1394.0
236.00	874.0
237.00	1768.0
238.90	694.0
240.00	647.0
240.90	1036.0
242.00	2381.0
243.00	2828.0
244.00	40160.0
245.00	5509.0
246.00	7379.0
247.00	1449.0
248.00	312.0
249.00	1609.0
251.00	312.0
252.00	357.0
253.00	871.0
255.00	204672.0
256.00	30896.0
257.00	2445.0
258.00	11145.0
259.00	1856.0
259.90	339.0
260.90	358.0
264.00	312.0
265.00	4370.0
266.00	632.0

Instrument :
BNA_N
ClientSampleId :
DFTPP51

271.00	370.0
272.00	663.0
273.00	6549.0
274.00	17144.0
275.00	102400.0
276.00	13586.0
277.00	7536.0
278.00	1353.0
278.90	317.0
283.00	924.0
284.00	584.0
285.00	1426.0
288.90	335.0
292.00	386.0
293.00	1829.0
294.00	560.0
295.00	577.0
296.00	25712.0
297.00	3476.0
302.10	436.0
303.00	2972.0
304.00	860.0
307.90	358.0
310.00	394.0
314.10	1356.0
315.00	2986.0
316.00	1649.0
317.00	356.0
320.90	849.0
322.00	532.0
323.00	9005.0
324.00	1623.0
327.00	1525.0
328.00	892.0
332.00	731.0
333.00	925.0
334.00	5928.0
335.00	1537.0
341.00	996.0
342.00	306.0
345.90	2000.0
347.00	450.0
352.00	2686.0
353.00	1953.0
354.00	2953.0
355.00	580.0
365.00	11301.0
366.00	1672.0
371.10	648.0
372.00	4972.0
373.00	1251.0
383.00	1304.0
383.90	360.0
390.00	574.0
391.00	422.0
391.90	352.0
402.00	1854.0
403.00	2967.0
404.00	1009.0
421.00	2464.0
422.10	2359.0
423.10	15715.0
424.10	4224.0
425.10	477.0
441.10	39312.0
442.10	252480.0
443.10	49464.0
444.10	4525.0

Instrument :

BNA_N

ClientSampleId :

DFTPP51

m/z	Abundance
38.00	772.0
39.00	4697.0
40.00	494.0
44.00	1510.0
49.00	509.0
50.00	17440.0
51.00	67544.0
52.00	3526.0
54.90	349.0
56.00	2239.0
57.00	4910.0
61.00	880.0
62.00	1099.0
63.00	2818.0
64.00	371.0
65.00	1458.0
68.10	1343.0
69.00	76160.0
70.00	500.0
73.00	731.0
74.00	7959.0
75.00	12270.0
76.10	4209.0
77.00	90920.0
78.00	6397.0
79.00	5882.0
80.00	4515.0
81.00	6296.0
82.00	1554.0
83.00	1423.0
85.00	1347.0
86.00	1944.0
87.00	809.0
87.90	352.0
91.00	1390.0
92.00	1676.0
93.00	10410.0
94.00	654.0
96.10	508.0
98.00	8378.0
99.00	6330.0
100.00	561.0
101.00	3722.0
103.00	1239.0
104.00	2500.0
105.00	2188.0
106.10	764.0
107.00	28728.0
108.00	4555.0
109.10	894.0
110.00	58872.0
111.00	8609.0
111.90	951.0
116.00	1549.0
117.00	22736.0
118.00	1817.0
120.00	367.0
122.00	2350.0
123.00	3077.0
124.00	1318.0
125.00	1283.0
127.00	114328.0
128.00	8573.0
129.00	42800.0
130.00	3614.0
131.10	653.0
132.10	459.0
134.00	1337.0
135.00	3400.0
136.00	1388.0
137.00	1973.0
137.90	525.0
140.00	581.0
141.00	5725.0
142.00	1842.0
143.00	1289.0
143.90	344.0
145.00	370.0
146.00	970.0
147.00	3169.0
148.00	6591.0
149.00	1408.0
150.00	417.0
151.00	812.0
151.70	596.0
153.00	1830.0
154.00	1502.0
155.00	3513.0
156.00	5131.0
157.00	1151.0
158.00	1173.0

Instrument :
BNA_N
ClientSampleId :
DFTPP51

159.00	829.0
160.00	1912.0
161.00	2904.0
161.90	730.0
163.00	306.0
163.90	349.0
165.00	2250.0
166.00	1924.0
167.00	12668.0
168.00	6914.0
169.00	1284.0
170.00	469.0
170.90	496.0
172.00	1093.0
173.00	1443.0
174.00	2635.0
175.00	4922.0
176.00	1465.0
177.00	2522.0
178.00	690.0
179.00	9243.0
180.00	6789.0
181.00	3168.0
182.00	546.0
183.00	350.0
184.00	782.0
185.00	4680.0
186.00	38296.0
187.00	10811.0
188.00	1150.0
189.00	2153.0
189.90	347.0
191.00	1056.0
192.00	3146.0
193.00	3527.0
194.00	785.0
195.10	500.0
196.00	9122.0
198.00	299904.0
199.00	19832.0
200.00	1388.0
201.40	1685.0
203.00	1930.0
204.00	9764.0
205.00	16800.0
206.00	74936.0
207.00	9866.0
208.00	2444.0
208.90	725.0
211.00	3043.0
215.00	808.0
216.00	1663.0
217.00	19328.0
218.00	2624.0
221.00	15152.0
223.00	4888.0
224.00	43480.0
225.00	10883.0
226.00	1125.0
227.00	16640.0
228.00	2285.0
229.00	3721.0
229.90	486.0
231.00	1771.0
234.00	1033.0
235.00	1279.0
236.00	764.0
237.00	1561.0
238.90	606.0
240.00	534.0
240.90	907.0
242.00	2358.0
243.10	2548.0
244.00	37128.0
245.00	4864.0
246.00	6560.0
247.00	1232.0
248.90	1321.0
251.90	367.0
252.90	746.0
255.00	192384.0
256.00	27632.0
257.00	2374.0
258.00	10122.0
259.00	1534.0
263.90	336.0
265.00	3965.0
266.00	581.0
271.00	399.0
272.00	542.0
273.00	5989.0
274.00	15877.0
275.00	95392.0
276.00	12734.0
277.00	7041.0

Instrument :
BNA_N
ClientSampleId :
DFTPP51

278.00	1247.0
278.90	326.0
283.00	753.0
284.00	593.0
285.00	1282.0
288.90	305.0
292.00	450.0
293.00	1830.0
294.00	539.0
296.00	24872.0
297.00	3482.0
301.00	338.0
302.00	480.0
303.00	2960.0
304.00	839.0
308.00	314.0
310.00	339.0
314.00	1300.0
315.00	2744.0
316.00	1603.0
321.00	902.0
322.00	519.0
323.10	8911.0
324.00	1507.0
327.00	1681.0
328.00	943.0
332.00	661.0
333.00	971.0
334.00	5889.0
335.00	1619.0
341.00	975.0
341.90	323.0
346.00	1962.0
346.90	428.0
352.00	2774.0
353.00	2194.0
354.00	3132.0
355.00	559.0
365.00	11987.0
366.00	1706.0
371.00	852.0
372.00	4763.0
373.00	1248.0
383.10	1484.0
384.00	356.0
389.90	641.0
391.00	490.0
392.10	392.0
402.00	2048.0
403.10	3163.0
404.00	1141.0
421.00	2404.0
422.00	2190.0
423.00	18224.0
424.10	4418.0
424.90	508.0
441.10	46112.0
442.10	297792.0
443.10	55568.0
444.00	4962.0

Instrument :
BNA_N
ClientSampleId :
DFTPP51

Scan 2484 (17.598 min): BN003551.D
DFTPP51

m/z	Abundance
39.90	328.0
44.00	1570.0
207.00	525.0

Instrument :
BNA_N
ClientSampleId :
DFTPP51