

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111323\
 Data File : BM042812.D
 Acq On : 16 Nov 2023 19:08
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020146

Quant Time: Nov 16 23:55:22 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM111323.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 15 22:28:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.822	152	34964	20.000	ng/u1	0.00
20) Naphthalene-d8	10.633	136	142562	20.000	ng/u1	# 0.00
38) Acenaphthene-d10	14.480	164	92282	20.000	ng/u1	0.00
64) Phenanthrene-d10	17.233	188	200178	20.000	ng/u1	0.00
79) Chrysene-d12	21.415	240	216386	20.000	ng/u1	0.00
88) Perylene-d12	23.780	264	244593	20.000	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.175	96	8388	7.528	ng/uL	0.00
4) Pyridine-d5	3.610	84	53825	19.419	ng/u1	0.00
7) Phenol-d5	6.987	99	64937	19.083	ng/u1	0.00
9) Bis-(2-Chloroethyl)eth...	7.169	67	49689	20.781	ng/u1	0.00
11) 2-Chlorophenol-d4	7.345	132	48458	18.455	ng/u1	0.00
15) 4-Methylphenol-d8	8.545	113	48656	18.160	ng/u1	0.00
21) Nitrobenzene-d5	9.028	128	22171	17.738	ng/u1	0.00
24) 2-Nitrophenol-d4	9.734	143	24655	16.862	ng/u1	0.00
28) 2,4-Dichlorophenol-d3	10.263	165	49274	18.124	ng/u1	0.00
31) 4-Chloroaniline-d4	10.810	131	58278	16.883	ng/u1	0.00
46) Dimethylphthalate-d6	13.904	166	155843	17.844	ng/u1	0.00
49) Acenaphthylene-d8	14.180	160	173363	18.552	ng/u1	0.00
54) 4-Nitrophenol-d4	14.727	143	12704	12.875	ng/u1	-0.01
60) Fluorene-d10	15.474	176	131518	18.185	ng/u1	0.00
65) 4,6-Dinitro-2-methylph...	15.621	200	21887	14.435	ng/u1	0.00
73) Anthracene-d10	17.333	188	202845	18.081	ng/u1	0.00
81) Pyrene-d10	19.627	212	255708	18.046	ng/u1	0.00
92) Benzo(a)pyrene-d12	23.621	264	264290	18.064	ng/u1	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.210	88	9194	7.748	ng/uL#	80
5) Pyridine	3.634	79	62822	20.844	ng/u1	97
6) Benzaldehyde	6.993	77	44416	22.750	ng/u1	92
8) Phenol	7.016	94	65221	19.269	ng/u1	96
10) Bis(2-Chloroethyl)ether	7.263	93	54278	19.413	ng/u1	96
12) 2-Chlorophenol	7.375	128	50254	19.232	ng/u1#	88
13) 2-Methylphenol	8.275	108	46013	18.230	ng/u1	98
14) 2,2'-oxybis(1-Chloropr...	8.334	45	66387	13.043	ng/u1	94
16) Acetophenone	8.675	105	81952	18.190	ng/u1	95
17) N-Nitroso-di-n-propyla...	8.640	70	52247	19.050	ng/u1	95
18) 4-Methylphenol	8.610	108	48716	17.752	ng/u1	98
19) Hexachloroethane	8.875	117	28554	19.605	ng/u1	91
22) Nitrobenzene	9.069	77	74235	19.651	ng/u1	98
23) Isophorone	9.575	82	139058	18.482	ng/u1	99
25) 2-Nitrophenol	9.769	139	26161	17.610	ng/u1	93
26) 2,4-Dimethylphenol	9.810	107	55947	17.602	ng/u1	98
27) Bis(2-Chloroethoxy)met...	10.063	93	73358	19.294	ng/u1	99
29) 2,4-Dichlorophenol	10.286	162	47502	18.454	ng/u1	96
30) Naphthalene	10.686	128	163956	18.357	ng/u1	99
32) 4-Chloroaniline	10.839	127	54689	16.514	ng/u1	97
33) Hexachlorobutadiene	10.904	225	42171	18.547	ng/u1	97
34) Caprolactam	11.657	113	8055	10.554	ng/u1	90
35) 4-Chloro-3-methylphenol	11.945	107	49428	17.920	ng/u1	91
36) 2-Methylnaphthalene	12.298	142	107233	17.914	ng/u1	100

Data Path : Z:\svoasrv\HPCHEM1\BNA_M\Data\BM111323\
 Data File : BM042812.D
 Acq On : 16 Nov 2023 19:08
 Operator : MA/JU
 Sample : SSTDCCC020
 Misc :
 ALS Vial : 54 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD020146


Quant Time: Nov 16 23:55:22 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM111323.MA.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 15 22:28:07 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
37) 1-Methylnaphthalene	12.516	142	111506	17.717	ng/ul	97
39) 1,2,4,5-Tetrachloroben...	12.651	216	70612	19.511	ng/ul	98
40) Hexachlorocyclopentadiene	12.592	237	26401	15.408	ng/ul	94
41) 2,4,6-Trichlorophenol	12.916	196	40926	18.513	ng/ul	95
42) 2,4,5-Trichlorophenol	12.992	196	41885	19.622	ng/ul	99
43) 1,1'-Biphenyl	13.316	154	146550	19.042	ng/ul	98
44) 2-Chloronaphthalene	13.363	162	117198	18.930	ng/ul	98
45) 2-Nitroaniline	13.616	65	38356	18.996	ng/ul	93
47) Dimethylphthalate	13.951	163	154740	18.190	ng/ul	99
48) 2,6-Dinitrotoluene	14.104	165	28433	17.791	ng/ul	97
50) Acenaphthylene	14.210	152	195952	18.714	ng/ul	99
51) 3-Nitroaniline	14.451	138	21448	16.412	ng/ul	99
52) Acenaphthene	14.545	153	132829	18.757	ng/ul	96
53) 2,4-Dinitrophenol	14.663	184	8451	10.635	ng/ul	89
55) 4-Nitrophenol	14.745	109	15672	10.093	ng/ul	95
56) Dibenzofuran	14.880	168	180228	18.336	ng/ul	97
57) 2,4-Dinitrotoluene	14.886	165	41068	17.504	ng/ul	92
58) 2,3,4,6-Tetrachlorophenol	15.104	232	39967	18.037	ng/ul#	94
59) Diethylphthalate	15.298	149	157022	17.603	ng/ul	99
61) Fluorene	15.527	166	154032	18.458	ng/ul	97
62) 4-Chlorophenyl-phenyle...	15.521	204	84960	18.891	ng/ul	95
63) 4-Nitroaniline	15.616	138	19665	16.506	ng/ul	97
66) 4,6-Dinitro-2-methylph...	15.633	198	24470	15.852	ng/ul#	87
67) N-Nitrosodiphenylamine	15.751	169	118016	18.797	ng/ul	98
68) 4-Bromophenyl-phenylether	16.421	248	49862	18.713	ng/ul	86
69) Hexachlorobenzene	16.498	284	59999	18.420	ng/ul	97
70) Atrazine	16.698	200	43993	18.191	ng/ul	99
71) Pentachlorophenol	16.868	266	30559	16.292	ng/ul	99
72) Phenanthrene	17.274	178	235621	18.841	ng/ul	98
74) Anthracene	17.368	178	233092	18.531	ng/ul	99
75) 1,2,3,4-Tetrachloroben...	13.269	216	71913	20.317	ng/uL	98
76) Pentachlorobenzene	14.774	250	73721	19.849	ng/uL	98
77) Carbazole	17.662	167	179694	17.301	ng/ul	99
78) Di-n-butylphthalate	18.174	149	265105	17.455	ng/ul	98
80) Fluoranthene	19.292	202	289039	17.890	ng/ul	98
82) Pyrene	19.656	202	306428	17.888	ng/ul	99
83) Butylbenzylphthalate	20.539	149	118800	16.236	ng/ul	91
84) 3,3'-Dichlorobenzidine	21.350	252	92882	16.771	ng/ul	100
85) Benzo(a)anthracene	21.398	228	311794	18.356	ng/ul	99
86) Bis(2-ethylhexyl)phtha...	21.286	149	178108	16.044	ng/ul	99
87) Chrysene	21.456	228	299144	18.159	ng/ul	100
89) Di-n-octyl phthalate	22.192	149	311888	15.786	ng/ul	100
90) Benzo(b)fluoranthene	23.050	252	312670	18.732	ng/ul	99
91) Benzo(k)fluoranthene	23.097	252	328470	18.486	ng/ul	99
93) Benzo(a)pyrene	23.674	252	298867	18.374	ng/ul	99
94) Indeno(1,2,3-cd)pyrene	26.238	276	378138	18.706	ng/ul	97
95) Dibenzo(a,h)anthracene	26.250	278	304944	18.196	ng/ul	98
96) Benzo(g,h,i)perylene	27.009	276	296605	18.478	ng/ul	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Nov 16 23:55:22 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_M\Methods\SFAM-EPA-BM111323.MA.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Nov 15 22:28:07 2023
Response via : Initial Calibration









#2
1,4-Dioxane
Concen: 7.748 ng/uL
RT: 3.210 min Scan# 38
Delta R.T. -0.006 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08




Instrument :
BNA_N
ClientSampleId :
SSTD020146


Tgt Ion: 88 Resp: 9194
Ion Ratio Lower Upper
88 100
43 54.5 34.3 51.5#
58 92.7 59.4 89.0#



#5
Pyridine
Concen: 20.844 ng/uL
RT: 3.634 min Scan# 110
Delta R.T. -0.006 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08

Tgt Ion: 79 Resp: 62822
Ion Ratio Lower Upper
79 100
52 92.9 74.6 111.8
51 60.5 44.6 67.0









#6
Benzaldehyde
Concen: 22.750 ng/ul
RT: 6.993 min Scan# 681
Delta R.T. -0.000 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08




Instrument :
BNA_N
ClientSampleId :
SSTD020146


Tgt Ion: 77 Resp: 44416
Ion Ratio Lower Upper
77 100
105 77.8 69.6 104.4
106 75.5 64.2 96.2



#8
Phenol
Concen: 19.269 ng/ul
RT: 7.016 min Scan# 685
Delta R.T. -0.006 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08

Tgt Ion: 94 Resp: 65221
Ion Ratio Lower Upper
94 100
65 39.5 30.3 45.5
66 53.4 39.9 59.9









#10
Bis(2-Chloroethyl)ether
Concen: 19.413 ng/ul
RT: 7.263 min Scan# 71
Delta R.T. -0.006 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08




Instrument :
BNA_N
ClientSampleId :
SSTD020146


Tgt Ion: 93 Resp: 54278
Ion Ratio Lower Upper
93 100
63 90.0 68.6 102.8
95 31.7 24.8 37.2



#12
2-Chlorophenol
Concen: 19.232 ng/ul
RT: 7.375 min Scan# 746
Delta R.T. -0.006 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08

Tgt Ion:128 Resp: 50254
Ion Ratio Lower Upper
128 100
64 66.7 42.9 64.3#
130 30.3 23.9 35.9





#13

2-Methylphenol

Concen: 18.230 ng/ul

RT: 8.275 min Scan# 899

Delta R.T. -0.006 min

Lab File: BM042812.D

Acq: 16 Nov 2023 19:08

Instrument :

BNA_N

ClientSampleId :





SSTD020146

Tgt Ion:108 Resp: 46013

Ion Ratio Lower Upper

108 100

107 89.6 70.4 105.6



#14

2,2'-oxybis(1-Chloropropane)

Concen: 13.043 ng/ul

RT: 8.334 min Scan# 909

Delta R.T. -0.006 min

Lab File: BM042812.D

Acq: 16 Nov 2023 19:08




Tgt Ion: 45 Resp: 66387


Ion Ratio Lower Upper

45 100

77 12.6 12.2 18.2

79 8.2 8.2 12.2





#16

Acetophenone

Concen: 18.190 ng/ul

RT: 8.675 min Scan# 968

Delta R.T. -0.006 min

Lab File: BM042812.D

Acq: 16 Nov 2023 19:08

Instrument :

BNA_N

ClientSampleId :

SSTD020146





Tgt Ion:105 Resp: 81952

Ion Ratio Lower Upper

105 100

77 93.7 71.6 107.4

51 48.8 35.4 53.2



#17

N-Nitroso-di-n-propylamine

Concen: 19.050 ng/ul

RT: 8.640 min Scan# 961

Delta R.T. -0.006 min

Lab File: BM042812.D

Acq: 16 Nov 2023 19:08

Tgt Ion: 70 Resp: 52247




Ion Ratio Lower Upper


70 100

42 84.0 63.0 94.6

101 7.2 7.0 10.4

130 16.1 11.8 17.8





#18

4-Methylphenol

Concen: 17.752 ng/ul

RT: 8.610 min Scan# 91

Delta R.T. -0.006 min

Lab File: BM042812.D

Acq: 16 Nov 2023 19:08

Instrument :

BNA_N

ClientSampleId :





SSTD020146

Tgt Ion:108 Resp: 48716

Ion Ratio Lower Upper

108 100

107 113.6 89.4 134.0



#19

Hexachloroethane

Concen: 19.605 ng/ul

RT: 8.875 min Scan# 1001

Delta R.T. -0.006 min

Lab File: BM042812.D

Acq: 16 Nov 2023 19:08




Tgt Ion:117 Resp: 28554


Ion Ratio Lower Upper

117 100

201 98.1 90.0 135.0

199 64.1 52.2 78.4









#22
Nitrobenzene
Concen: 19.651 ng/ul
RT: 9.069 min Scan# 1103
Delta R.T. -0.006 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08




Instrument :
BNA_N
ClientSampleId :
SSTD020146


Tgt Ion	Ratio	Lower	Upper
77	100		
123	34.1	28.4	42.6
65	14.8	12.4	18.6



#23
Isophorone
Concen: 18.482 ng/ul
RT: 9.575 min Scan# 1120
Delta R.T. -0.006 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08

Tgt Ion	Ratio	Lower	Upper
82	100		
95	7.0	6.4	9.6
138	15.3	12.1	18.1









#25
2-Nitrophenol
Concen: 17.610 ng/ul
RT: 9.769 min Scan# 1153
Delta R.T. -0.006 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08




Instrument :
BNA_N
ClientSampleId :
SSTD020146


Tgt Ion:139	Resp:	26161
Ion Ratio	Lower	Upper
139	100	
65	81.2	59.1 88.7
109	45.1	33.9 50.9



#26
2,4-Dimethylphenol
Concen: 17.602 ng/ul
RT: 9.810 min Scan# 1160
Delta R.T. -0.006 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08

Tgt Ion:107	Resp:	55947
Ion Ratio	Lower	Upper
107	100	
121	44.8	38.0 57.0
122	74.8	60.5 90.7





#27

Bis(2-Chloroethoxy)methane

Concen: 19.294 ng/ul

RT: 10.063 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM042812.D

Acq: 16 Nov 2023 19:08

Instrument :

BNA_N

ClientSampleId :

SSTD020146





Tgt Ion: 93 Resp: 73358

Ion Ratio Lower Upper

93 100

95 32.8 26.5 39.7

123 10.6 8.2 12.2



#29

2,4-Dichlorophenol

Concen: 18.454 ng/ul

RT: 10.286 min Scan# 1241

Delta R.T. -0.006 min

Lab File: BM042812.D

Acq: 16 Nov 2023 19:08




Tgt Ion:162 Resp: 47502


Ion Ratio Lower Upper

162 100

164 63.9 53.4 80.2

98 38.4 28.4 42.6





#30

Naphthalene

Concen: 18.357 ng/ul

RT: 10.686 min Scan# 11

Delta R.T. -0.006 min

Lab File: BM042812.D

Acq: 16 Nov 2023 19:08

Instrument :

BNA_N

ClientSampleId :

SSTD020146





Tgt Ion:128 Resp: 163956

Ion Ratio Lower Upper

128 100

129 11.3 8.6 13.0

127 13.2 10.5 15.7



#32

4-Chloroaniline

Concen: 16.514 ng/ul

RT: 10.839 min Scan# 1335

Delta R.T. -0.000 min

Lab File: BM042812.D




Acq: 16 Nov 2023 19:08


Tgt Ion:127 Resp: 54689

Ion Ratio Lower Upper

127 100

129 31.5 26.7 40.1





#33

Hexachlorobutadiene

Concen: 18.547 ng/ul

RT: 10.904 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM042812.D

Acq: 16 Nov 2023 19:08

Instrument :

BNA_N

ClientSampleId :

SSTD020146





Tgt Ion:225 Resp: 42171

Ion Ratio Lower Upper

225 100

223 64.6 53.0 79.4

227 65.7 50.5 75.7



#34

Caprolactam

Concen: 10.554 ng/ul

RT: 11.657 min Scan# 1474

Delta R.T. -0.006 min

Lab File: BM042812.D

Acq: 16 Nov 2023 19:08




Tgt Ion:113 Resp: 8055


Ion Ratio Lower Upper

113 100

55 254.9 184.2 276.2

56 171.8 134.3 201.5









#35
4-Chloro-3-methylphenol
Concen: 17.920 ng/ul
RT: 11.945 min Scan# 1523
Delta R.T. -0.006 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08




Instrument :
BNA_N
ClientSampleId :
SSTD020146


Tgt Ion:107 Resp: 49428
Ion Ratio Lower Upper
107 100
144 25.1 18.3 27.5
142 79.2 56.3 84.5



#36
2-Methylnaphthalene
Concen: 17.914 ng/ul
RT: 12.298 min Scan# 1583
Delta R.T. -0.006 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08

Tgt Ion:142 Resp: 107233
Ion Ratio Lower Upper
142 100
141 91.6 73.6 110.4





#37

1-Methylnaphthalene

Concen: 17.717 ng/ul

RT: 12.516 min Scan# 1

Delta R.T. -0.006 min

Lab File: BM042812.D

Acq: 16 Nov 2023 19:08

Instrument :

BNA_N

ClientSampleId :

SSTD020146





Tgt Ion:142 Resp: 111506

Ion Ratio Lower Upper

142 100

141 94.8 73.3 109.9

116 4.3 3.9 5.9



#39

1,2,4,5-Tetrachlorobenzene

Concen: 19.511 ng/ul

RT: 12.651 min Scan# 1643

Delta R.T. -0.006 min

Lab File: BM042812.D

Acq: 16 Nov 2023 19:08

Tgt Ion:216 Resp: 70612




Ion Ratio Lower Upper


216 100

214 78.1 64.1 96.1

179 18.4 15.5 23.3

108 17.9 15.1 22.7









#40
Hexachlorocyclopentadiene
Concen: 15.408 ng/ul
RT: 12.592 min Scan# 1633
Delta R.T. -0.006 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08




Instrument :
BNA_N
ClientSampleId :
SSTD020146


Tgt Ion:237 Resp: 26401
Ion Ratio Lower Upper
237 100
235 61.0 53.3 79.9
272 12.9 10.7 16.1



#41
2,4,6-Trichlorophenol
Concen: 18.513 ng/ul
RT: 12.916 min Scan# 1688
Delta R.T. -0.006 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08

Tgt Ion:196 Resp: 40926
Ion Ratio Lower Upper
196 100
198 96.5 81.0 121.4
200 33.5 24.1 36.1









#42
2,4,5-Trichlorophenol
Concen: 19.622 ng/ul
RT: 12.992 min Scan# 1701
Delta R.T. -0.006 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08




Instrument :
BNA_N
ClientSampleId :
SSTD020146


Tgt Ion:196 Resp: 41885
Ion Ratio Lower Upper
196 100
198 95.5 77.4 116.0
200 32.4 26.2 39.4



#43
1,1'-Biphenyl
Concen: 19.042 ng/ul
RT: 13.316 min Scan# 1756
Delta R.T. -0.006 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08

Tgt Ion:154 Resp: 146550
Ion Ratio Lower Upper
154 100
153 39.8 32.4 48.6
76 15.8 10.7 16.1









#44
2-Chloronaphthalene
Concen: 18.930 ng/ul
RT: 13.363 min Scan# 11
Delta R.T. -0.006 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08




Instrument :
BNA_N
ClientSampleId :
SSTD020146


Tgt Ion:162 Resp: 117198
Ion Ratio Lower Upper
162 100
164 32.1 24.2 36.2
127 39.6 30.8 46.2



#45
2-Nitroaniline
Concen: 18.996 ng/ul
RT: 13.616 min Scan# 1807
Delta R.T. -0.000 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08

Tgt Ion: 65 Resp: 38356
Ion Ratio Lower Upper
65 100
92 57.6 49.1 73.7
138 69.6 61.8 92.8









#47
Dimethylphthalate
Concen: 18.190 ng/ul
RT: 13.951 min Scan# 1864
Delta R.T. -0.006 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08




Instrument :
BNA_N
ClientSampleId :
SSTD020146


Tgt Ion:163 Resp: 154740
Ion Ratio Lower Upper
163 100
194 4.5 3.2 4.8
164 9.5 7.9 11.9



#48
2,6-Dinitrotoluene
Concen: 17.791 ng/ul
RT: 14.104 min Scan# 1890
Delta R.T. -0.000 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08

Tgt Ion:165 Resp: 28433
Ion Ratio Lower Upper
165 100
89 83.4 64.1 96.1
121 23.8 18.4 27.6









#50
Acenaphthylene
Concen: 18.714 ng/ul
RT: 14.210 min Scan# 1908
Delta R.T. -0.006 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08




Instrument :
BNA_N
ClientSampleId :
SSTD020146


Tgt Ion:152 Resp: 195952
Ion Ratio Lower Upper
152 100
151 19.6 16.0 24.0
153 13.0 10.5 15.7



#51
3-Nitroaniline
Concen: 16.412 ng/ul
RT: 14.451 min Scan# 1949
Delta R.T. -0.000 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08

Tgt Ion:138 Resp: 21448
Ion Ratio Lower Upper
138 100
108 8.9 8.7 13.1
92 137.4 109.4 164.0





#52

Acenaphthene

Concen: 18.757 ng/ul

RT: 14.545 min Scan# 1965

Delta R.T. -0.006 min

Lab File: BM042812.D

Acq: 16 Nov 2023 19:08

Instrument :

BNA_N

ClientSampleId :

SSTD020146





Tgt Ion:153 Resp: 132829

Ion Ratio Lower Upper

153 100

152 49.3 38.2 57.2

154 85.3 71.3 106.9



#53

2,4-Dinitrophenol

Concen: 10.635 ng/ul

RT: 14.663 min Scan# 1985

Delta R.T. -0.000 min

Lab File: BM042812.D

Acq: 16 Nov 2023 19:08




Tgt Ion:184 Resp: 8451


Ion Ratio Lower Upper

184 100

63 112.7 79.1 118.7

154 75.2 65.7 98.5









#55
4-Nitrophenol
Concen: 10.093 ng/ul
RT: 14.745 min Scan# 1999
Delta R.T. -0.006 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08




Instrument :
BNA_N
ClientSampleId :
SSTD020146


Tgt Ion:109 Resp: 15672
Ion Ratio Lower Upper
109 100
139 85.8 67.1 100.7
65 142.8 107.1 160.7



#56
Dibenzofuran
Concen: 18.336 ng/ul
RT: 14.880 min Scan# 2022
Delta R.T. -0.006 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08

Tgt Ion:168 Resp: 180228
Ion Ratio Lower Upper
168 100
139 40.1 30.8 46.2









#57
2,4-Dinitrotoluene
Concen: 17.504 ng/ul
RT: 14.886 min Scan# 20
Delta R.T. -0.006 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08




Instrument :
BNA_N
ClientSampleId :
SSTD020146


Tgt Ion:165 Resp: 41068
Ion Ratio Lower Upper
165 100
63 112.8 83.5 125.3
182 3.6 2.4 3.6



#58
2,3,4,6-Tetrachlorophenol
Concen: 18.037 ng/ul
RT: 15.104 min Scan# 2060
Delta R.T. -0.000 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08

Tgt Ion:232 Resp: 39967
Ion Ratio Lower Upper
232 100
131 44.9 38.9 58.3
130 3.5 1.9 2.9#
166 26.6 23.9 35.9









#59
Diethylphthalate
Concen: 17.603 ng/ul
RT: 15.298 min Scan# 2093
Delta R.T. -0.006 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08




Instrument :
BNA_N
ClientSampleId :
SSTD020146


Tgt Ion:149 Resp: 157022
Ion Ratio Lower Upper
149 100
177 21.9 17.0 25.6
150 11.8 9.7 14.5



#61
Fluorene
Concen: 18.458 ng/ul
RT: 15.527 min Scan# 2132
Delta R.T. -0.006 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08

Tgt Ion:166 Resp: 154032
Ion Ratio Lower Upper
166 100
165 102.5 79.3 118.9
167 13.6 10.6 15.8









#62
4-Chlorophenyl-phenylether
Concen: 18.891 ng/ul
RT: 15.521 min Scan# 2132
Delta R.T. -0.006 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08




Instrument :
BNA_N
ClientSampleId :
SSTD020146


Tgt Ion:204 Resp: 84960
Ion Ratio Lower Upper
204 100
206 30.6 27.0 40.4
141 53.8 46.0 69.0



#63
4-Nitroaniline
Concen: 16.506 ng/ul
RT: 15.616 min Scan# 2147
Delta R.T. -0.006 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08

Tgt Ion:138 Resp: 19665
Ion Ratio Lower Upper
138 100
92 67.4 47.4 71.0
108 161.9 129.5 194.3









#66
4,6-Dinitro-2-methylphenol
Concen: 15.852 ng/ul
RT: 15.633 min Scan# 2151
Delta R.T. -0.006 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08




Instrument :
BNA_N
ClientSampleId :
SSTD020146


Tgt Ion:198 Resp: 24470
Ion Ratio Lower Upper
198 100
182 5.4 3.4 5.0#
77 41.5 26.9 40.3#



#67
N-Nitrosodiphenylamine
Concen: 18.797 ng/ul
RT: 15.751 min Scan# 2170
Delta R.T. -0.006 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08

Tgt Ion:169 Resp: 118016
Ion Ratio Lower Upper
169 100
168 68.5 55.8 83.6
167 37.9 31.4 47.2









#68
4-Bromophenyl-phenylether
Concen: 18.713 ng/ul
RT: 16.421 min Scan# 21
Delta R.T. -0.000 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08




Instrument :
BNA_N
ClientSampleId :
SSTD020146


Tgt Ion:248 Resp: 49862
Ion Ratio Lower Upper
248 100
250 92.0 84.3 126.5
141 61.2 60.1 90.1



#69
Hexachlorobenzene
Concen: 18.420 ng/ul
RT: 16.498 min Scan# 2297
Delta R.T. -0.006 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08


Tgt Ion:284 Resp: 59999
Ion Ratio Lower Upper
284 100
142 36.7 30.5 45.7
249 34.3 25.3 37.9








#70
Atrazine
Concen: 18.191 ng/ul
RT: 16.698 min Scan# 21
Delta R.T. -0.006 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08

Instrument :
BNA_N
ClientSampleId :
SSTD020146







Tgt Ion:200 Resp: 43993
Ion Ratio Lower Upper
200 100
173 23.8 19.8 29.6
215 47.1 38.2 57.2



#71
Pentachlorophenol
Concen: 16.292 ng/ul
RT: 16.868 min Scan# 2360
Delta R.T. -0.006 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08

Tgt Ion:266 Resp: 30559
Ion Ratio Lower Upper
266 100
264 63.1 50.3 75.5
268 63.4 52.1 78.1









#72
Phenanthrene
Concen: 18.841 ng/ul
RT: 17.274 min Scan# 2430
Delta R.T. -0.006 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08




Instrument :
BNA_N
ClientSampleId :
SSTD020146


Tgt Ion:178 Resp: 235621
Ion Ratio Lower Upper
178 100
179 15.0 11.4 17.2
176 20.0 15.5 23.3



#74
Anthracene
Concen: 18.531 ng/ul
RT: 17.368 min Scan# 2445
Delta R.T. -0.006 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08

Tgt Ion:178 Resp: 233092
Ion Ratio Lower Upper
178 100
179 15.5 12.0 18.0
176 19.4 15.7 23.5









#75
1,2,3,4-Tetrachlorobenzene
Concen: 20.317 ng/uL
RT: 13.269 min Scan# 1
Delta R.T. -0.006 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08




Instrument :
BNA_N
ClientSampleId :
SSTD020146


Tgt Ion:216 Resp: 71913
Ion Ratio Lower Upper
216 100
214 76.9 62.6 94.0
218 48.1 39.3 58.9



#76
Pentachlorobenzene
Concen: 19.849 ng/uL
RT: 14.774 min Scan# 2004
Delta R.T. -0.000 min
Lab File: BM042812.D
Acq: 16 Nov 2023 19:08

Tgt Ion:250 Resp: 73721
Ion Ratio Lower Upper
250 100
248 60.9 51.0 76.6
252 64.3 51.8 77.8









#77
 Carbazole
 Concen: 17.301 ng/ul
 RT: 17.662 min Scan# 2495
 Delta R.T. -0.000 min
 Lab File: BM042812.D
 Acq: 16 Nov 2023 19:08




Instrument :
 BNA_N
 ClientSampleId :
 SSTD020146


Tgt Ion:167 Resp: 179694
 Ion Ratio Lower Upper
 167 100
 166 22.0 17.2 25.8
 139 13.2 10.8 16.2



#78
 Di-n-butylphthalate
 Concen: 17.455 ng/ul
 RT: 18.174 min Scan# 2582
 Delta R.T. -0.006 min
 Lab File: BM042812.D
 Acq: 16 Nov 2023 19:08

Tgt Ion:149 Resp: 265105
 Ion Ratio Lower Upper
 149 100
 150 9.1 7.7 11.5
 104 7.4 5.3 7.9





#80

Fluoranthene

Concen: 17.890 ng/ul

RT: 19.292 min Scan# 21

Delta R.T. -0.006 min

Lab File: BM042812.D

Acq: 16 Nov 2023 19:08

Instrument :

BNA_N

ClientSampleId :

SSTD020146





Tgt Ion:202 Resp: 289039

Ion Ratio Lower Upper

202 100

101 9.4 6.6 9.8

100 7.1 5.6 8.4



#82

Pyrene

Concen: 17.888 ng/ul

RT: 19.656 min Scan# 2834

Delta R.T. -0.006 min

Lab File: BM042812.D

Acq: 16 Nov 2023 19:08




Tgt Ion:202 Resp: 306428


Ion Ratio Lower Upper

202 100

101 10.6 8.1 12.1

100 8.5 6.9 10.3





#83

Butylbenzylphthalate

Concen: 16.236 ng/ul

RT: 20.539 min Scan# 2984

Delta R.T. -0.006 min

Lab File: BM042812.D

Acq: 16 Nov 2023 19:08

Instrument :

BNA_N

ClientSampleId :

SSTD020146





Tgt Ion:149 Resp: 118800

Ion Ratio Lower Upper

149 100

91 95.3 69.3 103.9

206 18.4 16.6 25.0



#84

3,3'-Dichlorobenzidine

Concen: 16.771 ng/ul

RT: 21.350 min Scan# 3122

Delta R.T. -0.006 min

Lab File: BM042812.D

Acq: 16 Nov 2023 19:08




Tgt Ion:252 Resp: 92882


Ion Ratio Lower Upper

252 100

254 65.0 51.8 77.6

126 9.7 7.8 11.8





#85

Benzo(a)anthracene

Concen: 18.356 ng/ul

RT: 21.398 min Scan# 3111

Delta R.T. -0.006 min

Lab File: BM042812.D

Acq: 16 Nov 2023 19:08

Instrument :

BNA_N

ClientSampleId :

SSTD020146





Tgt Ion:228 Resp: 311794

Ion Ratio Lower Upper

228 100

229 19.1 15.8 23.6

226 27.3 22.5 33.7



#86

Bis(2-ethylhexyl)phthalate

Concen: 16.044 ng/ul

RT: 21.286 min Scan# 3111

Delta R.T. -0.006 min

Lab File: BM042812.D

Acq: 16 Nov 2023 19:08




Tgt Ion:149 Resp: 178108


Ion Ratio Lower Upper

149 100

167 25.5 20.6 31.0

279 4.3 3.5 5.3





#87

Chrysene

Concen: 18.159 ng/ul

RT: 21.456 min Scan# 3140

Delta R.T. -0.000 min

Lab File: BM042812.D

Acq: 16 Nov 2023 19:08

Instrument :

BNA_N

ClientSampleId :

SSTD020146





Tgt Ion:228 Resp: 299144

Ion Ratio Lower Upper

228 100

226 30.7 24.3 36.5

229 19.4 15.7 23.5



#89

Di-n-octyl phthalate

Concen: 15.786 ng/ul




RT: 22.192 min Scan# 3265


Delta R.T. -0.012 min

Lab File: BM042812.D

Acq: 16 Nov 2023 19:08

Tgt Ion:149 Resp: 311888





#90

Benzo(b)fluoranthene

Concen: 18.732 ng/ul

RT: 23.050 min Scan# 3411

Delta R.T. -0.006 min

Lab File: BM042812.D

Acq: 16 Nov 2023 19:08

Instrument :

BNA_N

ClientSampleId :

SSTD020146





Tgt Ion:252 Resp: 312670

Ion Ratio Lower Upper

252 100

253 20.8 17.0 25.6

125 7.4 5.2 7.8



#91

Benzo(k)fluoranthene

Concen: 18.486 ng/ul

RT: 23.097 min Scan# 3419

Delta R.T. -0.006 min

Lab File: BM042812.D

Acq: 16 Nov 2023 19:08




Tgt Ion:252 Resp: 328470


Ion Ratio Lower Upper

252 100

253 21.6 17.5 26.3

125 7.4 5.5 8.3





#93

Benzo(a)pyrene

Concen: 18.374 ng/ul

RT: 23.674 min Scan# 3517

Delta R.T. -0.006 min

Lab File: BM042812.D

Acq: 16 Nov 2023 19:08

Instrument :

BNA_N

ClientSampleId :

SSTD020146





Tgt Ion:252 Resp: 298867

Ion Ratio Lower Upper

252 100

253 21.0 17.0 25.6

125 9.0 6.3 9.5



#94

Indeno(1,2,3-cd)pyrene

Concen: 18.706 ng/ul

RT: 26.238 min Scan# 3953

Delta R.T. -0.006 min

Lab File: BM042812.D

Acq: 16 Nov 2023 19:08




Tgt Ion:276 Resp: 378138


Ion Ratio Lower Upper

276 100

138 18.4 13.8 20.8

277 23.0 19.7 29.5





#95

Dibenzo(a,h)anthracene

Concen: 18.196 ng/ul

RT: 26.250 min Scan# 3955

Delta R.T. -0.012 min

Lab File: BM042812.D

Acq: 16 Nov 2023 19:08

Instrument :

BNA_N

ClientSampleId :

SSTD020146





Tgt Ion:278 Resp: 304944

Ion Ratio Lower Upper

278 100

139 13.1 9.0 13.4

279 23.4 18.8 28.2



#96

Benzo(g,h,i)perylene

Concen: 18.478 ng/ul

RT: 27.009 min Scan# 4084

Delta R.T. -0.012 min

Lab File: BM042812.D

Acq: 16 Nov 2023 19:08

Tgt Ion:276 Resp: 296605

Ion Ratio Lower Upper

276 100

138 15.8 13.0 19.6

277 22.8 18.4 27.6

