

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111523\
 Data File : BN028639.D
 Acq On : 15 Nov 2023 21:57
 Operator : MA/JU
 Sample : 05337-09
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GCDJ4

Quant Time: Nov 15 23:02:28 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 15 18:38:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.700	152	5365	0.400	ng/ul	0.02
4) Naphthalene-d8	10.480	136	24744	0.400	ng/ul	# 0.02
9) Acenaphthene-d10	14.326	164	16281	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	17.082	188	32087	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.299	240	23246	0.400	ng/ul	# 0.00
23) Perylene-d12	23.578	264	30784	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.118	96	30142	5.083	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.085	152	6892	0.185	ng/ul	0.02
18) Fluoranthene-d10	19.117	212	37248	0.498	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.152	88	13201	2.014	ng/ul#	85

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111523\
Data File : BN028639.D
Acq On : 15 Nov 2023 21:57
Operator : MA/JU
Sample : 05337-09
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
GCDJ4

Quant Time: Nov 15 23:02:28 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 15 18:38:25 2023
Response via : Initial Calibration

