

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111523\
 Data File : BN028647.D
 Acq On : 16 Nov 2023 03:20
 Operator : MA/JU
 Sample : 05337-15
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GCDK0

Quant Time: Nov 16 04:39:13 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 16 04:34:21 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.713	152	6543	0.400	ng/ul	0.03
4) Naphthalene-d8	10.474	136	31676	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.326	164	19282	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	17.082	188	37424	0.400	ng/ul	0.00
17) Chrysene-d12	21.297	240	25801	0.400	ng/ul	0.00
23) Perylene-d12	23.574	264	32522	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.118	96	31466	4.351	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.096	152	4905	0.103	ng/ul	0.03
18) Fluoranthene-d10	19.118	212	29943	0.361	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.152	88	1684	0.211	ng/ul#	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Quant Time: Nov 16 04:39:13 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 16 04:34:21 2023
Response via : Initial Calibration

