

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111523\
 Data File : BN028684.D
 Acq On : 17 Nov 2023 03:43
 Operator : MA/JU
 Sample : 05338-10
 Misc :
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GCDL5

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/17/2023
 Supervised By :mohammad ahmed 11/20/2023

Quant Time: Nov 17 04:19:47 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 16 22:38:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.719	152	9851m	0.400	ng/ul	-0.08
4) Naphthalene-d8	10.473	136	37367	0.400	ng/ul	#-0.02
9) Acenaphthene-d10	14.325	164	20428	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	17.081	188	37676m	0.400	ng/ul	0.00
17) Chrysene-d12	21.292	240	23348	0.400	ng/ul	0.00
23) Perylene-d12	23.572	264	29075	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.116	96	40265	3.698	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.112	152	6133	0.109	ng/ul	0.00
18) Fluoranthene-d10	19.117	212	26805	0.357	ng/ul	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.154	88	985	0.082	ng/ul#	72

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111523\
Data File : BN028684.D
Acq On : 17 Nov 2023 03:43
Operator : MA/JU
Sample : 05338-10
Misc :
ALS Vial : 52 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
GCDL5

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/17/2023
Supervised By :mohammad ahmed 11/20/2023

Quant Time: Nov 17 04:19:47 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 16 22:38:14 2023
Response via : Initial Calibration

