

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111523\
 Data File : BN028710.D
 Acq On : 17 Nov 2023 20:41
 Operator : MA/JU
 Sample : 05158-04
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BH371

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/19/2023
 Supervised By :mohammad ahmed 11/20/2023

Quant Time: Nov 17 23:26:13 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 17 09:39:27 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.728	152	7003m	0.400	ng/ul	-0.07
4) Naphthalene-d8	10.479	136	22952m	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.321	164	16379m	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.078	188	31676m	0.400	ng/ul	-0.01
17) Chrysene-d12	21.291	240	24081m	0.400	ng/ul	0.00
23) Perylene-d12	23.570	264	26115m	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.116	96	26646	3.443	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.118	152	7128m	0.206	ng/ul	0.00
18) Fluoranthene-d10	19.114	212	28994m	0.374	ng/ul	0.00
Target Compounds						Qvalue
2) 1,4-Dioxane	3.150	88	754	0.088	ng/ul#	69

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111523\
Data File : BN028710.D
Acq On : 17 Nov 2023 20:41
Operator : MA/JU
Sample : 05158-04
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BH371

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/19/2023
Supervised By :mohammad ahmed 11/20/2023

Quant Time: Nov 17 23:26:13 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 17 09:39:27 2023
Response via : Initial Calibration

