

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111523\
 Data File : BN028717.D
 Acq On : 18 Nov 2023 00:54
 Operator : MA/JU
 Sample : PB156818BS
 Misc :
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS818

Quant Time: Nov 18 02:00:40 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 17 09:39:27 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.808	152	8	0.400	ng/ul	# 0.01
4) Naphthalene-d8	10.511	136	29	0.400	ng/ul	# 0.02
9) Acenaphthene-d10	14.333	164	13	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	17.101	188	18	0.400	ng/ul	# 0.01
17) Chrysene-d12	21.305	240	67	0.400	ng/ul	# 0.01
23) Perylene-d12	23.568	264	3	0.400	ng/ul	#-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.120	96	6234	705.067	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.111	152	8621	197.465	ng/ul	0.00
18) Fluoranthene-d10	19.120	212	28300	131.256	ng/ul	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.150	88	18638	1907.079	ng/ul#	80
5) Naphthalene	10.555	128	18961	233.308	ng/ul	97
7) 2-Methylnaphthalene	12.177	142	9736	181.836	ng/ul	96
8) 1-Methylnaphthalene	12.364	142	13368	245.871	ng/ul	100
10) Acenaphthylene	14.051	152	25059	386.232	ng/ul	98
11) Acenaphthene	14.389	153	20079	420.536	ng/ul	99
12) Fluorene	15.393	166	18836	339.075	ng/ul	99
14) Pentachlorophenol	16.742	266	3936	911.145	ng/ul	99
15) Phenanthrene	17.130	178	26675	482.343	ng/ul	98
16) Anthracene	17.240	178	18557	365.927	ng/ul	96
19) Fluoranthene	19.148	202	36268	127.277	ng/ul	100
20) Pyrene	19.511	202	40404	136.811	ng/ul	97
21) Benzo(a)anthracene	21.285	228	21190	79.986	ng/ul	98
22) Chrysene	21.329	228	30481	117.688	ng/ul	99
24) Benzo(b)fluoranthene	22.886	252	24729	2196.037	ng/ul	89
25) Benzo(k)fluoranthene	22.933	252	30301	2570.015	ng/ul	93
26) Benzo(a)pyrene	23.480	252	25552	2404.049	ng/ul#	81
27) Indeno(1,2,3-cd)pyrene	25.925	276	29536	2504.014	ng/ul#	99
28) Dibenzo(a,h)anthracene	25.959	278	22458	2382.302	ng/ul	96
29) Benzo(g,h,i)perylene	26.636	276	26136	2702.960	ng/ul	97

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111523\
Data File : BN028717.D
Acq On : 18 Nov 2023 00:54
Operator : MA/JU
Sample : PB156818BS
Misc :
ALS Vial : 16 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SLCS818

Quant Time: Nov 18 02:00:40 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 17 09:39:27 2023
Response via : Initial Calibration

