

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111523\
 Data File : BN028737.D
 Acq On : 18 Nov 2023 14:06
 Operator : MA/JU
 Sample : 05158-07
 Misc :
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BH367

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/19/2023
 Supervised By :mohammad ahmed 11/20/2023

Quant Time: Nov 18 23:02:23 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 17 09:39:27 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.736	152	6772m	0.400	ng/ul	-0.06
4) Naphthalene-d8	10.479	136	21284m	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.320	164	16069m	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.081	188	30165m	0.400	ng/ul	0.00
17) Chrysene-d12	21.295	240	20798m	0.400	ng/ul	0.00
23) Perylene-d12	23.569	264	28181m	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.112	96	39873	5.327	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.134	152	6534m	0.204	ng/ul	0.02
18) Fluoranthene-d10	19.113	212	26054m	0.389	ng/ul	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111523\
Data File : BN028737.D
Acq On : 18 Nov 2023 14:06
Operator : MA/JU
Sample : 05158-07
Misc :
ALS Vial : 37 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BH367

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/19/2023
Supervised By :mohammad ahmed 11/20/2023

Quant Time: Nov 18 23:02:23 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 17 09:39:27 2023
Response via : Initial Calibration

