

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111523\
 Data File : BN028641.D
 Acq On : 15 Nov 2023 23:10
 Operator : MA/JU
 Sample : 05337-11
 Misc :
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GCDJ6

Quant Time: Nov 15 23:40:50 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 15 18:38:25 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.700	152	6193	0.400	ng/ul	0.02
4) Naphthalene-d8	10.470	136	28115	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.323	164	17505	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	17.079	188	33653	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.293	240	23531	0.400	ng/ul	# 0.00
23) Perylene-d12	23.578	264	28263	0.400	ng/ul	# 0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.114	96	28234	4.125	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.092	152	5691	0.134	ng/ul	0.03
18) Fluoranthene-d10	19.115	212	29080	0.384	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.152	88	1866	0.247	ng/ul#	86


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111523\
Data File : BN028641.D
Acq On : 15 Nov 2023 23:10
Operator : MA/JU
Sample : 05337-11
Misc :
ALS Vial : 10 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
GCDJ6

Quant Time: Nov 15 23:40:50 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 15 18:38:25 2023
Response via : Initial Calibration





#2
 1,4-Dioxane
 Concen: 0.247 ng/ul
 RT: 3.152 min Scan# 40
 Delta R.T. 0.004 min
 Lab File: BN028641.D
 Acq: 15 Nov 2023 23:10

Instrument :
 BNA_N
 ClientSampleId :
 GCDJ6

Tgt Ion: 88 Resp: 1866

Ion	Ratio	Lower	Upper
88	100		
43	42.0	32.9	49.3
58	35.3	41.8	62.8

