

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111523\
 Data File : BN028651.D
 Acq On : 16 Nov 2023 05:42
 Operator : MA/JU
 Sample : 05337-19
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GCDK5

Quant Time: Nov 16 06:09:46 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 16 04:48:33 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.709	152	6981	0.400	ng/ul	0.03
4) Naphthalene-d8	10.470	136	31735	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.323	164	19735	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	17.079	188	39393	0.400	ng/ul	0.00
17) Chrysene-d12	21.293	240	26873	0.400	ng/ul	# 0.00
23) Perylene-d12	23.573	264	31710	0.400	ng/ul	-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.118	96	40107	5.198	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.092	152	6920	0.145	ng/ul	0.03
18) Fluoranthene-d10	19.115	212	37158	0.430	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.156	88	1746	0.205	ng/ul#	78


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111523\
Data File : BN028651.D
Acq On : 16 Nov 2023 05:42
Operator : MA/JU
Sample : 05337-19
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
GCDK5

Quant Time: Nov 16 06:09:46 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 16 04:48:33 2023
Response via : Initial Calibration





#2
 1,4-Dioxane
 Concen: 0.205 ng/ul
 RT: 3.156 min Scan# 41
 Delta R.T. 0.008 min
 Lab File: BN028651.D
 Acq: 16 Nov 2023 05:42

Instrument :
 BNA_N
 ClientSampleId :
 GCDK5

Tgt Ion:	88	Resp:	1746
Ion	Ratio	Lower	Upper
88	100		
43	45.2	32.9	49.3
58	28.2	41.8	62.8#

