

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111523\
 Data File : BN028655.D
 Acq On : 16 Nov 2023 08:05
 Operator : MA/JU
 Sample : 05337-17
 Misc :
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GCDK3

Quant Time: Nov 16 11:02:07 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 16 07:25:53 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.692	152	6472	0.400	ng/ul	0.01
4) Naphthalene-d8	10.470	136	28458	0.400	ng/ul	# 0.00
9) Acenaphthene-d10	14.323	164	17690	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	17.079	188	36441	0.400	ng/ul	# 0.00
17) Chrysene-d12	21.290	240	26958	0.400	ng/ul	# 0.00
23) Perylene-d12	23.575	264	29928	0.400	ng/ul	#-0.01
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.114	96	25944	3.627	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.081	152	6994	0.163	ng/ul	0.02
18) Fluoranthene-d10	19.115	212	34745	0.401	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.152	88	2442	0.309	ng/ul#	89


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111523\
Data File : BN028655.D
Acq On : 16 Nov 2023 08:05
Operator : MA/JU
Sample : 05337-17
Misc :
ALS Vial : 24 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
GCDK3


Quant Time: Nov 16 11:02:07 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 16 07:25:53 2023
Response via : Initial Calibration





#2
 1,4-Dioxane
 Concen: 0.309 ng/ul
 RT: 3.152 min Scan# 40
 Delta R.T. 0.004 min
 Lab File: BN028655.D
 Acq: 16 Nov 2023 08:05

Instrument :
 BNA_N
 ClientSampleId :
 GCDK3



Tgt Ion:	88	Resp:	2442
Ion	Ratio	Lower	Upper
88	100		
43	41.3	32.9	49.3
58	38.5	41.8	62.8#

