

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111523\
 Data File : BN028688.D
 Acq On : 17 Nov 2023 06:06
 Operator : MA/JU
 Sample : 05338-14
 Misc :
 ALS Vial : 56 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 GCDL9

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/17/2023
 Supervised By :mohammad ahmed 11/20/2023

Quant Time: Nov 17 06:33:39 2023
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110923.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Thu Nov 16 22:38:14 2023
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.711	152	9908m	0.400	ng/ul	-0.08
4) Naphthalene-d8	10.473	136	36060	0.400	ng/ul	#-0.02
9) Acenaphthene-d10	14.320	164	19018	0.400	ng/ul	# 0.00
13) Phenanthrene-d10	17.077	188	35073	0.400	ng/ul	#-0.01
17) Chrysene-d12	21.292	240	23346	0.400	ng/ul	0.00
23) Perylene-d12	23.572	264	30074	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.116	96	37167	3.394	ng/ul	0.00
6) 2-Methylnaphthalene-d10	12.101	152	6955	0.128	ng/ul	-0.01
18) Fluoranthene-d10	19.112	212	28204	0.375	ng/ul	0.00
Target Compounds					Qvalue	
2) 1,4-Dioxane	3.154	88	1549	0.128	ng/ul#	67

(#) = qualifier out of range (m) = manual integration (+) = signals summed


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111523\
Data File : BN028688.D
Acq On : 17 Nov 2023 06:06
Operator : MA/JU
Sample : 05338-14
Misc :
ALS Vial : 56 Sample Multiplier: 1


Instrument :
BNA_N
ClientSampleId :
GCDL9

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/17/2023
Supervised By :mohammad ahmed 11/20/2023


Quant Time: Nov 17 06:33:39 2023
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110923.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Thu Nov 16 22:38:14 2023
Response via : Initial Calibration





#2
1,4-Dioxane
Concen: 0.128 ng/ul
RT: 3.154 min Scan# 183
Delta R.T. -0.000 min
Lab File: BN028688.D
Acq: 17 Nov 2023 06:06

Instrument :
BNA_N
ClientSampleId :
GCDL9



Tgt Ion: 88 Resp: 1549
Ion Ratio Lower Upper
88 100
43 32.9 32.9 49.3
58 18.6 41.8 62.8

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/17/2023
Supervised By :mohammad ahmed 11/20/2023

