

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111524\
 Data File : BN035113.D
 Acq On : 15 Nov 2024 19:24
 Operator : RC/JU
 Sample : PB164839BS
 Misc :
 ALS Vial : 18 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SLCS839

Quant Time: Nov 15 22:27:15 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 13 00:41:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.550	152	2963	0.400	ng/ul	0.00
4) Naphthalene-d8	10.311	136	7595	0.400	ng/ul	-0.01
9) Acenaphthene-d10	14.185	164	6103	0.400	ng/ul	0.00
13) Phenanthrene-d10	16.937	188	14932	0.400	ng/ul	0.00
17) Chrysene-d12	21.137	240	12108	0.400	ng/ul	0.00
23) Perylene-d12	23.299	264	10240	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.140	96	2180	0.773	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.912	152	5495	0.472	ng/ul	0.00
18) Fluoranthene-d10	18.972	212	17086	0.550	ng/ul	0.00
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.170	88	5810	1.880	ng/ul#	91
5) Naphthalene	10.361	128	7902	0.416	ng/ul	97
7) 2-Methylnaphthalene	11.983	142	5894	0.415	ng/ul	100
8) 1-Methylnaphthalene	12.203	142	6917	0.486	ng/ul	99
10) Acenaphthylene	13.903	152	9825	0.394	ng/ul	99
11) Acenaphthene	14.249	153	7545	0.399	ng/ul	99
12) Fluorene	15.240	166	9393	0.391	ng/ul	99
14) Pentachlorophenol	16.591	266	2685	0.778	ng/ul	98
15) Phenanthrene	16.975	178	16210	0.406	ng/ul	99
16) Anthracene	17.068	178	14881	0.397	ng/ul	99
19) Fluoranthene	19.000	202	19587	0.484	ng/ul#	95
20) Pyrene	19.362	202	20265	0.483	ng/ul#	95
21) Benzo(a)anthracene	21.119	228	17639	0.402	ng/ul	99
22) Chrysene	21.172	228	17898	0.400	ng/ul	99
24) Benzo(b)fluoranthene	22.656	252	15304	0.425	ng/ul	98
25) Benzo(k)fluoranthene	22.700	252	15996	0.424	ng/ul#	99
26) Benzo(a)pyrene	23.206	252	13391	0.416	ng/ul#	98
27) Indeno(1,2,3-cd)pyrene	25.446	276	15589	0.396	ng/ul#	92
28) Dibenzo(a,h)anthracene	25.463	278	12138	0.396	ng/ul	98
29) Benzo(g,h,i)perylene	26.097	276	12184	0.405	ng/ul#	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111524\
Data File : BN035113.D
Acq On : 15 Nov 2024 19:24
Operator : RC/JU
Sample : PB164839BS
Misc :
ALS Vial : 18 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SLCS839

Quant Time: Nov 15 22:27:15 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 13 00:41:17 2024
Response via : Initial Calibration

