

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111524\
 Data File : BN035115.D
 Acq On : 15 Nov 2024 21:12
 Operator : RC/JU
 Sample : PB164841BL
 Misc :
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SBLK841

Manual Integrations
 APPROVED

Reviewed By :Yogesh Patel 11/18/2024
 Supervised By :mohammad ahmed 11/18/2024

Quant Time: Nov 15 22:27:27 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110924.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 13 00:41:17 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.550	152	2891	0.400	ng/u1	0.00
4) Naphthalene-d8	10.312	136	7872	0.400	ng/u1	-0.01
9) Acenaphthene-d10	14.185	164	6471	0.400	ng/u1	0.00
13) Phenanthrene-d10	16.937	188	15034m	0.400	ng/u1	0.00
17) Chrysene-d12	21.134	240	10172	0.400	ng/u1	# 0.00
23) Perylene-d12	23.296	264	8405m	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.140	96	17426	6.335	ng/u1	0.00
6) 2-Methylnaphthalene-d10	11.912	152	4857	0.403	ng/u1	0.00
18) Fluoranthene-d10	18.968	212	13787	0.528	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111524\
Data File : BN035115.D
Acq On : 15 Nov 2024 21:12
Operator : RC/JU
Sample : PB164841BL
Misc :
ALS Vial : 21 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SBLK841

Quant Time: Nov 15 22:27:27 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110924.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Wed Nov 13 00:41:17 2024
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :Yogesh Patel 11/18/2024
Supervised By :mohammad ahmed 11/18/2024

