


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111524\
Data File : BN035096.D
Acq On : 15 Nov 2024 08:58
Operator : RC/JU
Sample : DFTPP241
Misc :
ALS Vial : 1 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
DFTPP241

Integration File: rteint.p

Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN110924.M
Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
Last Update : Wed Nov 13 00:41:17 2024



AutoFind: Scans 2468, 2469, 2470; Background Corrected with Scan 2461

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	20.1	71160	PASS
68	69	0.00	2	1.9	2002	PASS
69	198	0.00	100	29.8	105461	PASS
70	69	0.00	2	0.5	548	PASS
127	198	10	80	37.6	133288	PASS
197	198	0.00	2	0.4	1445	PASS
198	198	100	100	100.0	354155	PASS
199	198	5	9	6.9	24416	PASS
275	198	10	60	29.5	104499	PASS
365	198	1	100	4.7	16572	PASS
441	198	0.01	100	10.4	36808	PASS
442	198	50	100	65.8	233195	PASS
443	442	15	24	18.6	43467	PASS

m/z	Abundance
39.00	228.0
44.00	355.0
72.00	284.0
147.00	167.0
193.00	185.0
207.00	1342.0
208.00	259.0
281.00	295.0

Instrument :

BNA_N

ClientSampleId :

DFTPP241

m/z	Abundance
37.00	572.0
38.00	1368.0
39.00	6109.0
39.90	440.0
41.00	181.0
44.00	297.0
44.90	170.0
49.10	764.0
50.00	25104.0
51.00	85560.0
52.00	4367.0
53.00	192.0
54.90	528.0
56.00	2797.0
57.00	5498.0
57.90	294.0
60.90	1306.0
62.00	1657.0
63.00	4222.0
64.00	720.0
65.00	1946.0
65.90	246.0
67.10	184.0
68.00	2529.0
69.00	122792.0
69.90	737.0
73.00	1629.0
74.00	14768.0
75.00	20840.0
76.00	7278.0
77.00	132608.0
78.00	8947.0
79.00	11022.0
80.00	7674.0
81.00	10771.0
81.90	2594.0
83.00	1953.0
83.90	317.0
85.00	1829.0
86.00	3579.0
87.00	1594.0
88.00	432.0
88.90	332.0
91.00	2718.0
92.00	3111.0
92.90	22488.0
94.00	1355.0
94.90	359.0
95.80	525.0
96.10	510.0
97.10	298.0
97.90	16984.0
98.90	11792.0
100.00	882.0
100.90	6284.0
101.90	430.0
102.90	2114.0
103.90	3922.0
105.00	3952.0
105.90	1382.0
107.00	51504.0
108.00	6925.0
109.00	1428.0
110.00	74440.0
111.00	13151.0
111.90	1578.0
113.10	545.0
114.90	271.0
115.90	3108.0
116.90	53488.0
118.00	3625.0
119.00	491.0
120.00	735.0
120.90	176.0
121.90	3315.0
123.00	5194.0
124.00	2236.0
124.90	2053.0
127.00	147200.0
128.00	11757.0
128.90	64624.0
130.00	5656.0
130.90	1317.0
131.90	654.0
132.90	371.0
134.00	1857.0
135.00	5122.0
135.90	2172.0
136.90	3002.0
138.00	676.0
139.00	338.0
140.00	987.0

Instrument :
BNA_N
ClientSampleId :
DFTPP241

141.00	10264.0
142.00	2838.0
142.90	2474.0
144.00	528.0
144.80	507.0
145.90	1720.0
146.90	5084.0
147.90	13076.0
149.00	2287.0
149.90	638.0
150.90	1164.0
151.80	609.0
152.90	2920.0
154.00	2185.0
154.90	5467.0
156.00	8014.0
156.90	1486.0
158.00	1605.0
158.90	1734.0
159.90	3265.0
160.90	4492.0
161.90	991.0
162.90	376.0
163.80	485.0
164.90	3480.0
166.00	3303.0
167.00	20792.0
168.00	10531.0
169.00	1832.0
170.00	616.0
171.00	800.0
171.90	1905.0
172.90	2520.0
174.00	3973.0
175.00	8034.0
176.00	1761.0
176.90	3636.0
177.90	1395.0
178.90	15768.0
179.90	9826.0
181.00	5184.0
182.00	681.0
182.90	323.0
183.90	1076.0
185.00	7879.0
186.00	56032.0
187.00	16092.0
188.00	1764.0
188.90	3952.0
190.00	623.0
190.90	1908.0
191.90	4682.0
193.00	5904.0
193.90	1389.0
194.90	569.0
196.00	9133.0
196.80	1970.0
197.90	361344.0
198.90	25192.0
199.90	2338.0
201.60	1222.0
202.90	3046.0
203.90	15417.0
205.00	25464.0
206.00	107680.0
207.00	14845.0
208.00	4186.0
209.00	1344.0
210.00	1801.0
210.90	4186.0
212.90	273.0
213.80	252.0
214.90	1267.0
215.90	2355.0
216.90	31896.0
217.90	4142.0
219.00	384.0
221.00	11456.0
221.80	4055.0
222.90	6574.0
224.00	59040.0
225.00	15850.0
226.10	1882.0
226.90	29200.0
227.90	4195.0
229.00	5233.0
229.90	709.0
231.00	2086.0
231.90	385.0
232.90	495.0
233.90	1706.0
234.90	1730.0
235.90	1360.0
237.00	2467.0
238.00	447.0
238.90	1083.0

Instrument :
BNA_N
ClientSampleId :
DFTPP241

240.00	847.0
240.90	1712.0
241.90	3444.0
243.00	3153.0
244.00	44176.0
245.00	5746.0
245.90	10578.0
246.90	2150.0
247.80	416.0
248.90	1447.0
249.90	358.0
250.90	437.0
251.90	465.0
253.00	1064.0
255.00	231040.0
255.90	34080.0
256.90	2683.0
257.90	14826.0
258.90	2862.0
259.90	527.0
261.00	386.0
263.80	376.0
264.90	6449.0
265.80	1007.0
267.00	249.0
269.80	268.0
270.90	614.0
271.90	684.0
272.90	6685.0
274.00	17512.0
275.00	99152.0
276.00	13455.0
276.90	9930.0
277.90	1559.0
278.90	328.0
280.80	435.0
282.00	256.0
283.00	1067.0
284.00	676.0
285.00	1666.0
285.90	313.0
289.00	349.0
289.90	330.0
291.00	316.0
292.00	480.0
292.90	2232.0
294.00	525.0
295.00	636.0
295.90	31488.0
296.90	4640.0
297.90	400.0
301.00	491.0
301.90	636.0
302.90	3210.0
304.00	1126.0
307.80	496.0
308.80	318.0
310.00	325.0
313.00	249.0
313.90	1314.0
314.90	3624.0
316.00	1390.0
317.00	402.0
320.90	1030.0
321.90	582.0
323.00	9753.0
324.00	1641.0
324.90	196.0
325.70	233.0
326.90	2118.0
328.00	849.0
329.00	237.0
331.90	535.0
332.90	1135.0
334.00	5980.0
335.00	1547.0
336.00	163.0
338.80	164.0
341.00	1280.0
342.00	304.0
345.90	2127.0
346.80	540.0
351.00	166.0
352.00	2438.0
352.90	2105.0
353.90	2495.0
355.00	567.0
359.00	226.0
364.90	13568.0
365.90	2071.0
369.90	265.0
371.00	868.0
372.00	4608.0
372.90	1201.0
382.90	1371.0
383.80	364.0

Instrument :
BNA_N
ClientSampleId :
DFTPP241

385.00	207.0
389.90	619.0
391.10	265.0
391.90	214.0
401.00	249.0
402.00	1977.0
403.10	2304.0
403.90	1155.0
404.80	186.0
414.90	181.0
421.00	1935.0
422.00	2002.0
423.00	11530.0
424.00	2977.0
424.90	302.0
441.00	28896.0
442.00	191168.0
443.00	36120.0
444.00	3427.0

Instrument :
BNA_N
ClientSampleId :
DFTPP241

m/z	Abundance
36.90	509.0
38.10	1165.0
39.00	6053.0
39.90	627.0
40.90	190.0
43.90	450.0
44.90	182.0
49.00	699.0
50.00	24304.0
51.00	84784.0
52.00	4975.0
53.00	211.0
55.00	513.0
55.90	2759.0
57.00	6004.0
57.90	301.0
61.00	1737.0
62.00	1516.0
63.00	4673.0
64.00	644.0
65.00	2211.0
66.90	297.0
68.00	2342.0
69.00	124600.0
70.00	645.0
73.00	1920.0
74.00	15571.0
75.00	22312.0
76.00	7602.0
77.00	137472.0
78.00	9323.0
79.00	11475.0
79.90	8838.0
81.00	11019.0
82.00	2719.0
82.90	2216.0
84.00	277.0
84.90	1806.0
86.00	3790.0
86.90	1636.0
88.00	554.0
89.00	340.0
90.90	2697.0
92.00	2670.0
93.00	22384.0
94.00	1277.0
94.90	268.0
96.00	906.0
97.10	366.0
97.90	17984.0
99.00	11777.0
100.00	969.0
101.00	6057.0
102.10	381.0
102.90	2012.0
104.00	4189.0
104.90	4154.0
106.10	1527.0
107.00	52824.0
108.00	7826.0
109.00	1397.0
110.00	79664.0
111.00	13401.0
111.90	1606.0
112.90	648.0
114.90	231.0
116.00	3058.0
116.90	58168.0
118.00	3748.0
118.90	487.0
120.00	782.0
120.80	333.0
121.90	3816.0
122.90	5534.0
123.90	2716.0
124.90	2394.0
127.00	157184.0
128.00	12641.0
129.00	67736.0
130.00	5900.0
130.90	992.0
131.90	675.0
132.80	351.0
134.00	2275.0
134.90	5343.0
136.00	2408.0
137.00	3051.0
137.90	630.0
138.90	334.0
140.00	885.0
141.00	10364.0
141.90	2939.0

Instrument :
BNA_N
ClientSampleId :
DFTPP241

142.90	2142.0
143.90	589.0
145.00	539.0
146.00	1878.0
146.90	5722.0
147.90	13798.0
149.00	2722.0
150.00	715.0
150.90	1220.0
151.90	736.0
153.00	3457.0
153.90	2329.0
155.00	6330.0
156.00	9700.0
157.00	1467.0
157.90	2117.0
158.90	1566.0
160.00	3628.0
160.90	5197.0
161.90	1172.0
163.00	360.0
163.90	582.0
165.00	4541.0
165.90	3107.0
167.00	24400.0
168.00	10970.0
169.00	1928.0
169.90	700.0
170.90	948.0
171.90	2201.0
173.00	2624.0
174.00	4628.0
175.00	8719.0
176.00	2373.0
176.90	4383.0
177.90	1565.0
178.90	17464.0
179.90	11758.0
180.90	5166.0
181.90	741.0
182.70	501.0
184.00	1636.0
185.00	8104.0
186.00	64576.0
187.00	19272.0
188.00	2377.0
188.90	4712.0
189.90	746.0
190.90	2272.0
192.00	5970.0
193.00	6706.0
193.90	1330.0
194.80	599.0
195.90	10500.0
196.90	2365.0
197.90	419328.0
198.90	30208.0
199.90	2182.0
201.30	1448.0
202.90	4271.0
203.90	17784.0
205.00	29680.0
206.00	119208.0
207.00	17808.0
207.90	5208.0
209.00	1692.0
210.10	2120.0
210.90	4656.0
212.80	524.0
214.90	1839.0
216.00	2918.0
216.90	35688.0
217.90	4882.0
219.00	422.0
219.90	688.0
221.00	13482.0
221.80	4689.0
222.90	7242.0
224.00	69832.0
225.00	18456.0
226.00	1813.0
226.90	36200.0
227.90	4462.0
229.00	6513.0
229.90	930.0
230.90	2415.0
232.00	372.0
232.80	566.0
233.90	1996.0
235.00	2182.0
235.90	1899.0
236.90	2583.0
238.00	354.0
238.90	1348.0
239.90	1202.0
241.00	1635.0

Instrument :
BNA_N
ClientSampleId :
DFTPP241

241.90	4106.0
243.00	3812.0
244.00	52784.0
245.00	7495.0
246.00	13148.0
247.00	2527.0
248.00	678.0
248.90	1881.0
249.80	469.0
250.90	660.0
251.90	521.0
252.90	1305.0
255.00	290496.0
256.00	40264.0
257.00	3679.0
257.90	19296.0
259.00	2941.0
260.00	439.0
260.80	393.0
262.70	207.0
263.90	423.0
265.00	8090.0
265.90	1141.0
266.90	239.0
269.90	296.0
270.90	835.0
272.00	749.0
273.00	8307.0
274.00	22832.0
275.00	126440.0
276.00	15565.0
276.90	13174.0
278.00	2294.0
279.00	437.0
280.90	441.0
281.90	376.0
282.90	1401.0
284.00	981.0
285.00	2496.0
285.90	434.0
288.90	390.0
290.00	567.0
291.00	322.0
291.90	489.0
292.90	3062.0
293.90	825.0
296.00	42304.0
296.90	6068.0
297.90	283.0
300.90	576.0
301.90	645.0
302.90	4574.0
304.00	1212.0
304.90	181.0
308.00	528.0
308.90	324.0
310.00	534.0
310.80	156.0
313.00	326.0
314.00	2013.0
315.00	4914.0
316.00	2403.0
316.80	341.0
320.90	1259.0
321.90	602.0
323.00	12350.0
324.00	2303.0
324.90	301.0
325.70	309.0
326.90	2793.0
327.90	1239.0
328.90	200.0
331.90	762.0
333.00	1373.0
334.00	8619.0
334.90	2419.0
335.90	258.0
339.80	174.0
341.00	1393.0
341.90	341.0
345.90	3148.0
346.90	521.0
350.90	260.0
352.00	3998.0
353.00	2568.0
354.00	3816.0
355.00	753.0
358.90	264.0
364.90	19992.0
365.90	3013.0
366.80	391.0
370.00	476.0
371.00	888.0
372.00	6840.0
372.90	1682.0
373.90	218.0

Instrument :
BNA_N
ClientSampleId :
DFTPP241

376.80	190.0
382.90	1781.0
384.00	465.0
384.80	195.0
390.10	768.0
390.90	562.0
392.00	447.0
400.90	351.0
402.00	2898.0
403.00	3169.0
404.00	1361.0
404.90	240.0
421.00	2549.0
421.90	2644.0
423.00	17056.0
424.00	3730.0
424.90	442.0
441.00	46448.0
442.00	267200.0
443.00	53208.0
443.90	4416.0
444.70	207.0

Instrument :
BNA_N
ClientSampleId :
DFTPP241

m/z	Abundance
36.90	222.0
38.00	618.0
39.10	3356.0
40.00	329.0
40.90	190.0
43.90	317.0
49.00	489.0
50.00	13107.0
51.00	43136.0
52.00	2052.0
55.00	291.0
55.90	1242.0
57.00	2995.0
57.90	166.0
61.00	724.0
62.10	872.0
63.00	2617.0
63.90	347.0
65.00	1119.0
67.90	1136.0
68.90	68992.0
70.00	262.0
73.00	1330.0
74.00	8148.0
75.00	11374.0
76.00	4357.0
77.00	75904.0
78.10	5342.0
79.00	6334.0
79.90	4600.0
81.00	6378.0
81.90	1595.0
82.90	1208.0
84.00	177.0
85.00	1099.0
85.90	2086.0
86.90	1072.0
88.00	293.0
88.80	191.0
90.90	1667.0
91.90	1766.0
93.00	12705.0
93.90	774.0
95.00	208.0
96.00	498.0
97.00	275.0
97.90	9244.0
99.00	7128.0
99.90	725.0
100.90	3613.0
102.90	1384.0
103.90	2573.0
105.00	2286.0
106.00	797.0
107.00	29744.0
108.00	4154.0
109.00	953.0
109.90	46472.0
111.00	7987.0
111.80	815.0
112.80	302.0
115.10	184.0
115.90	1713.0
116.90	34808.0
118.00	2149.0
118.90	279.0
120.00	392.0
121.00	192.0
121.90	2257.0
122.90	3386.0
124.00	1798.0
124.90	1302.0
127.00	95480.0
128.00	6252.0
128.90	41248.0
129.90	3063.0
131.00	671.0
132.00	384.0
132.90	273.0
133.90	1393.0
134.90	3366.0
136.00	1540.0
136.90	1997.0
137.90	390.0
139.00	266.0
140.00	721.0
140.90	6124.0
142.00	1990.0
142.90	1462.0
143.90	382.0
144.90	387.0
145.90	1140.0

Instrument :

BNA_N

ClientSampleId :

DFTPP241

146.90	3579.0
147.90	8444.0
148.90	1515.0
150.00	434.0
151.00	833.0
151.80	444.0
152.90	1805.0
154.00	1681.0
155.00	3918.0
156.00	5439.0
157.00	910.0
157.80	1257.0
158.90	1011.0
159.90	2150.0
160.90	3330.0
161.90	1018.0
162.90	341.0
164.10	282.0
164.90	2992.0
166.00	1857.0
167.00	13908.0
168.00	7053.0
169.00	1073.0
169.90	524.0
170.90	627.0
172.00	1598.0
172.90	2031.0
174.00	2815.0
175.00	5667.0
175.90	1563.0
176.90	2675.0
178.00	909.0
178.90	11301.0
179.90	7035.0
180.90	3224.0
182.00	557.0
182.80	393.0
183.90	1002.0
185.00	5539.0
186.00	40272.0
187.00	11790.0
188.00	1317.0
188.90	3089.0
190.00	564.0
190.90	1478.0
191.90	4036.0
193.00	4165.0
193.90	929.0
194.80	439.0
196.00	6740.0
197.90	281792.0
198.90	17848.0
199.90	1536.0
201.30	1148.0
202.90	2190.0
204.00	11349.0
205.00	18880.0
206.00	82656.0
207.00	11239.0
208.00	3300.0
208.90	1186.0
210.00	1619.0
210.90	3018.0
211.90	509.0
212.80	375.0
215.00	1265.0
216.00	2023.0
216.90	22728.0
217.90	2966.0
218.90	288.0
221.00	8372.0
221.80	3205.0
222.90	5026.0
224.00	46624.0
225.00	11655.0
226.10	1280.0
226.90	22040.0
227.90	3361.0
228.90	4181.0
229.90	547.0
230.90	1802.0
231.80	318.0
233.00	452.0
233.90	1328.0
234.90	1529.0
235.90	1150.0
237.00	1872.0
238.00	268.0
238.90	904.0
239.80	831.0
241.00	1095.0
242.00	2654.0
243.00	2632.0
244.00	36568.0
245.00	4395.0
245.90	9323.0

Instrument :
BNA_N
ClientSampleId :
DFTPP241

247.00	1623.0
247.90	412.0
248.80	1295.0
249.90	236.0
250.90	426.0
252.00	437.0
253.00	927.0
254.00	1504.0
254.90	194304.0
256.00	28056.0
257.00	2344.0
257.90	13413.0
259.00	1858.0
259.90	323.0
260.90	311.0
263.90	437.0
264.90	5254.0
266.00	760.0
266.80	182.0
269.90	331.0
270.90	502.0
271.90	535.0
272.90	5899.0
274.00	15482.0
274.90	87904.0
276.00	11236.0
276.90	8898.0
277.90	1540.0
279.00	308.0
281.00	484.0
281.90	393.0
282.90	990.0
284.00	846.0
284.90	1625.0
286.00	266.0
287.90	160.0
288.80	313.0
289.90	307.0
292.00	495.0
292.90	2033.0
293.90	634.0
295.00	647.0
295.90	31208.0
297.00	4085.0
298.00	344.0
301.10	367.0
302.00	372.0
302.90	3154.0
304.00	727.0
304.90	169.0
307.90	415.0
308.90	297.0
310.00	369.0
311.80	152.0
313.00	290.0
313.90	1273.0
314.90	3123.0
316.00	1715.0
316.90	463.0
319.90	156.0
321.00	740.0
321.90	426.0
323.00	9878.0
323.90	1264.0
324.80	177.0
326.00	159.0
326.90	1895.0
327.90	920.0
328.90	261.0
331.80	832.0
332.90	1031.0
334.00	6390.0
335.00	1416.0
335.90	180.0
338.90	216.0
340.90	1147.0
341.90	323.0
345.90	2260.0
346.90	390.0
352.00	2736.0
353.00	1978.0
353.90	2930.0
354.90	478.0
358.90	223.0
364.90	16157.0
366.00	1825.0
367.00	169.0
370.10	261.0
371.00	837.0
372.00	4650.0
373.00	1290.0
382.90	1430.0
383.90	422.0
384.90	170.0
389.90	599.0
390.90	336.0

Instrument :
BNA_N
ClientSampleId :
DFTPP241

392.00	292.0
401.00	425.0
402.00	1858.0
403.00	3329.0
404.00	966.0
421.00	2095.0
422.00	2065.0
423.00	14685.0
424.00	3014.0
425.00	319.0
441.00	35080.0
442.00	241216.0
443.00	41072.0
444.00	3581.0
445.00	200.0

Instrument :
BNA_N
ClientSampleId :
DFTPP241