





#70
Phenanthrene
Concen: 8.505 ng/ul
RT: 16.90 min Scan# 2450
Delta R.T. 0.00 min
Lab File: BN012617.D
Acq: 19 Nov 2020 05:45

Instrument :
BNA_N
ClientSampleId :

Tot Ion:178 Resp: 412639
Ion Ratio Lower Upper
178 100
179 15.6 13.1 19.7
176 19.6 15.4 23.0



#75
Carbazole
Concen: 44.209 ng/ul
RT: 17.27 min Scan# 2513
Delta R.T. 0.00 min
Lab File: BN012617.D
Acq: 19 Nov 2020 05:45

Tgt Ion:167 Resp: 1855744
Ion Ratio Lower Upper
167 100
166 21.6 17.8 26.8
139 13.5 10.4 15.6

