

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN111820\
 Data File : BN012651.D
 Acq On : 20 Nov 2020 02:45
 Operator : CG/JU
 Sample : SSTDC0020
 Misc :
 ALS Vial : 60 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD02008

Manual Integrations
 APPROVED

mohammad
 11/20/2020 11:31:29 AM

Quant Time: Nov 20 09:40:25 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN102220.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 18 14:31:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) 1,4-Dichlorobenzene-d4	7.44	152	119332	20.00	ng/ul	0.00
18) Naphthalene-d8	10.21	136	520528	20.00	ng/ul	0.00
36) Acenaphthene-d10	14.10	164	357011	20.00	ng/ul	0.00
62) Phenanthrene-d10	16.85	188	778712	20.00	ng/ul	0.00
78) Chrysene-d12	21.06	240	872604	20.00	ng/ul	-0.01
86) Perylene-d12	23.18	264	927764	20.00	ng/ul	-0.02

System Monitoring Compounds

3) 1,4-Dioxane-d8	2.98	96	24779	7.40	ng/uL	0.00
5) Phenol-d5	6.63	99	199695	18.37	ng/ul	0.00
7) Bis-(2-Chloroethyl)ether-d	6.79	67	130191	20.36	ng/ul	0.00
9) 2-Chlorophenol-d4	6.98	132	166396	19.60	ng/ul	-0.01
13) 4-Methylphenol-d8	8.15	113	161789	18.24	ng/ul	-0.01
19) Nitrobenzene-d5	8.59	128	79259	18.56	ng/ul	0.00
22) 2-Nitrophenol-d4	9.30	143	89592	18.88	ng/ul	0.00
26) 2,4-Dichlorophenol-d3	9.83	165	168768	18.72	ng/ul	-0.01
29) 4-Chloroaniline-d4	10.36	131	200348	22.48	ng/ul	0.00
44) Dimethylphthalate-d6	13.52	166	551780	18.92	ng/ul	0.00
47) Acenaphthylene-d8	13.79	160	658812	18.93	ng/ul	0.00
52) 4-Nitrophenol-d4	14.32	143	92055	16.86	ng/ul	0.00
58) Fluorene-d10	15.10	176	491057	19.14	ng/ul	0.00
63) 4,6-Dinitro-2-methylphenol	15.23	200	92241	17.36	ng/ul	0.00
71) Anthracene-d10	16.95	188	759006	19.19	ng/ul	0.00
79) Pyrene-d10	19.26	212	830303	18.19	ng/ul	0.00
90) Benzo(a)pyrene-d12	23.04	264	961391	18.96	ng/ul	-0.02

Target Compounds

					Ovalue
2) 1,4-Dioxane	3.01	88	25801	7.125	ng/uL 88
4) Benzaldehyde	6.59	77	119894	19.145	ng/ul 96
6) Phenol	6.65	94	208234	18.753	ng/ul 94
8) Bis(2-Chloroethyl)ether	6.88	93	165632	19.495	ng/ul 96
10) 2-Chlorophenol	7.00	128	164703	19.135	ng/ul 97
11) 2-Methylphenol	7.89	108	154409	18.609	ng/ul 99
12) 2,2'-oxybis(1-Chloropropan	7.97	45	239170	21.356	ng/ul 98
14) Acetophenone	8.26	105	264868	19.441	ng/ul 97
15) N-Nitroso-di-n-propylamine	8.25	70	145540	21.091	ng/ul 95
16) 4-Methylphenol	8.22	108	164715	18.168	ng/ul 92
17) Hexachloroethane	8.50	117	80184	21.238	ng/ul 97
20) Nitrobenzene	8.63	77	227115	19.510	ng/ul 98
21) Isophorone	9.16	82	394460	20.051	ng/ul 98
23) 2-Nitrophenol	9.33	139	95682	18.713	ng/ul 93
24) 2,4-Dimethylphenol	9.40	107	210515	19.310	ng/ul 97
25) Bis(2-Chloroethoxy)methane	9.65	93	231182	18.989	ng/ul 97
27) 2,4-Dichlorophenol	9.86	162	168092	19.119	ng/ul 100
28) Naphthalene	10.26	128	572029	19.010	ng/ul 99
30) 4-Chloroaniline	10.38	127	197033	22.975	ng/ul 99
31) Hexachlorobutadiene	10.54	225	116541	20.211	ng/ul 96
32) Caprolactam	11.15	113	49196m	17.612	ng/ul
33) 4-Chloro-3-methylphenol	11.51	107	189213	18.792	ng/ul 97
34) 2-Methylnaphthalene	11.88	142	396313	18.871	ng/ul 100

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN111820\
 Data File : BN012651.D
 Acq On : 20 Nov 2020 02:45
 Operator : CG/JU
 Sample : SSTDC0020
 Misc :
 ALS Vial : 60 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD02008

Manual Integrations
 APPROVED

mohammad
 11/20/2020 11:31:29 AM

Quant Time: Nov 20 09:40:25 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN102220.M
 Quant Title : SVOA CALIBRATION
 QLast Update : Wed Nov 18 14:31:59 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
35) 1-Methylnaphthalene	12.10	142	486686	19.446	ng/uL	98
37) 1,2,4,5-Tetrachlorobenzene	12.26	216	209133	18.971	ng/uL	97
38) Hexachlorocyclopentadiene	12.23	237	126760	18.329	ng/uL	93
39) 2,4,6-Trichlorophenol	12.51	196	134489	18.420	ng/uL	95
40) 2,4,5-Trichlorophenol	12.58	196	145541	18.923	ng/uL	89
41) 1,1'-Biphenyl	12.92	154	540513	18.267	ng/uL#	95
42) 2-Chloronaphthalene	12.96	162	426999	18.471	ng/uL	96
43) 2-Nitroaniline	13.17	65	133885	19.460	ng/uL	95
45) Dimethylphthalate	13.57	163	569531	19.392	ng/uL	98
46) 2,6-Dinitrotoluene	13.69	165	107252	18.126	ng/uL#	88
48) Acenaphthylene	13.82	152	650595	18.666	ng/uL	99
49) 3-Nitroaniline	14.02	138	84728	18.497	ng/uL	95
50) Acenaphthene	14.16	153	477718	18.767	ng/uL	98
51) 2,4-Dinitrophenol	14.23	184	52865	14.439	ng/uL#	84
53) 4-Nitrophenol	14.33	109	95576	19.156	ng/uL	91
54) Dibenzofuran	14.50	168	671336	19.024	ng/uL	99
55) 2,4-Dinitrotoluene	14.48	165	168634	19.539	ng/uL	89
56) 2,3,4,6-Tetrachlorophenol	14.73	232	128008	19.191	ng/uL	98
57) Diethylphthalate	14.95	149	595242	19.830	ng/uL	99
59) Fluorene	15.16	166	558493	18.941	ng/uL	96
60) 4-Chlorophenyl-phenylether	15.16	204	281694	19.681	ng/uL	96
61) 4-Nitroaniline	15.19	138	96846	16.670	ng/uL	97
64) 4,6-Dinitro-2-methylphenol	15.25	198	97964	17.377	ng/uL#	90
65) N-Nitrosodiphenylamine	15.37	169	452689	18.284	ng/uL	99
66) 4-Bromophenyl-phenylether	16.05	248	168428	19.255	ng/uL	94
67) Hexachlorobenzene	16.16	284	192898	18.502	ng/uL	99
68) Atrazine	16.34	200	172059	19.324	ng/uL	96
69) Pentachlorophenol	16.50	266	110507	17.970	ng/uL	97
70) Phenanthrene	16.89	178	920142	19.219	ng/uL	97
72) Anthracene	16.99	178	929184	19.144	ng/uL	98
73) 1,2,3,4-Tetrachlorobenzene	12.88	216	203769	18.430	ng/uL	98
74) Pentachlorobenzene	14.42	250	215734	18.689	ng/uL	93
75) Carbazole	17.26	167	787618	19.014	ng/uL	98
76) Di-n-butylphthalate	17.84	149	1064097	20.766	ng/uL	99
77) Fluoranthene	18.92	202	1054188	19.145	ng/uL	95
80) Pvrene	19.29	202	1118523	18.004	ng/uL	99
81) Butylbenzylphthalate	20.22	149	472999	18.814	ng/uL	95
82) 3,3'-Dichlorobenzidine	20.99	252	317375	18.316	ng/uL	90
83) Benzo(a)anthracene	21.05	228	1097078	18.654	ng/uL	99
84) Bis(2-ethylhexyl)phthalate	21.00	149	731756	19.312	ng/uL	99
85) Chrysene	21.10	228	1071678	18.693	ng/uL	98
87) Di-n-octyl phthalate	21.85	149	1263705	19.124	ng/uL	100
88) Benzo(b)fluoranthene	22.56	252	1166484m	18.816	ng/uL	
89) Benzo(k)fluoranthene	22.60	252	1147793	19.251	ng/uL	99
91) Benzo(a)pyrene	23.09	252	1058905	18.698	ng/uL	95
92) Indeno(1,2,3-cd)pyrene	25.27	276	1335651	18.924	ng/uL	97
93) Dibenzo(a,h)anthracene	25.27	278	1127404	18.654	ng/uL	96
94) Benzo(a,h,i)perylene	25.90	276	1116222	18.677	ng/uL	99

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN111820\
Data File : BN012651.D
Acq On : 20 Nov 2020 02:45
Operator : CG/JU
Sample : SSTDCCC020
Misc :
ALS Vial : 60 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD02008


Manual Integrations
APPROVED

mohammad
11/20/2020 11:31:29 AM

Quant Time: Nov 20 09:40:25 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN102220.M
Quant Title : SVOA CALIBRATION
QLast Update : Wed Nov 18 14:31:59 2020
Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
--------------------	------	------	----------	------	-------	----------

(#) = qualifier out of range (m) = manual integration (+) = signals summed



#2

1,4-Dioxane

Concen: 7.125 ng/uL

RT: 3.01 min Scan# 89

Delta R.T. -0.01 min

Lab File: BN012651.D

Acq: 20 Nov 2020 02:45

Instrument :

BNA_N





ClientSampleId :

SSTD02008

Tot Ion:	88	Resp:	25801
Ion Ratio	Lower	Upper	
88	100		
43	36.5	26.6	39.8
58	63.2	61.0	91.4

Manual Integrations
APPROVED

mohammad
11/20/2020 11:31:29 AM



#4

Benzaldehyde

Concen: 19.145 ng/uL




RT: 6.59 min Scan# 698


Delta R.T. -0.01 min

Lab File: BN012651.D

Acq: 20 Nov 2020 02:45

Tgt Ion:	77	Resp:	119894
Ion Ratio	Lower	Upper	
77	100		
105	86.6	71.0	106.6
106	84.9	63.9	95.9





#6

Phenol





Concen: 18.753 ng/ul
 RT: 6.65 min Scan# 708
 Delta R.T. -0.01 min
 Lab File: BN012651.D
 Acq: 20 Nov 2020 02:45

Instrument :
 BNA_N
 ClientSampleId :
 SST02008

Tot Ion	Ratio	Lower	Upper
94	100		
65	34.9	26.1	39.1
66	49.5	36.1	54.1

Manual Integrations
 APPROVED

mohammad
 11/20/2020 11:31:29 AM







#8

Bis(2-Chloroethyl)ether

Concen: 19.495 ng/ul
 RT: 6.88 min Scan# 747
 Delta R.T. -0.01 min
 Lab File: BN012651.D
 Acq: 20 Nov 2020 02:45

Tgt Ion	Ratio	Lower	Upper
93	100		
63	80.9	60.9	91.3
95	32.2	25.8	38.6





#10

2-Chlorophenol

Concen: 19.135 ng/ul

RT: 7.00 min Scan# 768

Delta R.T. -0.01 min

Lab File: BN012651.D

Acq: 20 Nov 2020 02:45

Instrument :

BNA_N





ClientSampleId :

SSTD02008

Tot Ion:	128	Resp:	164703
Ion Ratio	Lower	Upper	
128	100		
64	52.1	40.0	60.0
130	30.7	26.2	39.2

Manual Integrations
APPROVED

mohammad
11/20/2020 11:31:29 AM



#11

2-Methylphenol

Concen: 18.609 ng/ul




RT: 7.89 min Scan# 918


Delta R.T. -0.01 min

Lab File: BN012651.D

Acq: 20 Nov 2020 02:45

Tgt Ion:	108	Resp:	154409
Ion Ratio	Lower	Upper	
108	100		
107	90.8	71.9	107.9









#12
2,2'-oxvbis(1-Chloropropane)
Concen: 21.356 ng/ul
RT: 7.97 min Scan# 932
Delta R.T. -0.01 min
Lab File: BN012651.D
Acq: 20 Nov 2020 02:45

Instrument :
BNA_N
ClientSampleId :
SSTD02008

Tot Ion	Ratio	Lower	Upper
45	100		
77	15.7	11.8	17.6
79	11.2	9.3	13.9




Manual Integrations
APPROVED




mohammad
11/20/2020 11:31:29 AM



#14
Acetophenone
Concen: 19.441 ng/ul
RT: 8.26 min Scan# 982
Delta R.T. -0.01 min
Lab File: BN012651.D
Acq: 20 Nov 2020 02:45

Tgt Ion	Ratio	Lower	Upper
105	100		
77	85.2	65.7	98.5
51	33.2	25.2	37.8





#15

N-Nitroso-di-n-propylamine

Concen: 21.091 ng/ul

RT: 8.25 min Scan# 980

Delta R.T. -0.01 min

Lab File: BN012651.D

Acq: 20 Nov 2020 02:45

Tot Ion	Ion	Ratio	Lower	Upper
70	100			
42	66.6	50.0	75.0	
101	9.7	7.5	11.3	
130	19.2	17.6	26.4	

Instrument :





BNA_N

ClientSampleId :

SSTD02008

Manual Integrations
APPROVED

mohammad
11/20/2020 11:31:29 AM



#16

4-Methylphenol

Concen: 18.168 ng/ul


RT: 8.22 min Scan# 974


Delta R.T. -0.01 min

Lab File: BN012651.D

Acq: 20 Nov 2020 02:45

Tgt Ion	Ion	Ratio	Lower	Upper
108	100			
107	119.1	88.4	132.6	









#17
Hexachloroethane
Concen: 21.238 ng/ul
RT: 8.50 min Scan# 1022
Delta R.T. -0.01 min
Lab File: BN012651.D
Acq: 20 Nov 2020 02:45

Instrument :
BNA_N
ClientSampleId :
SSTD02008

Tot Ion	Ratio	Lower	Upper
117	100		
201	102.4	79.4	119.2
199	62.0	48.3	72.5




Manual Integrations
APPROVED


mohammad
11/20/2020 11:31:29 AM



#20
Nitrobenzene
Concen: 19.510 ng/ul
RT: 8.63 min Scan# 1045
Delta R.T. -0.01 min
Lab File: BN012651.D
Acq: 20 Nov 2020 02:45

Tgt Ion	Ratio	Lower	Upper
77	100		
123	39.6	33.0	49.6
65	14.7	12.0	18.0





#21

Isophorone

Concen: 20.051 ng/ul

RT: 9.16 min Scan# 1134

Delta R.T. -0.01 min

Lab File: BN012651.D

Acq: 20 Nov 2020 02:45

Instrument :

BNA_N





ClientSampleId :

SSTD02008

Tot Ion:	82	Resp:	394460
Ion Ratio		Lower	Upper
82	100		
95	8.3	6.6	10.0
138	16.1	13.8	20.8

Manual Integrations
APPROVED

mohammad
11/20/2020 11:31:29 AM



#23

2-Nitrophenol

Concen: 18.713 ng/ul




RT: 9.33 min Scan# 1164


Delta R.T. -0.01 min

Lab File: BN012651.D

Acq: 20 Nov 2020 02:45

Tgt Ion:	139	Resp:	95682
Ion Ratio		Lower	Upper
139	100		
65	63.5	44.9	67.3
109	36.2	28.5	42.7









#24
2,4-Dimethylphenol
Concen: 19.310 ng/ul
RT: 9.40 min Scan# 1176
Delta R.T. -0.01 min
Lab File: BN012651.D
Acq: 20 Nov 2020 02:45

Instrument :
BNA_N
ClientSampleId :
SSTD02008

Tot Ion	Ratio	Resp	Lower	Upper
107	100	210515		
121	46.1		39.3	58.9
122	78.5		61.7	92.5




Manual Integrations
APPROVED


mohammad
11/20/2020 11:31:29 AM



#25
Bis(2-Chloroethoxy)methane
Concen: 18.989 ng/ul
RT: 9.65 min Scan# 1217
Delta R.T. -0.01 min
Lab File: BN012651.D
Acq: 20 Nov 2020 02:45

Tgt Ion	Ratio	Resp	Lower	Upper
93	100	231182		
95	32.9		24.9	37.3
123	10.9		8.3	12.5









#27
 2,4-Dichlorophenol
 Concen: 19.119 ng/ul
 RT: 9.86 min Scan# 1254
 Delta R.T. -0.01 min
 Lab File: BN012651.D
 Acq: 20 Nov 2020 02:45

Instrument :
 BNA_N
 ClientSampleID :
 SST02008

Tot Ion	Ratio	Lower	Upper
162	100		
164	66.7	53.0	79.6
98	40.1	32.1	48.1




Manual Integrations
 APPROVED


mohammad
 11/20/2020 11:31:29 AM



#28
 Naphthalene
 Concen: 19.010 ng/ul
 RT: 10.26 min Scan# 1321
 Delta R.T. -0.01 min
 Lab File: BN012651.D
 Acq: 20 Nov 2020 02:45

Tgt Ion	Ratio	Lower	Upper
128	100		
129	10.8	8.6	12.8
127	13.7	11.6	17.4









#30
4-Chloroaniline
Concen: 22.975 ng/ul
RT: 10.38 min Scan# 1341
Delta R.T. -0.01 min
Lab File: BN012651.D
Acq: 20 Nov 2020 02:45

Instrument :
BNA_N
ClientSampleId :
SSTD02008

Tot Ion:127 Resp: 197033
Ion Ratio Lower Upper
127 100
129 31.2 25.5 38.3




Manual Integrations
APPROVED


mohammad
11/20/2020 11:31:29 AM



#31
Hexachlorobutadiene
Concen: 20.211 ng/ul
RT: 10.54 min Scan# 1369
Delta R.T. -0.01 min
Lab File: BN012651.D
Acq: 20 Nov 2020 02:45

Tgt Ion:225 Resp: 116541
Ion Ratio Lower Upper
225 100
223 64.0 49.1 73.7
227 61.4 46.0 69.0





#32

Caprolactam

Concen: 17.612 ng/ul m

RT: 11.15 min Scan# 1473

Delta R.T. -0.01 min

Lab File: BN012651.D

Acq: 20 Nov 2020 02:45

Instrument :

BNA_N





ClientSampleId :

SSTD02008

Tot Ion:	113	Resp:	49196
Ion	Ratio	Lower	Upper
113	100		
55	171.1	124.0	186.0
56	145.0	97.0	145.4

Manual Integrations
APPROVED

mohammad
11/20/2020 11:31:29 AM



#33

4-Chloro-3-methylphenol

Concen: 18.792 ng/ul




RT: 11.51 min Scan# 1534


Delta R.T. -0.01 min

Lab File: BN012651.D

Acq: 20 Nov 2020 02:45

Tgt Ion:	107	Resp:	189213
Ion	Ratio	Lower	Upper
107	100		
144	24.7	20.0	30.0
142	71.6	60.2	90.4









#34
2-Methylnaphthalene
Concen: 18.871 ng/uL
RT: 11.88 min Scan# 1597
Delta R.T. -0.01 min
Lab File: BN012651.D
Acq: 20 Nov 2020 02:45

Instrument :
BNA_N
ClientSampleId :
SSTD02008

Tot Ion	Ratio	Lower	Upper
142	100		
141	92.7	73.8	110.8




Manual Integrations
APPROVED


mohammad
11/20/2020 11:31:29 AM



#35
1-Methylnaphthalene
Concen: 19.446 ng/uL
RT: 12.10 min Scan# 1635
Delta R.T. -0.01 min
Lab File: BN012651.D
Acq: 20 Nov 2020 02:45

Tgt Ion	Ratio	Lower	Upper
142	100		
141	91.6	74.7	112.1
116	4.6	3.4	5.0





#37

1,2,4,5-Tetrachlorobenzene

Concen: 18.971 ng/ul

RT: 12.26 min Scan# 1661

Delta R.T. -0.01 min

Lab File: BN012651.D

Acq: 20 Nov 2020 02:45

Instrument :

BNA_N





ClientSampleId :

SSTD02008

Tot Ion	Ratio	Resp	Lower	Upper
216	100	209133		
214	81.9	63.4	95.0	
179	22.5	17.4	26.2	
108	20.8	14.8	22.2	

Manual Integrations
APPROVED

mohammad
11/20/2020 11:31:29 AM



#38

Hexachlorocyclopentadiene

Concen: 18.329 ng/ul




RT: 12.23 min Scan# 1657


Delta R.T. -0.01 min

Lab File: BN012651.D

Acq: 20 Nov 2020 02:45

Tgt Ion	Ratio	Resp	Lower	Upper
237	100	126760		
235	57.3	50.7	76.1	
272	12.8	10.2	15.4	









#39
 2,4,6-Trichlorophenol
 Concen: 18.420 ng/ul
 RT: 12.51 min Scan# 1704
 Delta R.T. -0.01 min
 Lab File: BN012651.D
 Acq: 20 Nov 2020 02:45

Instrument :
 BNA_N
 ClientSampled :
 SSTD02008

Tot Ion	Ratio	Resp	Lower	Upper
196	100	134489		
198	90.9	78.2	117.2	
200	31.9	25.4	38.2	




Manual Integrations
 APPROVED


mohammad
 11/20/2020 11:31:29 AM



#40
 2,4,5-Trichlorophenol
 Concen: 18.923 ng/ul
 RT: 12.58 min Scan# 1716
 Delta R.T. -0.01 min
 Lab File: BN012651.D
 Acq: 20 Nov 2020 02:45

Tgt Ion	Ratio	Resp	Lower	Upper
196	100	145541		
198	90.1	83.8	125.6	
200	29.7	23.8	35.6	









#41
1,1'-Biphenyl
Concen: 18.267 ng/ul
RT: 12.92 min Scan# 1774
Delta R.T. -0.01 min
Lab File: BN012651.D
Acq: 20 Nov 2020 02:45

Instrument :
BNA_N
ClientSampleId :
SSTD02008

Tot Ion:	154	Resp:	540513
Ion Ratio	Lower	Upper	
154	100		
153	42.3	31.4	47.0
76	13.9	9.3	13.9#




Manual Integrations
APPROVED


mohammad
11/20/2020 11:31:29 AM



#42
2-Chloronaphthalene
Concen: 18.471 ng/ul
RT: 12.96 min Scan# 1780
Delta R.T. -0.01 min
Lab File: BN012651.D
Acq: 20 Nov 2020 02:45

Tgt Ion:	162	Resp:	426999
Ion Ratio	Lower	Upper	
162	100		
164	33.8	23.5	35.3
127	40.8	32.9	49.3









#43
2-Nitroaniline
Concen: 19.460 ng/ul
RT: 13.17 min Scan# 1817
Delta R.T. -0.01 min
Lab File: BN012651.D
Acq: 20 Nov 2020 02:45

Instrument :
BNA_N
ClientSampleId :
SSTD02008

Tot Ion	Ratio	Lower	Upper
65	100		
92	68.4	53.0	79.6
138	91.4	78.6	118.0




Manual Integrations
APPROVED


mohammad
11/20/2020 11:31:29 AM



#45
Dimethylphthalate
Concen: 19.392 ng/ul
RT: 13.57 min Scan# 1884
Delta R.T. -0.01 min
Lab File: BN012651.D
Acq: 20 Nov 2020 02:45

Tgt Ion	Ratio	Lower	Upper
163	100		
194	3.7	3.0	4.4
164	9.8	8.6	13.0









#46
 2,6-Dinitrotoluene
 Concen: 18.126 ng/ul
 RT: 13.69 min Scan# 1904
 Delta R.T. -0.01 min
 Lab File: BN012651.D
 Acq: 20 Nov 2020 02:45

Instrument :
 BNA_N
 ClientSampleId :
 SSTD02008

Tot Ion	Ratio	Lower	Upper
165	100		
89	73.8	48.7	73.1#
121	22.1	17.8	26.8




Manual Integrations
 APPROVED


mohammad
 11/20/2020 11:31:29 AM



#48
 Acenaphthylene
 Concen: 18.666 ng/ul
 RT: 13.82 min Scan# 1926
 Delta R.T. -0.01 min
 Lab File: BN012651.D
 Acq: 20 Nov 2020 02:45

Tgt Ion	Ratio	Lower	Upper
152	100		
151	20.5	15.5	23.3
153	13.2	10.5	15.7









#49
3-Nitroaniline
Concen: 18.497 ng/ul
RT: 14.02 min Scan# 1960
Delta R.T. -0.01 min
Lab File: BN012651.D
Acq: 20 Nov 2020 02:45

Instrument :
BNA_N
ClientSampleId :
SSTD02008

Tot Ion	Ratio	Lower	Upper
138	100		
108	12.6	9.1	13.7
92	130.5	99.4	149.2




Manual Integrations
APPROVED


mohammad
11/20/2020 11:31:29 AM



#50
Acenaphthene
Concen: 18.767 ng/ul
RT: 14.16 min Scan# 1985
Delta R.T. -0.01 min
Lab File: BN012651.D
Acq: 20 Nov 2020 02:45

Tgt Ion	Ratio	Lower	Upper
153	100		
152	48.2	37.1	55.7
154	88.8	69.7	104.5









#51
2,4-Dinitrophenol
Concen: 14.439 ng/ul
RT: 14.23 min Scan# 1997
Delta R.T. 0.00 min
Lab File: BN012651.D
Acq: 20 Nov 2020 02:45

Instrument :
BNA_N
ClientSampleId :
SSTD02008

Tot Ion	Ratio	Resp	Lower	Upper
184	100			
63	81.6	56.6	84.8	
154	76.7	49.8	74.8	




Manual Integrations
APPROVED


mohammad
11/20/2020 11:31:29 AM



#53
4-Nitrophenol
Concen: 19.156 ng/ul
RT: 14.33 min Scan# 2014
Delta R.T. -0.01 min
Lab File: BN012651.D
Acq: 20 Nov 2020 02:45

Tgt Ion	Ratio	Resp	Lower	Upper
109	100			
139	96.5	93.4	140.0	
65	129.3	103.7	155.5	





#54

Dibenzenofuran

Concen: 19.024 ng/ul

RT: 14.50 min Scan# 2043

Delta R.T. -0.01 min

Lab File: BN012651.D

Acq: 20 Nov 2020 02:45

Instrument :

BNA_N

ClientSampled :

SSTD02008

Tot Ion:168 Resp: 671336

Ion Ratio Lower Upper

168 100


139 37.3 30.5 45.7

Manual Integrations

APPROVED

mohammad

11/20/2020 11:31:29 AM



Abundance Ion 168.00 (167.70 to 168.70): E

Ion 139.00 (138.70 to 139.70): E

14.50

500000

400000



300000

200000

100000

0

Time--> 14.40 14.50 14.60



#55

2,4-Dinitrotoluene

Concen: 19.539 ng/ul

RT: 14.48 min Scan# 2039

Delta R.T. -0.01 min

Lab File: BN012651.D

Acq: 20 Nov 2020 02:45


Tgt Ion:165 Resp: 168634

Ion Ratio Lower Upper

165 100

63 57.6 39.4 59.0

182 3.4 2.8 4.2



Abundance Ion 165.00 (164.70 to 165.70): E

Ion 63.00 (62.70 to 63.70): BNO

Ion 182.00 (181.70 to 182.70): E

14.48


150000


100000

50000

0

Time--> 14.45 14.50 14.55





#56

2,3,4,6-Tetrachlorophenol

Concen: 19.191 ng/ul

RT: 14.73 min Scan# 2082

Delta R.T. -0.01 min

Lab File: BN012651.D

Acq: 20 Nov 2020 02:45

Instrument :

BNA_N

ClientSampleId :

SSTD02008

Tot Ion:232 Resp: 128008

Ion Ratio Lower Upper

232 100

131 52.3 42.9 64.3


130 3.2 2.5 3.7

166 35.2 27.1 40.7

Manual Integrations
APPROVED

mohammad

11/20/2020 11:31:29 AM






Abundance Ion 232.00 (231.70 to 232.70): E

Ion 131.00 (130.70 to 131.70): E

Ion 130.00 (129.70 to 130.70): E

Ion 166.00 (165.70 to 166.70): E



#57

Diethylphthalate

Concen: 19.830 ng/ul

RT: 14.95 min Scan# 2118

Delta R.T. -0.01 min

Lab File: BN012651.D

Acq: 20 Nov 2020 02:45


Tgt Ion:149 Resp: 595242

Ion Ratio Lower Upper

149 100

177 20.9 17.3 25.9



150 13.1 10.2 15.4




Abundance Ion 149.00 (148.70 to 149.70): E

Ion 177.00 (176.70 to 177.70): E

Ion 150.00 (149.70 to 150.70): E





#59

Fluorene

Concen: 18.941 ng/ul

RT: 15.16 min Scan# 2154

Delta R.T. -0.01 min

Lab File: BN012651.D

Acq: 20 Nov 2020 02:45

Instrument :

BNA_N





ClientSampleID :

SSTD02008

Tot Ion:	166	Resp:	558493
Ion Ratio	Lower	Upper	
166	100		
165	99.6	83.3	124.9
167	14.3	11.0	16.6

Manual Integrations
APPROVED

mohammad
11/20/2020 11:31:29 AM



#60

4-Chlorophenyl-phenylether

Concen: 19.681 ng/ul




RT: 15.16 min Scan# 2154


Delta R.T. -0.01 min

Lab File: BN012651.D

Acq: 20 Nov 2020 02:45

Tgt Ion:	204	Resp:	281694
Ion Ratio	Lower	Upper	
204	100		
206	35.1	27.7	41.5
141	64.8	55.8	83.8









#61
4-Nitroaniline
Concen: 16.670 ng/ul
RT: 15.19 min Scan# 2159
Delta R.T. -0.01 min
Lab File: BN012651.D
Acq: 20 Nov 2020 02:45

Instrument :
BNA_N
ClientSampleId :
SSTD02008

Tot Ion	Ratio	Lower	Upper
138	100		
92	58.5	47.0	70.4
108	104.1	79.8	119.6




Manual Integrations
APPROVED


mohammad
11/20/2020 11:31:29 AM



#64
4,6-Dinitro-2-methylphenol
Concen: 17.377 ng/ul
RT: 15.25 min Scan# 2169
Delta R.T. -0.01 min
Lab File: BN012651.D
Acq: 20 Nov 2020 02:45

Tgt Ion	Ratio	Lower	Upper
198	100		
182	5.0	3.8	5.6
77	33.5	22.0	33.0#





#65





N-Nitrosodiphenylamine
Concen: 18.284 ng/ul
RT: 15.37 min Scan# 2191
Delta R.T. -0.01 min
Lab File: BN012651.D
Acq: 20 Nov 2020 02:45

Instrument :
BNA_N
ClientSampleId :
SSTD02008

Tot Ion	Ratio	Lower	Upper
169	100		
168	69.1	56.0	84.0
167	37.0	29.8	44.6

Manual Integrations
APPROVED




mohammad
11/20/2020 11:31:29 AM




#66

4-Bromophenyl-phenylether
Concen: 19.255 ng/ul
RT: 16.05 min Scan# 2306
Delta R.T. -0.01 min
Lab File: BN012651.D
Acq: 20 Nov 2020 02:45

Tgt Ion	Ratio	Lower	Upper
248	100		
250	89.5	75.1	112.7
141	84.1	61.2	91.8





#67

Hexachlorobenzene

Concen: 18.502 ng/ul

RT: 16.16 min Scan# 2324

Delta R.T. -0.01 min

Lab File: BN012651.D

Acq: 20 Nov 2020 02:45

Instrument :

BNA_N


ClientSampleId :

SSTD02008




Tot Ion:	284	Resp:	192898
Ion Ratio	Lower	Upper	
284	100		
142	39.3	31.0	46.6
249	31.6	25.1	37.7

Manual Integrations
APPROVED

mohammad
11/20/2020 11:31:29 AM



Abundance Ion 284.00 (283.70 to 284.70): E
Ion 142.00 (141.70 to 142.70): E
Ion 249.00 (248.70 to 249.70): E



#68

Atrazine

Concen: 19.324 ng/ul


RT: 16.34 min Scan# 2355

Delta R.T. -0.01 min



Lab File: BN012651.D


Acq: 20 Nov 2020 02:45

Tgt Ion:	200	Resp:	172059
Ion Ratio	Lower	Upper	
200	100		
173	27.6	23.4	35.0
215	47.8	35.8	53.6



Abundance Ion 200.00 (199.70 to 200.70): E
Ion 173.00 (172.70 to 173.70): E
Ion 215.00 (214.70 to 215.70): E









#69
 Pentachlorophenol
 Concen: 17.970 ng/ul
 RT: 16.50 min Scan# 2383
 Delta R.T. -0.01 min
 Lab File: BN012651.D
 Acq: 20 Nov 2020 02:45

Instrument :
 BNA_N
 ClientSampled :
 SSTD02008

Tot Ion	Ratio	Resp	Lower	Upper
266	100	110507		
264	65.6	55.0	82.6	
268	69.3	56.9	85.3	




Manual Integrations
 APPROVED


mohammad
 11/20/2020 11:31:29 AM



#70
 Phenanthrene
 Concen: 19.219 ng/ul
 RT: 16.89 min Scan# 2449
 Delta R.T. -0.01 min
 Lab File: BN012651.D
 Acq: 20 Nov 2020 02:45

Tgt Ion	Ratio	Resp	Lower	Upper
178	100	920142		
179	14.6	13.1	19.7	
176	18.7	15.4	23.0	









#72
 Anthracene
 Concen: 19.144 ng/ul
 RT: 16.99 min Scan# 2465
 Delta R.T. -0.01 min
 Lab File: BN012651.D
 Acq: 20 Nov 2020 02:45

Instrument :
 BNA_N
 ClientSampleId :
 SST02008

Tot Ion	Ratio	Lower	Upper
178	100		
179	15.7	12.4	18.6
176	16.9	14.8	22.2




Manual Integrations
 APPROVED


mohammad
 11/20/2020 11:31:29 AM



#73
 1,2,3,4-Tetrachlorobenzene
 Concen: 18.430 ng/uL
 RT: 12.88 min Scan# 1767
 Delta R.T. -0.01 min
 Lab File: BN012651.D
 Acq: 20 Nov 2020 02:45

Tgt Ion	Ratio	Lower	Upper
216	100		
214	74.2	58.5	87.7
218	43.4	36.0	54.0









#74
 Pentachlorobenzene
 Concen: 18.689 ng/uL
 RT: 14.42 min Scan# 2029
 Delta R.T. -0.01 min
 Lab File: BN012651.D
 Acq: 20 Nov 2020 02:45

Instrument :
 BNA_N
 ClientSampleId :
 SST02008

Tot Ion	Ratio	Lower	Upper
250	100		
248	59.3	50.7	76.1
252	68.1	49.6	74.4




Manual Integrations
 APPROVED


mohammad
 11/20/2020 11:31:29 AM



#75
 Carbazole
 Concen: 19.014 ng/uL
 RT: 17.26 min Scan# 2512
 Delta R.T. -0.01 min
 Lab File: BN012651.D
 Acq: 20 Nov 2020 02:45

Tgt Ion	Ratio	Lower	Upper
167	100		
166	21.9	17.8	26.8
139	14.5	10.4	15.6









#76
Di-n-butylphthalate
Concen: 20.766 ng/ul
RT: 17.84 min Scan# 2610
Delta R.T. -0.01 min
Lab File: BN012651.D
Acq: 20 Nov 2020 02:45

Instrument :
BNA_N
ClientSampleId :
SSTD02008

Tot Ion:149 Resp: 1064097
Ion Ratio Lower Upper
149 100
150 8.7 7.0 10.6
104 5.5 4.6 7.0




Manual Integrations
APPROVED


mohammad
11/20/2020 11:31:29 AM



#77
Fluoranthene
Concen: 19.145 ng/ul
RT: 18.92 min Scan# 2794
Delta R.T. -0.01 min
Lab File: BN012651.D
Acq: 20 Nov 2020 02:45

Tgt Ion:202 Resp: 1054188
Ion Ratio Lower Upper
202 100
101 10.4 10.2 15.2
100 8.5 7.7 11.5









#80
Pvrene
Concen: 18.004 ng/ul
RT: 19.29 min Scan# 2856
Delta R.T. -0.01 min
Lab File: BN012651.D
Acq: 20 Nov 2020 02:45

Instrument :
BNA_N
ClientSampleId :
SSTD02008

Tot Ion	Ratio	Lower	Upper
202	100		
101	12.3	10.1	15.1
100	10.8	8.5	12.7




Manual Integrations
APPROVED


mohammad
11/20/2020 11:31:29 AM



#81
Butylbenzylphthalate
Concen: 18.814 ng/ul
RT: 20.22 min Scan# 3014
Delta R.T. -0.01 min
Lab File: BN012651.D
Acq: 20 Nov 2020 02:45

Tgt Ion	Ratio	Lower	Upper
149	100		
91	78.2	58.6	88.0
206	19.5	16.9	25.3









#82
 3,3'-Dichlorobenzidine
 Concen: 18.316 ng/ul
 RT: 20.99 min Scan# 3146
 Delta R.T. -0.01 min
 Lab File: BN012651.D
 Acq: 20 Nov 2020 02:45

Instrument :
 BNA_N
 ClientSampleId :
 SST02008

Tot Ion	Ratio	Lower	Upper
252	100		
254	69.1	48.4	72.6
126	11.4	8.8	13.2




Manual Integrations
 APPROVED


mohammad
 11/20/2020 11:31:29 AM



#83
 Benzo(a)anthracene
 Concen: 18.654 ng/ul
 RT: 21.05 min Scan# 3156
 Delta R.T. -0.01 min
 Lab File: BN012651.D
 Acq: 20 Nov 2020 02:45

Tgt Ion	Ratio	Lower	Upper
228	100		
229	19.3	15.0	22.6
226	26.0	21.5	32.3









#84
 Bis(2-ethylhexyl)phthalate
 Concen: 19.312 ng/ul
 RT: 21.00 min Scan# 3147
 Delta R.T. -0.01 min
 Lab File: BN012651.D
 Acq: 20 Nov 2020 02:45

Instrument :
 BNA_N
 ClientSampleId :
 SST02008

Tot Ion	Ratio	Lower	Upper
149	100		
167	27.6	22.6	33.8
279	3.8	3.4	5.2




Manual Integrations
 APPROVED


mohammad
 11/20/2020 11:31:29 AM



#85
 Chrysene
 Concen: 18.693 ng/ul
 RT: 21.10 min Scan# 3164
 Delta R.T. -0.01 min
 Lab File: BN012651.D
 Acq: 20 Nov 2020 02:45

Tgt Ion	Ratio	Lower	Upper
228	100		
226	30.1	24.3	36.5
229	20.1	14.8	22.2









#87
 Di-n-octyl phthalate
 Concen: 19.124 ng/ul
 RT: 21.85 min Scan# 3292
 Delta R.T. -0.01 min
 Lab File: BN012651.D
 Acq: 20 Nov 2020 02:45

Instrument :
 BNA_N
 ClientSampleId :
 SST02008

Tgt Ion:149 Resp: 1263705




Manual Integrations
 APPROVED


mohammad
 11/20/2020 11:31:29 AM



#88
 Benzo(b)fluoranthene
 Concen: 18.816 ng/ul m
 RT: 22.56 min Scan# 3412
 Delta R.T. -0.01 min
 Lab File: BN012651.D
 Acq: 20 Nov 2020 02:45

Tgt Ion:252 Resp: 1166484
 Ion Ratio Lower Upper
 252 100
 253 20.1 17.2 25.8
 125 8.2 8.2 12.4#









#89
Benzo(k)fluoranthene
Concen: 19.251 ng/ul
RT: 22.60 min Scan# 3419
Delta R.T. -0.01 min
Lab File: BN012651.D
Acq: 20 Nov 2020 02:45

Instrument :
BNA_N
ClientSampleId :
SSTD02008

Tot Ion:252 Resp: 1147793
Ion Ratio Lower Upper
252 100
253 21.2 17.0 25.6
125 8.8 8.0 12.0




Manual Integrations
APPROVED


mohammad
11/20/2020 11:31:29 AM



#91
Benzo(a)pyrene
Concen: 18.698 ng/ul
RT: 23.09 min Scan# 3502
Delta R.T. -0.02 min
Lab File: BN012651.D
Acq: 20 Nov 2020 02:45

Tgt Ion:252 Resp: 1058905
Ion Ratio Lower Upper
252 100
253 23.1 15.6 23.4
125 11.2 9.2 13.8









#92
 Indeno(1,2,3-cd)pyrene
 Concen: 18.924 ng/ul
 RT: 25.27 min Scan# 3873
 Delta R.T. -0.02 min
 Lab File: BN012651.D
 Acq: 20 Nov 2020 02:45

Instrument :
 BNA_N
 ClientSampleId :
 SST02008

Tot Ion	Ratio	Lower	Upper
276	100		
138	21.8	19.2	28.8
277	25.7	21.4	32.0




Manual Integrations
 APPROVED


mohammad
 11/20/2020 11:31:29 AM



#93
 Dibenzo(a,h)anthracene
 Concen: 18.654 ng/ul
 RT: 25.27 min Scan# 3874
 Delta R.T. -0.02 min
 Lab File: BN012651.D
 Acq: 20 Nov 2020 02:45

Tgt Ion	Ratio	Lower	Upper
278	100		
139	14.0	12.4	18.6
279	24.4	17.7	26.5





#94

Benzo(a,h,i)perylene

Concen: 18.677 ng/ul

RT: 25.90 min Scan# 3980

Delta R.T. -0.02 min

Lab File: BN012651.D

Acq: 20 Nov 2020 02:45

Instrument :

BNA_N

ClientSampleId :

SSTD02008

Tot Ion: 276 Resp: 1116222

Ion Ratio Lower Upper

276 100

138 19.1 15.3 22.9



277 22.9 17.3 25.9

Manual Integrations

APPROVED

mohammad

11/20/2020 11:31:29 AM



Abundance Ion 276.00 (275.70 to 276.70): E

Ion 138.00 (137.70 to 138.70): E

Ion 277.00 (276.70 to 277.70): E

