


Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111820\
 Data File : BN012591.D
 Acq On : 18 Nov 2020 12:57
 Operator : CG/JU
 Sample : DFTPP17
 Misc :
 ALS Vial : 1 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 DFTPP17

Integration File: LSCINT.P

Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-BN102220.M
 Title : SVOA CALIBRATION
 Last Update : Wed Nov 18 14:31:59 2020



AutoFind: Scans 2521, 2522, 2523; Background Corrected with Scan 2515

Target Mass	Rel. to Mass	Lower Limit%	Upper Limit%	Rel. Abn%	Raw Abn	Result Pass/Fail
51	198	10	80	48.4	168472	PASS
68	69	0.00	2	1.3	2131	PASS
69	198	0.00	100	47.6	165512	PASS
70	69	0.00	2	0.5	891	PASS
127	198	10	80	51.2	178218	PASS
197	198	0.00	2	0.0	0	PASS
198	198	100	100	100.0	348053	PASS
199	198	5	9	6.7	23320	PASS
275	198	10	60	28.6	99469	PASS
365	198	1	100	4.3	14807	PASS
441	198	0.01	100	15.1	52421	PASS
442	198	50	100	97.0	337493	PASS
443	442	15	24	20.7	69693	PASS

Scan 2515 (17.280 min): BN012591.D
DFTPP17

m/z	Abundance
35.00	568.0
35.90	312.0
39.90	815.0
41.10	557.0
41.90	509.0
43.00	649.0
44.00	2293.0
47.00	1516.0
48.00	504.0
49.00	4578.0
50.10	306.0
51.00	1288.0
57.10	339.0
73.00	949.0
83.10	471.0
83.90	3031.0
85.90	1752.0
147.10	357.0
207.00	1594.0
207.90	501.0
281.10	462.0

Instrument :
BNA_N
ClientSampleId :
DFTPP17

m/z	Abundance
37.00	1391.0
38.00	2887.0
39.10	14287.0
40.10	1868.0
41.10	1054.0
43.10	488.0
44.00	2224.0
45.10	610.0
47.00	982.0
48.10	699.0
49.10	6170.0
50.10	53024.0
51.10	182592.0
52.10	8434.0
52.90	431.0
55.00	1472.0
56.00	5457.0
57.00	10574.0
58.00	605.0
59.90	323.0
60.90	2250.0
62.00	2773.0
63.00	7511.0
64.10	1148.0
65.00	3138.0
67.00	319.0
68.10	4011.0
69.00	175232.0
70.00	987.0
73.00	2529.0
74.00	19736.0
75.00	30392.0
76.10	10752.0
77.10	186816.0
78.10	12111.0
79.00	15705.0
80.00	13094.0
81.00	15672.0
82.00	4323.0
83.00	3179.0
84.00	3649.0
85.00	2581.0
86.00	6499.0
87.00	2091.0
87.90	826.0
89.00	425.0
91.00	3107.0
92.00	3360.0
93.00	24552.0
94.00	1525.0
94.90	514.0
95.90	1021.0
97.00	502.0
98.00	20720.0
99.00	14585.0
100.00	1124.0
101.00	7397.0
101.90	503.0
103.00	2300.0
104.00	4954.0
105.00	4625.0
106.10	1524.0
107.00	58992.0
108.00	7785.0
109.00	1817.0
110.00	97160.0
111.00	17344.0
111.90	2139.0
113.00	876.0
115.00	466.0
116.00	2935.0
117.00	61688.0
117.90	4968.0
119.10	558.0
120.00	830.0
122.00	3412.0
122.90	6200.0
123.90	3175.0
125.00	2684.0
127.00	181248.0
128.00	13771.0
129.00	74408.0
130.00	6124.0
130.90	1302.0
132.00	465.0
132.90	540.0
134.00	3068.0
135.00	6394.0
136.00	2911.0
137.00	3698.0
138.00	707.0
138.90	566.0

Instrument :
BNA_N
ClientSampleId :
DFTPP17

140.00	1159.0
141.00	11183.0
142.00	3400.0
143.00	1921.0
143.90	926.0
145.00	658.0
146.00	2092.0
147.00	5920.0
148.00	13043.0
149.00	2766.0
150.00	650.0
151.00	1625.0
151.70	951.0
153.00	3675.0
154.00	3001.0
155.00	4997.0
156.00	7196.0
157.00	1922.0
158.00	1897.0
159.00	1347.0
160.00	2797.0
161.00	4417.0
162.00	1044.0
163.00	703.0
164.00	945.0
165.00	4626.0
166.00	2861.0
167.00	19128.0
168.00	11345.0
169.00	1964.0
169.90	1042.0
171.00	1048.0
171.90	1803.0
173.00	2224.0
174.00	4044.0
175.00	7969.0
176.30	1928.0
177.00	3736.0
178.00	1509.0
179.00	14432.0
180.00	12200.0
181.00	4138.0
181.90	598.0
184.00	1177.0
185.00	6500.0
186.00	50088.0
187.00	14351.0
188.00	2018.0
189.00	4202.0
190.00	532.0
190.90	1755.0
192.00	4671.0
193.00	6166.0
194.00	1137.0
194.80	564.0
196.00	10119.0
198.00	337472.0
199.00	22048.0
200.00	2017.0
201.40	1840.0
203.10	3133.0
204.00	14748.0
205.00	20200.0
206.00	88320.0
207.00	13177.0
208.00	3397.0
209.10	1383.0
210.00	1322.0
211.10	2987.0
211.90	832.0
213.10	580.0
215.00	1089.0
216.00	2189.0
217.00	24592.0
218.00	3033.0
218.90	340.0
221.00	12790.0
221.90	3639.0
223.00	5251.0
224.10	45520.0
225.00	12944.0
226.00	1676.0
227.00	20344.0
228.00	3272.0
229.00	4302.0
230.10	533.0
231.10	1698.0
234.00	1211.0
235.00	1916.0
236.00	926.0
236.90	2132.0
239.00	921.0
240.10	943.0
241.00	1890.0
241.90	2635.0
243.00	3147.0

Instrument :
BNA_N
ClientSampleId :
DFTPP17

244.00	39368.0
245.00	4736.0
246.00	8386.0
247.00	1787.0
247.90	470.0
249.00	1060.0
251.90	359.0
253.00	1093.0
254.20	2808.0
255.00	193408.0
256.00	28760.0
257.00	2776.0
258.00	10553.0
259.00	1910.0
265.00	5021.0
266.00	722.0
267.10	388.0
271.00	458.0
271.90	514.0
273.00	5903.0
274.00	13634.0
275.00	89080.0
276.00	11225.0
277.10	8303.0
278.00	1775.0
279.00	330.0
280.90	735.0
282.90	774.0
284.10	769.0
285.10	1204.0
288.90	310.0
293.00	1454.0
294.00	483.0
295.20	756.0
296.00	26344.0
297.00	3664.0
301.10	669.0
301.90	568.0
303.00	2796.0
303.90	690.0
305.10	314.0
307.90	438.0
314.10	1493.0
315.00	2783.0
316.00	1914.0
321.00	867.0
321.90	453.0
323.00	6556.0
324.10	1460.0
327.00	1548.0
328.00	1088.0
331.90	521.0
333.00	718.0
334.00	5994.0
335.00	1262.0
336.00	452.0
340.90	872.0
341.80	319.0
346.00	1778.0
346.80	408.0
352.00	2349.0
352.90	1483.0
354.00	2595.0
355.00	659.0
365.00	11763.0
366.00	1511.0
367.00	483.0
371.20	431.0
372.00	4305.0
373.10	1228.0
383.00	730.0
384.00	374.0
390.20	368.0
391.00	300.0
401.90	1051.0
403.00	2323.0
403.80	637.0
404.80	369.0
421.10	1186.0
422.00	1516.0
423.00	10708.0
424.00	4756.0
425.00	625.0
441.00	38040.0
442.00	233728.0
443.00	47792.0
444.00	4471.0

Instrument :
BNA_N
ClientSampleId :
DFTPP17

m/z	Abundance
35.10	558.0
37.00	1012.0
38.00	3751.0
39.10	15877.0
40.00	1291.0
41.00	847.0
41.80	498.0
42.90	508.0
44.10	2512.0
45.00	587.0
46.90	1355.0
47.90	637.0
49.00	5831.0
50.00	54752.0
51.00	204160.0
52.10	9868.0
53.00	785.0
55.10	1471.0
56.00	5951.0
57.00	12725.0
58.00	880.0
60.00	501.0
61.00	2617.0
62.10	3553.0
63.00	7677.0
63.90	1601.0
64.90	3223.0
66.90	376.0
69.00	200704.0
70.00	1258.0
73.00	3541.0
74.00	24856.0
75.00	34160.0
76.10	11239.0
77.10	217792.0
78.00	14372.0
79.00	17576.0
80.00	11159.0
81.00	16792.0
82.00	4460.0
83.00	3672.0
84.00	3113.0
85.00	2642.0
86.00	8062.0
87.10	2426.0
87.90	1206.0
88.80	386.0
91.00	3659.0
92.00	3749.0
93.00	30880.0
94.00	1797.0
94.90	475.0
96.10	1245.0
97.10	517.0
98.00	25352.0
99.00	16448.0
99.90	1555.0
101.00	8213.0
102.10	511.0
103.00	2881.0
104.00	6350.0
105.00	5500.0
106.10	1906.0
107.00	69536.0
108.00	9809.0
109.10	2184.0
110.00	111488.0
111.00	18752.0
112.00	2388.0
113.00	651.0
115.00	613.0
116.00	5549.0
117.00	72000.0
118.00	4929.0
119.00	758.0
120.00	676.0
122.00	5090.0
123.00	7605.0
124.00	3628.0
125.00	2276.0
127.00	218944.0
128.00	16896.0
129.00	90024.0
130.00	7778.0
131.00	1468.0
132.10	536.0
132.80	577.0
133.20	558.0
134.00	3275.0
135.00	7545.0
136.00	2866.0
137.00	3644.0

Instrument :
BNA_N
ClientSampleId :
DFTPP17

137.90	949.0
138.90	461.0
140.00	1250.0
141.00	12883.0
141.90	3998.0
142.90	2760.0
144.00	787.0
145.00	714.0
145.90	2072.0
147.10	6800.0
148.00	16093.0
149.10	2909.0
150.10	796.0
151.00	2184.0
151.80	1191.0
153.00	3710.0
153.90	2465.0
155.00	6627.0
156.00	9128.0
157.00	1403.0
157.90	2359.0
158.90	1540.0
160.00	4318.0
161.00	5753.0
162.00	1811.0
162.90	673.0
164.00	565.0
165.00	5536.0
166.00	3583.0
167.00	27960.0
168.00	12739.0
169.00	1544.0
170.10	981.0
170.90	1008.0
171.90	2285.0
173.00	3370.0
174.00	5012.0
175.00	9321.0
175.90	3128.0
177.00	5166.0
178.00	1182.0
179.00	18736.0
180.00	11774.0
180.90	5132.0
182.10	1285.0
183.00	463.0
184.00	1339.0
185.00	8983.0
186.00	64448.0
187.00	17600.0
188.00	2465.0
188.90	4633.0
190.00	841.0
191.10	2505.0
192.00	6494.0
193.00	6463.0
194.10	1433.0
194.90	1057.0
196.00	13437.0
198.00	424256.0
199.00	29040.0
200.00	2343.0
201.60	2239.0
203.00	2872.0
204.00	18448.0
205.00	27392.0
206.00	116608.0
207.00	16235.0
208.00	4895.0
209.00	1475.0
210.00	2322.0
211.00	4613.0
211.80	1242.0
215.00	1576.0
216.00	3149.0
217.00	30920.0
218.00	3390.0
218.90	377.0
220.00	580.0
221.00	18480.0
221.80	4096.0
223.00	6741.0
224.00	64408.0
225.10	15553.0
226.10	2205.0
227.00	30912.0
228.00	4181.0
229.00	6406.0
230.00	759.0
231.00	2605.0
233.20	416.0
234.00	2325.0
235.10	1941.0
235.90	1451.0
237.00	2794.0
238.00	354.0

Instrument :
BNA_N
ClientSampleId :
DFTPP17

239.00	1368.0
239.90	848.0
241.10	1975.0
242.00	3475.0
243.00	3554.0
244.00	48728.0
245.00	6132.0
246.00	11171.0
247.00	2031.0
247.90	642.0
249.00	1925.0
249.80	426.0
251.00	512.0
251.90	453.0
253.00	1457.0
255.00	259584.0
256.00	39296.0
257.00	3938.0
258.00	16363.0
258.90	2962.0
260.00	482.0
260.90	549.0
263.90	498.0
265.00	6778.0
266.00	833.0
270.10	458.0
270.90	633.0
271.90	947.0
273.00	9723.0
274.00	20080.0
275.00	115080.0
276.00	15171.0
277.00	11686.0
278.00	1725.0
278.80	414.0
281.00	1016.0
282.00	573.0
282.90	1035.0
284.10	1011.0
285.10	2192.0
286.00	512.0
288.90	469.0
289.90	661.0
292.00	565.0
293.10	2262.0
294.10	765.0
296.00	40896.0
297.00	4516.0
298.00	354.0
301.00	456.0
302.00	790.0
303.00	4470.0
304.00	1085.0
307.90	390.0
309.00	373.0
309.90	636.0
314.00	1567.0
315.00	4378.0
316.00	2552.0
317.00	473.0
321.00	1210.0
321.90	514.0
323.00	9160.0
324.10	1986.0
325.90	411.0
327.00	2893.0
328.10	888.0
332.00	815.0
333.00	1178.0
334.00	7472.0
335.00	1771.0
341.10	1415.0
342.00	532.0
345.90	2521.0
346.90	851.0
352.10	3064.0
353.00	2619.0
354.10	3642.0
355.00	472.0
359.10	405.0
365.00	15916.0
366.00	2412.0
369.80	398.0
371.00	1122.0
372.00	5111.0
372.90	1514.0
383.00	1125.0
384.10	506.0
390.00	621.0
391.00	627.0
391.90	414.0
400.90	311.0
402.10	2644.0
403.00	3009.0
403.90	886.0
421.00	2804.0

Instrument :
BNA_N
ClientSampleId :
DFTPP17

421.90	3357.0
423.00	19648.0
424.10	6869.0
425.10	714.0
441.10	57336.0
442.00	417408.0
443.00	85024.0
444.00	6707.0
445.20	620.0

Instrument :
BNA_N
ClientSampleId :
DFTPP17

m/z	Abundance
35.90	317.0
37.00	812.0
38.10	2080.0
39.10	9313.0
40.00	1163.0
41.00	703.0
43.10	549.0
44.00	2237.0
45.10	528.0
47.00	856.0
48.00	701.0
49.00	5444.0
50.10	32296.0
51.10	122528.0
52.10	5792.0
53.00	549.0
55.10	1045.0
56.00	3462.0
57.00	8372.0
58.10	341.0
60.10	321.0
61.00	1617.0
62.10	1756.0
63.10	4274.0
64.10	718.0
65.00	2148.0
67.00	391.0
68.10	2383.0
69.00	120600.0
70.10	430.0
71.10	300.0
73.10	1566.0
74.00	14446.0
75.00	19216.0
76.10	6197.0
77.10	123736.0
78.10	7740.0
79.00	10894.0
80.00	7521.0
81.00	11272.0
82.00	2639.0
83.10	2430.0
84.00	3596.0
85.00	1740.0
85.90	5637.0
87.00	1466.0
88.00	934.0
91.00	2869.0
92.00	2534.0
93.00	18056.0
94.00	1055.0
96.10	997.0
98.00	14289.0
99.00	10376.0
100.00	805.0
101.00	5861.0
103.10	1586.0
104.00	3628.0
105.00	3444.0
106.10	858.0
107.00	39264.0
108.00	5733.0
109.00	1543.0
110.00	75792.0
111.00	10461.0
111.90	1603.0
113.00	765.0
114.90	401.0
116.10	2834.0
117.00	45432.0
118.00	2659.0
118.90	410.0
120.10	444.0
122.00	3293.0
123.00	5122.0
124.00	2010.0
125.00	1693.0
127.00	134464.0
128.00	9462.0
129.00	55792.0
130.00	4748.0
131.00	883.0
132.00	599.0
133.10	466.0
134.00	2581.0
135.00	5204.0
135.90	2256.0
137.00	3179.0
138.00	561.0
139.90	821.0
141.00	8559.0
142.00	3384.0

Instrument :
BNA_N
ClientSampleId :
DFTPP17

142.90	1329.0
144.00	521.0
145.10	517.0
146.00	1799.0
147.00	4401.0
148.00	11358.0
149.00	2271.0
150.10	640.0
151.20	1105.0
151.90	460.0
152.90	2130.0
154.00	1935.0
155.00	4416.0
156.00	7693.0
157.00	977.0
157.90	1945.0
159.00	1393.0
160.00	2432.0
161.00	3106.0
162.10	1204.0
162.90	345.0
164.10	399.0
165.00	2736.0
166.10	2715.0
167.00	17216.0
168.00	8033.0
169.00	1391.0
170.10	425.0
170.90	815.0
172.00	1191.0
173.00	1759.0
174.10	3151.0
175.00	6617.0
176.10	1493.0
177.00	2816.0
178.00	1036.0
179.00	12024.0
180.00	7769.0
181.00	4033.0
182.00	550.0
182.80	536.0
183.80	1064.0
184.10	1055.0
185.00	6177.0
186.00	41736.0
187.00	12623.0
188.00	1620.0
188.90	3206.0
190.10	568.0
191.10	1155.0
192.00	3904.0
193.00	4723.0
194.00	942.0
194.90	817.0
196.00	9459.0
198.00	282432.0
199.00	18872.0
200.00	1566.0
200.90	859.0
201.60	1223.0
203.00	2541.0
204.00	12867.0
205.10	19520.0
206.10	83216.0
207.10	11976.0
207.90	3292.0
209.00	1530.0
209.90	1295.0
210.30	1459.0
211.00	3780.0
214.90	1487.0
216.00	2091.0
217.00	22760.0
217.90	2783.0
220.10	589.0
221.00	11251.0
221.80	3363.0
223.00	4402.0
224.00	44480.0
225.00	11687.0
226.00	1049.0
227.00	21376.0
228.00	3279.0
229.00	4046.0
230.00	431.0
231.00	1422.0
233.00	489.0
234.00	1645.0
235.00	1285.0
236.00	755.0
237.00	2051.0
239.00	1242.0
240.00	794.0
240.90	1052.0
242.00	3100.0
243.00	2738.0

Instrument :
BNA_N
ClientSampleId :
DFTPP17

244.10	38128.0
245.10	6518.0
246.00	7741.0
247.00	1632.0
249.00	1547.0
249.80	347.0
251.00	397.0
252.90	686.0
255.00	198016.0
256.00	26192.0
257.00	2830.0
258.00	12163.0
259.00	1932.0
260.00	438.0
260.90	457.0
265.00	4943.0
265.90	1133.0
267.00	311.0
271.00	582.0
272.00	550.0
273.00	5451.0
274.00	16032.0
275.00	94248.0
276.00	10842.0
277.00	9589.0
277.90	1446.0
281.00	861.0
282.10	388.0
283.10	1298.0
284.00	874.0
285.00	1465.0
288.90	493.0
290.00	316.0
293.00	2107.0
294.10	535.0
295.20	791.0
296.00	32176.0
297.00	4368.0
301.00	385.0
301.90	615.0
303.10	2939.0
304.10	926.0
308.00	508.0
310.00	313.0
314.10	1167.0
315.00	2923.0
316.00	1616.0
317.00	414.0
320.00	380.0
321.00	1181.0
322.00	462.0
323.10	8720.0
323.90	1675.0
326.90	2441.0
328.00	934.0
329.00	432.0
331.80	496.0
332.90	529.0
334.10	5512.0
335.00	1778.0
341.00	1435.0
345.90	1935.0
346.90	470.0
352.00	2734.0
353.00	1570.0
354.00	3361.0
355.00	1050.0
358.90	398.0
365.00	16744.0
365.90	1463.0
369.90	338.0
370.90	795.0
372.00	4710.0
373.00	1418.0
383.00	1437.0
384.10	354.0
389.90	671.0
391.00	410.0
392.00	315.0
402.00	1608.0
402.90	2113.0
403.90	653.0
421.00	2636.0
422.30	1930.0
423.00	19088.0
424.00	4110.0
424.90	512.0
441.00	61888.0
442.00	361344.0
443.10	76264.0
444.00	5194.0
445.10	399.0

Instrument :
BNA_N
ClientSampleId :
DFTPP17