

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111921\
 Data File : BN017516.D
 Acq On : 19 Nov 2021 10:29
 Operator : CG/JU
 Sample : SSTDICC0.1
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Quant Time: Nov 19 12:30:32 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN111921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 19 12:29:22 2021
 Response via : Initial Calibration


Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)
Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.865	152	28211	0.400	ng	0.00
7) Naphthalene-d8	10.648	136	127464	0.400	ng	#-0.01
13) Acenaphthene-d10	14.485	164	67933	0.400	ng	0.00
19) Phenanthrene-d10	17.225	188	130892	0.400	ng	# 0.00
29) Chrysene-d12	21.409	240	112442	0.400	ng	# 0.00
35) Perylene-d12	23.776	264	97956	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.440	112	6856	0.102	ng	0.00
5) Phenol-d6	7.026	99	7604	0.105	ng	0.00
8) Nitrobenzene-d5	9.001	82	7610	0.096	ng	-0.01
11) 2-Methylnaphthalene-d10	12.239	152	20579	0.098	ng	0.00
14) 2,4,6-Tribromophenol	15.970	330	1863	0.108	ng	0.00
15) 2-Fluorobiphenyl	13.115	172	24742	0.097	ng	0.00
27) Fluoranthene-d10	19.258	212	36321	0.093	ng	0.00
31) Terphenyl-d14	19.858	244	24229	0.088	ng	0.00
Target Compounds						
2) 1,4-Dioxane	3.383	88	1348	0.109	ng	# 1
3) n-Nitrosodimethylamine	3.715	42	2537	0.107	ng	# 90
6) bis(2-Chloroethyl)ether	7.284	93	8160	0.105	ng	95
9) Naphthalene	10.699	128	32816	0.102	ng	99
10) Hexachlorobutadiene	11.003	225	6150	0.092	ng	# 99
12) 2-Methylnaphthalene	12.315	142	20492	0.101	ng	98
16) Acenaphthylene	14.206	152	22602	0.089	ng	100
17) Acenaphthene	14.549	154	17792	0.097	ng	99
18) Fluorene	15.526	166	21028	0.100	ng	99
20) 4,6-Dinitro-2-methylph...	15.615	198	263	0.051	ng	# 8
21) 4-Bromophenyl-phenylether	16.422	248	6511	0.088	ng	# 82
22) Hexachlorobenzene	16.544	284	7464	0.089	ng	95
23) Atrazine	16.690	200	4070	0.096	ng	# 96
24) Pentachlorophenol	16.885	266	1728	0.103	ng	98
25) Phenanthrene	17.262	178	35144	0.098	ng	99
26) Anthracene	17.359	178	26832	0.090	ng	100
28) Fluoranthene	19.287	202	36716	0.095	ng	100
30) Pyrene	19.650	202	36998	0.092	ng	100
32) Benzo(a)anthracene	21.398	228	30881	0.089	ng	99
33) Chrysene	21.451	228	36576	0.102	ng	99
34) Bis(2-ethylhexyl)phtha...	21.335	149	12414	0.068	ng	100
36) Indeno(1,2,3-cd)pyrene	26.220	276	25087	0.111	ng	95
37) Benzo(b)fluoranthene	23.058	252	31163	0.085	ng	# 96
38) Benzo(k)fluoranthene	23.106	252	31899	0.094	ng	# 95
39) Benzo(a)pyrene	23.674	252	24882	0.087	ng	# 94
40) Dibenzo(a,h)anthracene	26.241	278	20472	0.119	ng	# 90
41) Benzo(g,h,i)perylene	26.967	276	21820	0.103	ng	96


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111921\
Data File : BN017516.D
Acq On : 19 Nov 2021 10:29
Operator : CG/JU
Sample : SSTDICC0.1
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1

Quant Time: Nov 19 12:30:32 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN111921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 19 12:29:22 2021
Response via : Initial Calibration









#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.865 min Scan# 49
 Delta R.T. -0.000 min
 Lab File: BN017516.D
 Acq: 19 Nov 2021 10:29




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:152 Resp: 28211
 Ion Ratio Lower Upper
 152 100
 150 155.9 123.0 184.4
 115 61.2 45.6 68.4



#2
 1,4-Dioxane
 Concen: 0.109 ng
 RT: 3.383 min Scan# 6
 Delta R.T. -0.000 min
 Lab File: BN017516.D
 Acq: 19 Nov 2021 10:29

Tgt Ion: 88 Resp: 1348
 Ion Ratio Lower Upper
 88 100
 43 235.9 24.7 37.1#
 58 81.3 63.0 94.6









#3
n-Nitrosodimethylamine
Concen: 0.107 ng
RT: 3.715 min Scan# 41
Delta R.T. 0.009 min
Lab File: BN017516.D
Acq: 19 Nov 2021 10:29




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion: 42 Resp: 2537
Ion Ratio Lower Upper
42 100
74 134.2 97.8 146.6
44 4.4 5.5 8.3#



#4
2-Fluorophenol
Concen: 0.102 ng
RT: 5.440 min Scan# 229
Delta R.T. -0.009 min
Lab File: BN017516.D
Acq: 19 Nov 2021 10:29

Tgt Ion: 112 Resp: 6856
Ion Ratio Lower Upper
112 100
64 60.5 47.6 71.4
63 32.0 24.6 37.0









#5
Phenol-d6
Concen: 0.105 ng
RT: 7.026 min Scan# 401
Delta R.T. -0.000 min
Lab File: BN017516.D
Acq: 19 Nov 2021 10:29




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion: 99 Resp: 7604
Ion Ratio Lower Upper
99 100
42 20.7 17.0 25.6
71 34.2 24.8 37.2



#6
bis(2-Chloroethyl)ether
Concen: 0.105 ng
RT: 7.284 min Scan# 429
Delta R.T. -0.009 min
Lab File: BN017516.D
Acq: 19 Nov 2021 10:29

Tgt Ion: 93 Resp: 8160
Ion Ratio Lower Upper
93 100
63 76.7 65.8 98.6
95 32.0 26.6 39.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.648 min Scan# 71
Delta R.T. -0.013 min
Lab File: BN017516.D
Acq: 19 Nov 2021 10:29




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:136 Resp: 127464
Ion Ratio Lower Upper
136 100
137 11.0 8.8 13.2
54 9.4 4.0 6.0#
68 7.0 3.0 4.6#



#8
Nitrobenzene-d5
Concen: 0.096 ng
RT: 9.001 min Scan# 595
Delta R.T. -0.013 min
Lab File: BN017516.D
Acq: 19 Nov 2021 10:29

Tgt Ion: 82 Resp: 7610
Ion Ratio Lower Upper
82 100
128 32.3 27.7 41.5
54 58.7 49.2 73.8





#9

Naphthalene

Concen: 0.102 ng

RT: 10.699 min Scan# 729

Delta R.T. -0.000 min

Lab File: BN017516.D

Acq: 19 Nov 2021 10:29

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:128 Resp: 32816

Ion Ratio Lower Upper

128 100

129 11.3 8.8 13.2

127 13.5 10.6 15.8



#10

Hexachlorobutadiene

Concen: 0.092 ng

RT: 11.003 min Scan# 753

Delta R.T. -0.000 min

Lab File: BN017516.D

Acq: 19 Nov 2021 10:29




Tgt Ion:225 Resp: 6150


Ion Ratio Lower Upper

225 100

223 0.0 0.0 0.0

227 64.7 51.1 76.7









#11
2-Methylnaphthalene-d10
Concen: 0.098 ng
RT: 12.239 min Scan# 911
Delta R.T. -0.005 min
Lab File: BN017516.D
Acq: 19 Nov 2021 10:29




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:152 Resp: 20579
Ion Ratio Lower Upper
152 100
151 20.3 16.2 24.4



#12
2-Methylnaphthalene
Concen: 0.101 ng
RT: 12.315 min Scan# 974
Delta R.T. -0.000 min
Lab File: BN017516.D
Acq: 19 Nov 2021 10:29

Tgt Ion:142 Resp: 20492
Ion Ratio Lower Upper
142 100
141 88.1 70.6 106.0
115 36.5 25.8 38.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.485 min Scan# 1173
Delta R.T. -0.000 min
Lab File: BN017516.D
Acq: 19 Nov 2021 10:29




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:164 Resp: 67933
Ion Ratio Lower Upper
164 100
162 100.3 81.4 122.2
160 44.9 36.8 55.2



#14
2,4,6-Tribromophenol
Concen: 0.108 ng
RT: 15.970 min Scan# 1290
Delta R.T. -0.000 min
Lab File: BN017516.D
Acq: 19 Nov 2021 10:29

Tgt Ion:330 Resp: 1863
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 41.7 28.9 43.3









#15
2-Fluorobiphenyl
Concen: 0.097 ng
RT: 13.115 min Scan# 1065
Delta R.T. -0.000 min
Lab File: BN017516.D
Acq: 19 Nov 2021 10:29




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:172 Resp: 24742
Ion Ratio Lower Upper
172 100
171 34.2 28.7 43.1
170 22.6 19.4 29.2



#16
Acenaphthylene
Concen: 0.089 ng
RT: 14.206 min Scan# 1151
Delta R.T. -0.000 min
Lab File: BN017516.D
Acq: 19 Nov 2021 10:29

Tgt Ion:152 Resp: 22602
Ion Ratio Lower Upper
152 100
151 19.4 15.5 23.3
153 13.0 10.4 15.6









#17
Acenaphthene
Concen: 0.097 ng
RT: 14.549 min Scan# 1178
Delta R.T. -0.000 min
Lab File: BN017516.D
Acq: 19 Nov 2021 10:29




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:154 Resp: 17792
Ion Ratio Lower Upper
154 100
153 112.8 91.2 136.8
152 55.3 44.3 66.5



#18
Fluorene
Concen: 0.100 ng
RT: 15.526 min Scan# 1255
Delta R.T. -0.000 min
Lab File: BN017516.D
Acq: 19 Nov 2021 10:29

Tgt Ion:166 Resp: 21028
Ion Ratio Lower Upper
166 100
165 98.4 77.9 116.9
167 13.7 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.225 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN017516.D
Acq: 19 Nov 2021 10:29




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:188 Resp: 130892
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 9.8 6.1 9.1#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.051 ng
RT: 15.615 min Scan# 1262
Delta R.T. -0.000 min
Lab File: BN017516.D
Acq: 19 Nov 2021 10:29

Tgt Ion:198 Resp: 263
Ion Ratio Lower Upper
198 100
182 51.9 11.4 17.2#
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.088 ng
RT: 16.422 min Scan# 1327
Delta R.T. -0.000 min
Lab File: BN017516.D
Acq: 19 Nov 2021 10:29




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:248 Resp: 6511
Ion Ratio Lower Upper
248 100
250 96.2 71.8 107.8
141 60.4 70.9 106.3#



#22
Hexachlorobenzene
Concen: 0.089 ng
RT: 16.544 min Scan# 1337
Delta R.T. -0.000 min
Lab File: BN017516.D
Acq: 19 Nov 2021 10:29

Tgt Ion:284 Resp: 7464
Ion Ratio Lower Upper
284 100
142 39.2 28.2 42.2
249 31.9 24.4 36.6









#23
Atrazine
Concen: 0.096 ng
RT: 16.690 min Scan# 1349
Delta R.T. -0.000 min
Lab File: BN017516.D
Acq: 19 Nov 2021 10:29




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:200 Resp: 4070
Ion Ratio Lower Upper
200 100
173 26.4 17.6 26.4#
215 49.6 40.2 60.4



#24
Pentachlorophenol
Concen: 0.103 ng
RT: 16.885 min Scan# 1365
Delta R.T. -0.000 min
Lab File: BN017516.D
Acq: 19 Nov 2021 10:29

Tgt Ion:266 Resp: 1728
Ion Ratio Lower Upper
266 100
264 61.2 49.8 74.8
268 65.5 51.3 76.9









#25
Phenanthrene
Concen: 0.098 ng
RT: 17.262 min Scan# 1396
Delta R.T. -0.000 min
Lab File: BN017516.D
Acq: 19 Nov 2021 10:29




Instrument :
BNA_N
ClientSampleId :
SSTDIC0.1


Tgt Ion:178 Resp: 35144
Ion Ratio Lower Upper
178 100
176 19.1 15.0 22.6
179 15.6 12.3 18.5



#26
Anthracene
Concen: 0.090 ng
RT: 17.359 min Scan# 1404
Delta R.T. -0.000 min
Lab File: BN017516.D
Acq: 19 Nov 2021 10:29

Tgt Ion:178 Resp: 26832
Ion Ratio Lower Upper
178 100
176 18.2 14.6 22.0
179 15.2 12.2 18.4









#27
Fluoranthene-d10
Concen: 0.093 ng
RT: 19.258 min Scan# 1708
Delta R.T. -0.000 min
Lab File: BN017516.D
Acq: 19 Nov 2021 10:29




Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1


Tgt Ion:212 Resp: 36321
Ion Ratio Lower Upper
212 100
106 13.7 10.5 15.7
104 7.6 5.8 8.6



#28
Fluoranthene
Concen: 0.095 ng
RT: 19.287 min Scan# 1714
Delta R.T. -0.000 min
Lab File: BN017516.D
Acq: 19 Nov 2021 10:29

Tgt Ion:202 Resp: 36716
Ion Ratio Lower Upper
202 100
101 10.9 8.9 13.3
203 17.2 13.8 20.6





#29

Chrysene-d12

Concen: 0.400 ng

RT: 21.409 min Scan# 1991

Delta R.T. -0.000 min

Lab File: BN017516.D

Acq: 19 Nov 2021 10:29

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:240 Resp: 112442

Ion Ratio Lower Upper

240 100

120 13.0 4.2 6.2#

236 27.2 20.0 30.0



#30

Pyrene

Concen: 0.092 ng

RT: 19.650 min Scan# 1787

Delta R.T. -0.000 min

Lab File: BN017516.D

Acq: 19 Nov 2021 10:29




Tgt Ion:202 Resp: 36998


Ion Ratio Lower Upper

202 100

200 20.4 16.5 24.7

203 17.2 14.0 21.0









#31
 Terphenyl-d14
 Concen: 0.088 ng
 RT: 19.858 min Scan# 1829
 Delta R.T. -0.000 min
 Lab File: BN017516.D
 Acq: 19 Nov 2021 10:29




Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1


Tgt Ion:244 Resp: 24229
 Ion Ratio Lower Upper
 244 100
 212 8.2 5.7 8.5
 122 13.5 8.6 13.0#



#32
 Benzo(a)anthracene
 Concen: 0.089 ng
 RT: 21.398 min Scan# 1990
 Delta R.T. -0.000 min
 Lab File: BN017516.D
 Acq: 19 Nov 2021 10:29

Tgt Ion:228 Resp: 30881
 Ion Ratio Lower Upper
 228 100
 226 26.2 21.3 31.9
 229 19.7 15.7 23.5





#33

Chrysene

Concen: 0.102 ng

RT: 21.451 min Scan# 1995

Delta R.T. -0.000 min

Lab File: BN017516.D

Acq: 19 Nov 2021 10:29

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:228 Resp: 36576

Ion Ratio Lower Upper

228 100

226 29.2 23.3 34.9

229 20.2 15.6 23.4



#34

Bis(2-ethylhexyl)phthalate

Concen: 0.068 ng

RT: 21.335 min Scan# 1984

Delta R.T. -0.000 min

Lab File: BN017516.D

Acq: 19 Nov 2021 10:29




Tgt Ion:149 Resp: 12414


Ion Ratio Lower Upper

149 100

167 27.2 21.8 32.6

279 3.6 3.4 5.0





#35

Perylene-d12

Concen: 0.400 ng

RT: 23.776 min Scan# 2486

Delta R.T. -0.003 min

Lab File: BN017516.D


Acq: 19 Nov 2021 10:29

Instrument :

BNA_N

ClientSampleId :

SSTDIC0.1






Tgt Ion:264 Resp: 97956

Ion Ratio Lower Upper

264 100

260 25.5 19.4 29.2

265 24.0 18.6 28.0



#36

Indeno(1,2,3-cd)pyrene

Concen: 0.111 ng

RT: 26.220 min Scan# 3199

Delta R.T. 0.003 min

Lab File: BN017516.D

Acq: 19 Nov 2021 10:29




Tgt Ion:276 Resp: 25087


Ion Ratio Lower Upper

276 100

138 29.5 19.7 29.5

277 24.9 20.0 30.0





#37

Benzo(b)fluoranthene

Concen: 0.085 ng

RT: 23.058 min Scan# 21

Delta R.T. -0.000 min

Lab File: BN017516.D

Acq: 19 Nov 2021 10:29

Instrument :

BNA_N

ClientSampleId :

SSTDICC0.1





Tgt Ion:252 Resp: 31163

Ion Ratio Lower Upper

252 100

253 23.6 17.7 26.5

125 12.4 7.8 11.8#



#38

Benzo(k)fluoranthene

Concen: 0.094 ng

RT: 23.106 min Scan# 2289

Delta R.T. -0.000 min

Lab File: BN017516.D

Acq: 19 Nov 2021 10:29




Tgt Ion:252 Resp: 31899


Ion Ratio Lower Upper

252 100

253 23.9 17.7 26.5


125 13.4 8.2 12.2#








#39
Benzo(a)pyrene
Concen: 0.087 ng
RT: 23.674 min Scan# 2455
Delta R.T. -0.000 min
Lab File: BN017516.D
Acq: 19 Nov 2021 10:29


Instrument :
BNA_N
ClientSampleId :
SSTDICC0.1





Tgt Ion:252 Resp: 24882
Ion Ratio Lower Upper
252 100
253 25.0 17.9 26.9
125 14.5 9.2 13.8#




#40
Dibenzo(a,h)anthracene
Concen: 0.119 ng
RT: 26.241 min Scan# 3205
Delta R.T. 0.003 min
Lab File: BN017516.D
Acq: 19 Nov 2021 10:29



Tgt Ion:278 Resp: 20472
Ion Ratio Lower Upper
278 100
139 24.2 13.0 19.6#
279 26.1 19.2 28.8





#41
 Benzo(g,h,i)perylene
 Concen: 0.103 ng
 RT: 26.967 min Scan# 34
 Delta R.T. 0.003 min
 Lab File: BN017516.D
 Acq: 19 Nov 2021 10:29

Instrument :
 BNA_N
 ClientSampleId :
 SSTDICC0.1

Tgt Ion:	276	Resp:	21820
Ion Ratio	Lower	Upper	
276	100		
277	24.7	19.0	28.4
138	24.5	17.1	25.7

