

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111921\
 Data File : BN017523.D
 Acq On : 19 Nov 2021 14:54
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN111921

Quant Time: Nov 19 15:28:18 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN111921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 19 14:57:14 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	99	0.00
2	1,4-Dioxane	0.192	0.201	-4.7	109	0.00
3	n-Nitrosodimethylamine	0.395	0.389	1.5	99	0.00
4 S	2-Fluorophenol	0.991	0.987	0.4	98	0.00
5 S	Phenol-d6	1.118	1.100	1.6	97	0.00
6	bis(2-Chloroethyl)ether	1.142	1.188	-4.0	99	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	99	-0.01
8 S	Nitrobenzene-d5	0.263	0.244	7.2	100	-0.01
9	Naphthalene	1.015	1.027	-1.2	99	0.00
10	Hexachlorobutadiene	0.200	0.204	-2.0	100	0.00
11 SURR	2-Methylnaphthalene-d10	0.666	0.672	-0.9	99	0.00
12	2-Methylnaphthalene	0.653	0.670	-2.6	99	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	99	0.00
14 S	2,4,6-Tribromophenol	0.123	0.112	8.9	96	0.00
15 S	2-Fluorobiphenyl	1.490	1.527	-2.5	99	0.00
16	Acenaphthylene	1.596	1.514	5.1	100	0.00
17	Acenaphthene	1.102	1.108	-0.5	99	0.00
18	Fluorene	1.328	1.334	-0.5	99	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	99	0.00
20	4,6-Dinitro-2-methylphenol	0.019	0.013	31.6#	117	0.00
21	4-Bromophenyl-phenylether	0.220	0.212	3.6	99	0.00
22	Hexachlorobenzene	0.237	0.240	-1.3	101	0.00
23	Atrazine	0.146	0.136	6.8	99	0.00
24	Pentachlorophenol	0.072	0.047	34.7#	87	0.00
25	Phenanthrene	1.120	1.133	-1.2	100	0.00
26	Anthracene	0.957	0.910	4.9	99	0.00
27 SURR	Fluoranthene-d10	1.230	1.204	2.1	99	0.00
28	Fluoranthene	1.238	1.267	-2.3	99	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	100	0.00
30	Pyrene	1.353	1.353	0.0	101	0.00
31 S	Terphenyl-d14	0.879	0.877	0.2	100	0.00
32	Benzo(a)anthracene	1.245	1.169	6.1	99	0.00
33	Chrysene	1.330	1.346	-1.2	99	0.00
34	Bis(2-ethylhexyl)phthalate	0.480	0.408	15.0	97	0.00
35 I	Perylene-d12	1.000	1.000	0.0	105	0.00
36	Indeno(1,2,3-cd)pyrene	1.221	1.189	2.6	107	0.00
37	Benzo(b)fluoranthene	1.455	1.398	3.9	101	0.00
38	Benzo(k)fluoranthene	1.412	1.419	-0.5	104	0.00
39 C	Benzo(a)pyrene	1.201	1.134	5.6	104	0.00
40	Dibenzo(a,h)anthracene	1.002	0.976	2.6	107	0.00
41	Benzo(g,h,i)perylene	1.045	0.986	5.6	104	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

ClientSampleId :

ICVBN111921