

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111921\
 Data File : BN017529.D
 Acq On : 19 Nov 2021 21:15
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 20 01:32:31 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN111921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 19 14:57:14 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	101	0.00
2	1,4-Dioxane	0.400	0.375	6.3	100	0.00
3	n-Nitrosodimethylamine	0.400	0.380	5.0	98	0.00
4 S	2-Fluorophenol	0.400	0.405	-1.3	103	0.00
5 S	Phenol-d6	0.400	0.407	-1.7	103	0.00
6	bis(2-Chloroethyl)ether	0.400	0.421	-5.2	102	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	102	-0.01
8 S	Nitrobenzene-d5	0.400	0.374	6.5	104	-0.01
9	Naphthalene	0.400	0.406	-1.5	102	0.00
10	Hexachlorobutadiene	0.400	0.404	-1.0	102	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.406	-1.5	102	0.00
12	2-Methylnaphthalene	0.400	0.413	-3.2	102	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	104	0.00
14 S	2,4,6-Tribromophenol	0.400	0.373	6.8	103	0.00
15 S	2-Fluorobiphenyl	0.400	0.405	-1.3	103	0.00
16	Acenaphthylene	0.400	0.380	5.0	105	0.00
17	Acenaphthene	0.400	0.397	0.8	102	0.00
18	Fluorene	0.400	0.403	-0.8	104	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	105	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.294	26.5#	76	0.00
21	4-Bromophenyl-phenylether	0.400	0.385	3.8	104	0.00
22	Hexachlorobenzene	0.400	0.391	2.3	103	0.00
23	Atrazine	0.400	0.391	2.3	110	0.00
24	Pentachlorophenol	0.400	0.373	6.8	91	0.00
25	Phenanthrene	0.400	0.399	0.3	104	0.00
26	Anthracene	0.400	0.385	3.8	106	0.00
27 SURR	Fluoranthene-d10	0.400	0.402	-0.5	107	0.00
28	Fluoranthene	0.400	0.414	-3.5	106	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	106	0.00
30	Pyrene	0.400	0.399	0.3	107	0.00
31 S	Terphenyl-d14	0.400	0.396	1.0	105	0.00
32	Benzo(a)anthracene	0.400	0.412	-3.0	115	0.00
33	Chrysene	0.400	0.409	-2.2	106	-0.01
34	Bis(2-ethylhexyl)phthalate	0.400	0.425	-6.2	137	0.00
35 I	Perylene-d12	0.400	0.400	0.0	112	0.00
36	Indeno(1,2,3-cd)pyrene	0.400	0.379	5.3	111	0.00
37	Benzo(b)fluoranthene	0.400	0.419	-4.7	118	0.00
38	Benzo(k)fluoranthene	0.400	0.400	0.0	110	0.00
39 C	Benzo(a)pyrene	0.400	0.401	-0.3	117	0.00
40	Dibenzo(a,h)anthracene	0.400	0.372	7.0	109	0.00
41	Benzo(g,h,i)perylene	0.400	0.373	6.8	109	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4