

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111921\
 Data File : BN017523.D
 Acq On : 19 Nov 2021 14:54
 Operator : CG/JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN111921

Quant Time: Nov 19 15:28:18 2021
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN111921.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Fri Nov 19 14:57:14 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.865	152	28310	0.400	ng	0.00
7) Naphthalene-d8	10.649	136	126305	0.400	ng	#-0.01
13) Acenaphthene-d10	14.486	164	66617	0.400	ng	0.00
19) Phenanthrene-d10	17.226	188	128208	0.400	ng	# 0.00
29) Chrysene-d12	21.409	240	119330	0.400	ng	# 0.00
35) Perylene-d12	23.780	264	103038	0.400	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.440	112	27941	0.398	ng	0.00
5) Phenol-d6	7.026	99	31135	0.393	ng	0.00
8) Nitrobenzene-d5	9.001	82	30829	0.371	ng	-0.01
11) 2-Methylnaphthalene-d10	12.244	152	84923	0.404	ng	0.00
14) 2,4,6-Tribromophenol	15.970	330	7487	0.366	ng	0.00
15) 2-Fluorobiphenyl	13.115	172	101753	0.410	ng	0.00
27) Fluoranthene-d10	19.258	212	154398	0.392	ng	0.00
31) Terphenyl-d14	19.858	244	104704	0.399	ng	0.00
Target Compounds						
						Qvalue
2) 1,4-Dioxane	3.384	88	5696	0.419	ng	# 93
3) n-Nitrosodimethylamine	3.706	42	11016	0.394	ng	97
6) bis(2-Chloroethyl)ether	7.284	93	33640	0.416	ng	96
9) Naphthalene	10.699	128	129679	0.405	ng	100
10) Hexachlorobutadiene	11.003	225	25803	0.408	ng	# 100
12) 2-Methylnaphthalene	12.315	142	84621	0.410	ng	98
16) Acenaphthylene	14.206	152	100858	0.379	ng	100
17) Acenaphthene	14.549	154	73811	0.402	ng	99
18) Fluorene	15.526	166	88894	0.402	ng	100
20) 4,6-Dinitro-2-methylph...	15.615	198	1655	0.396	ng	97
21) 4-Bromophenyl-phenylether	16.422	248	27244	0.387	ng	# 82
22) Hexachlorobenzene	16.544	284	30780	0.404	ng	96
23) Atrazine	16.690	200	17464	0.374	ng	96
24) Pentachlorophenol	16.885	266	5968	0.375	ng	99
25) Phenanthrene	17.262	178	145257	0.405	ng	100
26) Anthracene	17.359	178	116717	0.380	ng	100
28) Fluoranthene	19.287	202	162460	0.409	ng	99
30) Pyrene	19.650	202	161431	0.400	ng	100
32) Benzo(a)anthracene	21.399	228	139514	0.376	ng	99
33) Chrysene	21.452	228	160671	0.405	ng	100
34) Bis(2-ethylhexyl)phtha...	21.335	149	48687	0.344	ng	99
36) Indeno(1,2,3-cd)pyrene	26.224	276	122534	0.390	ng	# 95
37) Benzo(b)fluoranthene	23.061	252	144074	0.384	ng	99
38) Benzo(k)fluoranthene	23.109	252	146241	0.402	ng	99
39) Benzo(a)pyrene	23.674	252	116869	0.378	ng	99
40) Dibenzo(a,h)anthracene	26.241	278	100570	0.389	ng	# 91
41) Benzo(g,h,i)perylene	26.967	276	101628	0.377	ng	97


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN111921\
Data File : BN017523.D
Acq On : 19 Nov 2021 14:54
Operator : CG/JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN111921


Quant Time: Nov 19 15:28:18 2021
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN111921.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Fri Nov 19 14:57:14 2021
Response via : Initial Calibration








#1
 1,4-Dichlorobenzene-d4
 Concen: 0.400 ng
 RT: 7.865 min Scan# 492
 Delta R.T. -0.000 min
 Lab File: BN017523.D
 Acq: 19 Nov 2021 14:54

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN111921







Tgt Ion:152 Resp: 28310
 Ion Ratio Lower Upper
 152 100
 150 155.0 123.0 184.4
 115 61.1 45.6 68.4



#2
 1,4-Dioxane
 Concen: 0.419 ng
 RT: 3.384 min Scan# 6
 Delta R.T. -0.000 min
 Lab File: BN017523.D
 Acq: 19 Nov 2021 14:54

Tgt Ion: 88 Resp: 5696
 Ion Ratio Lower Upper
 88 100
 43 40.6 24.7 37.1#
 58 80.9 63.0 94.6









#3
 n-Nitrosodimethylamine
 Concen: 0.394 ng
 RT: 3.706 min Scan# 41
 Delta R.T. 0.000 min
 Lab File: BN017523.D
 Acq: 19 Nov 2021 14:54




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN11921


Tgt Ion: 42 Resp: 11016
 Ion Ratio Lower Upper
 42 100
 74 126.0 97.8 146.6
 44 5.7 5.5 8.3



#4
 2-Fluorophenol
 Concen: 0.398 ng
 RT: 5.440 min Scan# 229
 Delta R.T. -0.009 min
 Lab File: BN017523.D
 Acq: 19 Nov 2021 14:54

Tgt Ion: 112 Resp: 27941
 Ion Ratio Lower Upper
 112 100
 64 60.6 47.6 71.4
 63 31.5 24.6 37.0









#5
Phenol-d6
Concen: 0.393 ng
RT: 7.026 min Scan# 401
Delta R.T. -0.000 min
Lab File: BN017523.D
Acq: 19 Nov 2021 14:54




Instrument :
BNA_N
ClientSampleId :
ICVBN111921


Tgt Ion: 99 Resp: 31135
Ion Ratio Lower Upper
99 100
42 20.6 17.0 25.6
71 33.9 24.8 37.2



#6
bis(2-Chloroethyl)ether
Concen: 0.416 ng
RT: 7.284 min Scan# 429
Delta R.T. -0.009 min
Lab File: BN017523.D
Acq: 19 Nov 2021 14:54

Tgt Ion: 93 Resp: 33640
Ion Ratio Lower Upper
93 100
63 77.3 65.8 98.6
95 32.1 26.6 39.8









#7
Naphthalene-d8
Concen: 0.400 ng
RT: 10.649 min Scan# 71
Delta R.T. -0.012 min
Lab File: BN017523.D
Acq: 19 Nov 2021 14:54




Instrument :
BNA_N
ClientSampleId :
ICVBN111921


Tgt Ion:136 Resp: 126305
Ion Ratio Lower Upper
136 100
137 10.9 8.8 13.2
54 9.3 4.0 6.0#
68 6.8 3.0 4.6#



#8
Nitrobenzene-d5
Concen: 0.371 ng
RT: 9.001 min Scan# 595
Delta R.T. -0.013 min
Lab File: BN017523.D
Acq: 19 Nov 2021 14:54

Tgt Ion: 82 Resp: 30829
Ion Ratio Lower Upper
82 100
128 33.0 27.7 41.5
54 57.7 49.2 73.8









#9
Naphthalene
Concen: 0.405 ng
RT: 10.699 min Scan# 729
Delta R.T. 0.000 min
Lab File: BN017523.D
Acq: 19 Nov 2021 14:54




Instrument :
BNA_N
ClientSampleId :
ICVBN111921


Tgt Ion:128 Resp: 129679
Ion Ratio Lower Upper
128 100
129 10.9 8.8 13.2
127 13.1 10.6 15.8



#10
Hexachlorobutadiene
Concen: 0.408 ng
RT: 11.003 min Scan# 753
Delta R.T. 0.000 min
Lab File: BN017523.D
Acq: 19 Nov 2021 14:54

Tgt Ion:225 Resp: 25803
Ion Ratio Lower Upper
225 100
223 0.0 0.0 0.0
227 64.1 51.1 76.7





#11

2-Methylnaphthalene-d10

Concen: 0.404 ng

RT: 12.244 min Scan# 911

Delta R.T. 0.000 min

Lab File: BN017523.D

Acq: 19 Nov 2021 14:54

Instrument :

BNA_N

ClientSampleId :





ICVBN111921

Tgt Ion:152 Resp: 84923

Ion Ratio Lower Upper

152 100

151 20.3 16.2 24.4



#12

2-Methylnaphthalene

Concen: 0.410 ng

RT: 12.315 min Scan# 974

Delta R.T. 0.000 min

Lab File: BN017523.D

Acq: 19 Nov 2021 14:54




Tgt Ion:142 Resp: 84621


Ion Ratio Lower Upper

142 100

141 88.6 70.6 106.0

115 35.9 25.8 38.6









#13
Acenaphthene-d10
Concen: 0.400 ng
RT: 14.486 min Scan# 1173
Delta R.T. -0.000 min
Lab File: BN017523.D
Acq: 19 Nov 2021 14:54




Instrument :
BNA_N
ClientSampleId :
ICVBN111921


Tgt Ion:164 Resp: 66617
Ion Ratio Lower Upper
164 100
162 101.1 81.4 122.2
160 45.3 36.8 55.2



#14
2,4,6-Tribromophenol
Concen: 0.366 ng
RT: 15.970 min Scan# 1290
Delta R.T. 0.000 min
Lab File: BN017523.D
Acq: 19 Nov 2021 14:54

Tgt Ion:330 Resp: 7487
Ion Ratio Lower Upper
330 100
332 0.0 0.0 0.0
141 40.6 28.9 43.3









#15
2-Fluorobiphenyl
Concen: 0.410 ng
RT: 13.115 min Scan# 1065
Delta R.T. 0.000 min
Lab File: BN017523.D
Acq: 19 Nov 2021 14:54




Instrument :
BNA_N
ClientSampleId :
ICVBN111921


Tgt Ion:172 Resp: 101753
Ion Ratio Lower Upper
172 100
171 33.9 28.7 43.1
170 22.7 19.4 29.2



#16
Acenaphthylene
Concen: 0.379 ng
RT: 14.206 min Scan# 1151
Delta R.T. 0.000 min
Lab File: BN017523.D
Acq: 19 Nov 2021 14:54

Tgt Ion:152 Resp: 100858
Ion Ratio Lower Upper
152 100
151 19.4 15.5 23.3
153 13.0 10.4 15.6









#17
Acenaphthene
Concen: 0.402 ng
RT: 14.549 min Scan# 1178
Delta R.T. 0.000 min
Lab File: BN017523.D
Acq: 19 Nov 2021 14:54




Instrument :
BNA_N
ClientSampleId :
ICVBN111921


Tgt Ion:154 Resp: 73811
Ion Ratio Lower Upper
154 100
153 112.5 91.2 136.8
152 55.0 44.3 66.5



#18
Fluorene
Concen: 0.402 ng
RT: 15.526 min Scan# 1255
Delta R.T. -0.000 min
Lab File: BN017523.D
Acq: 19 Nov 2021 14:54

Tgt Ion:166 Resp: 88894
Ion Ratio Lower Upper
166 100
165 97.9 77.9 116.9
167 13.5 10.9 16.3









#19
Phenanthrene-d10
Concen: 0.400 ng
RT: 17.226 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN017523.D
Acq: 19 Nov 2021 14:54




Instrument :
BNA_N
ClientSampleId :
ICVBN111921


Tgt Ion:188 Resp: 128208
Ion Ratio Lower Upper
188 100
94 0.0 0.0 0.0
80 9.4 6.1 9.1#



#20
4,6-Dinitro-2-methylphenol
Concen: 0.396 ng
RT: 15.615 min Scan# 1262
Delta R.T. -0.000 min
Lab File: BN017523.D
Acq: 19 Nov 2021 14:54

Tgt Ion:198 Resp: 1655
Ion Ratio Lower Upper
198 100
182 15.4 11.4 17.2
77 0.0 0.0 0.0









#21
4-Bromophenyl-phenylether
Concen: 0.387 ng
RT: 16.422 min Scan# 1327
Delta R.T. 0.000 min
Lab File: BN017523.D
Acq: 19 Nov 2021 14:54




Instrument :
BNA_N
ClientSampleId :
ICVBN111921


Tgt Ion:248 Resp: 27244
Ion Ratio Lower Upper
248 100
250 97.0 71.8 107.8
141 61.9 70.9 106.3#



#22
Hexachlorobenzene
Concen: 0.404 ng
RT: 16.544 min Scan# 1337
Delta R.T. 0.000 min
Lab File: BN017523.D
Acq: 19 Nov 2021 14:54

Tgt Ion:284 Resp: 30780
Ion Ratio Lower Upper
284 100
142 38.0 28.2 42.2
249 31.9 24.4 36.6









#23
Atrazine
Concen: 0.374 ng
RT: 16.690 min Scan# 1349
Delta R.T. 0.000 min
Lab File: BN017523.D
Acq: 19 Nov 2021 14:54




Instrument :
BNA_N
ClientSampleId :
ICVBN111921


Tgt Ion:	200	Resp:	17464
Ion Ratio	Lower	Upper	
200	100		
173	23.8	17.6	26.4
215	47.8	40.2	60.4



#24
Pentachlorophenol
Concen: 0.375 ng
RT: 16.885 min Scan# 1365
Delta R.T. -0.000 min
Lab File: BN017523.D
Acq: 19 Nov 2021 14:54

Tgt Ion:	266	Resp:	5968
Ion Ratio	Lower	Upper	
266	100		
264	63.1	49.8	74.8
268	64.2	51.3	76.9









#25
Phenanthrene
Concen: 0.405 ng
RT: 17.262 min Scan# 1396
Delta R.T. 0.000 min
Lab File: BN017523.D
Acq: 19 Nov 2021 14:54




Instrument :
BNA_N
ClientSampleId :
ICVBN111921


Tgt Ion:178 Resp: 145257
Ion Ratio Lower Upper
178 100
176 18.9 15.0 22.6
179 15.3 12.3 18.5



#26
Anthracene
Concen: 0.380 ng
RT: 17.359 min Scan# 1404
Delta R.T. 0.000 min
Lab File: BN017523.D
Acq: 19 Nov 2021 14:54

Tgt Ion:178 Resp: 116717
Ion Ratio Lower Upper
178 100
176 18.4 14.6 22.0
179 15.2 12.2 18.4









#27
Fluoranthene-d10
Concen: 0.392 ng
RT: 19.258 min Scan# 1708
Delta R.T. -0.000 min
Lab File: BN017523.D
Acq: 19 Nov 2021 14:54




Instrument :
BNA_N
ClientSampleId :
ICVBN111921


Tgt Ion:212 Resp: 154398
Ion Ratio Lower Upper
212 100
106 13.7 10.5 15.7
104 7.7 5.8 8.6



#28
Fluoranthene
Concen: 0.409 ng
RT: 19.287 min Scan# 1714
Delta R.T. 0.000 min
Lab File: BN017523.D
Acq: 19 Nov 2021 14:54

Tgt Ion:202 Resp: 162460
Ion Ratio Lower Upper
202 100
101 11.5 8.9 13.3
203 17.3 13.8 20.6









#29
Chrysene-d12
Concen: 0.400 ng
RT: 21.409 min Scan# 1991
Delta R.T. 0.000 min
Lab File: BN017523.D
Acq: 19 Nov 2021 14:54




Instrument :
BNA_N
ClientSampleId :
ICVBN111921


Tgt Ion:240 Resp: 119330
Ion Ratio Lower Upper
240 100
120 14.0 4.2 6.2#
236 26.8 20.0 30.0



#30
Pyrene
Concen: 0.400 ng
RT: 19.650 min Scan# 1787
Delta R.T. -0.000 min
Lab File: BN017523.D
Acq: 19 Nov 2021 14:54

Tgt Ion:202 Resp: 161431
Ion Ratio Lower Upper
202 100
200 20.6 16.5 24.7
203 17.4 14.0 21.0









#31
 Terphenyl-d14
 Concen: 0.399 ng
 RT: 19.858 min Scan# 1829
 Delta R.T. 0.000 min
 Lab File: BN017523.D
 Acq: 19 Nov 2021 14:54




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN111921


Tgt Ion:244 Resp: 104704
 Ion Ratio Lower Upper
 244 100
 212 8.2 5.7 8.5
 122 13.7 8.6 13.0#



#32
 Benzo(a)anthracene
 Concen: 0.376 ng
 RT: 21.399 min Scan# 1990
 Delta R.T. 0.001 min
 Lab File: BN017523.D
 Acq: 19 Nov 2021 14:54

Tgt Ion:228 Resp: 139514
 Ion Ratio Lower Upper
 228 100
 226 26.2 21.3 31.9
 229 19.8 15.7 23.5









#33
Chrysene
Concen: 0.405 ng
RT: 21.452 min Scan# 1995
Delta R.T. -0.000 min
Lab File: BN017523.D
Acq: 19 Nov 2021 14:54




Instrument :
BNA_N
ClientSampleId :
ICVBN111921


Tgt Ion:228 Resp: 160671
Ion Ratio Lower Upper
228 100
226 29.1 23.3 34.9
229 19.5 15.6 23.4



#34
Bis(2-ethylhexyl)phthalate
Concen: 0.344 ng
RT: 21.335 min Scan# 1984
Delta R.T. -0.000 min
Lab File: BN017523.D
Acq: 19 Nov 2021 14:54

Tgt Ion:149 Resp: 48687
Ion Ratio Lower Upper
149 100
167 27.5 21.8 32.6
279 3.4 3.4 5.0









#35
 Perylene-d12
 Concen: 0.400 ng
 RT: 23.780 min Scan# 24
 Delta R.T. 0.000 min
 Lab File: BN017523.D
 Acq: 19 Nov 2021 14:54




Instrument :
 BNA_N
 ClientSampleId :
 ICVBN111921


Tgt Ion:264 Resp: 103038
 Ion Ratio Lower Upper
 264 100
 260 25.1 19.4 29.2
 265 23.2 18.6 28.0



#36
 Indeno(1,2,3-cd)pyrene
 Concen: 0.390 ng
 RT: 26.224 min Scan# 3200
 Delta R.T. 0.007 min
 Lab File: BN017523.D
 Acq: 19 Nov 2021 14:54

Tgt Ion:276 Resp: 122534
 Ion Ratio Lower Upper
 276 100
 138 29.8 19.7 29.5#
 277 25.0 20.0 30.0









#37
Benzo(b)fluoranthene
Concen: 0.384 ng
RT: 23.061 min Scan# 21
Delta R.T. 0.003 min
Lab File: BN017523.D
Acq: 19 Nov 2021 14:54




Instrument :
BNA_N
ClientSampleId :
ICVBN111921


Tgt Ion:252 Resp: 144074
Ion Ratio Lower Upper
252 100
253 21.9 17.7 26.5
125 11.0 7.8 11.8



#38
Benzo(k)fluoranthene
Concen: 0.402 ng
RT: 23.109 min Scan# 2290
Delta R.T. 0.003 min
Lab File: BN017523.D
Acq: 19 Nov 2021 14:54


Tgt Ion:252 Resp: 146241
Ion Ratio Lower Upper
252 100
253 22.0 17.7 26.5
125 11.4 8.2 12.2








#39
Benzo(a)pyrene
Concen: 0.378 ng
RT: 23.674 min Scan# 2455
Delta R.T. -0.000 min
Lab File: BN017523.D
Acq: 19 Nov 2021 14:54

Instrument :
BNA_N
ClientSampleId :
ICVBN111921







Tgt Ion:252 Resp: 116869
Ion Ratio Lower Upper
252 100
253 22.0 17.9 26.9
125 12.0 9.2 13.8



#40
Dibenzo(a,h)anthracene
Concen: 0.389 ng
RT: 26.241 min Scan# 3205
Delta R.T. 0.003 min
Lab File: BN017523.D
Acq: 19 Nov 2021 14:54

Tgt Ion:278 Resp: 100570
Ion Ratio Lower Upper
278 100
139 25.0 13.0 19.6#
279 23.8 19.2 28.8





#41
 Benzo(g,h,i)perylene
 Concen: 0.377 ng
 RT: 26.967 min Scan# 34
 Delta R.T. 0.004 min
 Lab File: BN017523.D
 Acq: 19 Nov 2021 14:54

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN111921

Tgt Ion:276 Resp: 101628

Ion	Ratio	Lower	Upper
276	100		
277	23.9	19.0	28.4
138	23.8	17.1	25.7

