

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112018\
 Data File : BN003624.D
 Acq On : 20 Nov 2018 09:20
 Operator : MJ/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTD0.471

Quant Time: Nov 20 10:40:45 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN111918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 20 03:27:39 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	88	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	100	0.00
3	Naphthalene	1.124	1.136	-1.1	104	0.00
4 SURR	2-Methylnaphthalene-d10	0.615	0.611	0.7	102	0.00
5	2-Methylnaphthalene	0.785	0.790	-0.6	105	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	90	0.00
7	Acenaphthylene	1.801	1.824	-1.3	93	0.00
8 C	Acenaphthene	1.578	1.589	-0.7	91	0.00
9	Fluorene	1.840	1.840	0.0	90	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	88	0.00
11	Pentachlorophenol	0.072	0.072	0.0	103	0.00
12	Phenanthrene	1.313	1.318	-0.4	89	0.00
13	Anthracene	1.079	1.095	-1.5	91	0.00
14 SURR	Fluoranthene-d10	1.170	1.194	-2.1	92	0.00
15 C	Fluoranthene	1.550	1.575	-1.6	92	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
17	Pyrene	1.435	1.406	2.0	92	0.00
18	Benzo(a)anthracene	1.263	1.291	-2.2	108	0.00
19	Chrysene	1.446	1.444	0.1	103	0.00
20 I	Perylene-d12	1.000	1.000	0.0	103	0.00
21	Benzo(b)fluoranthene	1.655	1.653	0.1	104	0.00
22	Benzo(k)fluoranthene	1.612	1.566	2.9	101	0.00
23 C	Benzo(a)pyrene	1.433	1.433	0.0	104	0.00
24	Indeno(1,2,3-cd)pyrene	1.643	1.584	3.6	100	0.00
25	Dibenzo(a,h)anthracene	1.322	1.273	3.7	101	0.00
26	Benzo(g,h,i)perylene	1.428	1.351	5.4	96	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0