

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN112018\
 Data File : BN003633.D
 Acq On : 20 Nov 2018 16:02
 Operator : MJ/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTD0.472

Quant Time: Nov 20 16:53:13 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN111918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 20 03:27:39 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	77	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	88	0.00
3	Naphthalene	1.124	1.120	0.4	91	0.00
4 SURR	2-Methylnaphthalene-d10	0.615	0.612	0.5	91	0.00
5	2-Methylnaphthalene	0.785	0.785	0.0	92	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	79	0.00
7	Acenaphthylene	1.801	1.814	-0.7	82	0.00
8 C	Acenaphthene	1.578	1.582	-0.3	80	0.00
9	Fluorene	1.840	1.832	0.4	79	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	79	0.00
11	Pentachlorophenol	0.072	0.078	-8.3	101	0.00
12	Phenanthrene	1.313	1.309	0.3	79	0.00
13	Anthracene	1.079	1.087	-0.7	81	0.00
14 SURR	Fluoranthene-d10	1.170	1.188	-1.5	82	0.00
15 C	Fluoranthene	1.550	1.556	-0.4	81	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	94	0.00
17	Pyrene	1.435	1.382	3.7	82	0.00
18	Benzo(a)anthracene	1.263	1.285	-1.7	98	0.00
19	Chrysene	1.446	1.430	1.1	93	0.00
20 I	Perylene-d12	1.000	1.000	0.0	94	0.00
21	Benzo(b)fluoranthene	1.655	1.621	2.1	93	0.00
22	Benzo(k)fluoranthene	1.612	1.589	1.4	94	0.00
23 C	Benzo(a)pyrene	1.433	1.418	1.0	94	0.00
24	Indeno(1,2,3-cd)pyrene	1.643	1.591	3.2	92	0.00
25	Dibenzo(a,h)anthracene	1.322	1.276	3.5	92	0.00
26	Benzo(g,h,i)perylene	1.428	1.361	4.7	88	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0