

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN112018\
 Data File : BN003643.D
 Acq On : 20 Nov 2018 22:04
 Operator : MJ/SJ
 Sample : SSTDCCC0.4EC
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTD0.473

Quant Time: Nov 21 00:13:24 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN111918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 21 00:09:42 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	84	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	97	0.00
3	Naphthalene	1.124	1.134	-0.9	101	0.00
4 SURR	2-Methylnaphthalene-d10	0.615	0.622	-1.1	101	0.00
5	2-Methylnaphthalene	0.785	0.797	-1.5	102	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	89	0.00
7	Acenaphthylene	1.801	1.843	-2.3	93	0.00
8 C	Acenaphthene	1.578	1.570	0.5	89	0.00
9	Fluorene	1.840	1.835	0.3	89	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	87	0.00
11	Pentachlorophenol	0.072	0.095	-31.9#	136	0.00
12	Phenanthrene	1.313	1.309	0.3	88	0.00
13	Anthracene	1.079	1.099	-1.9	91	0.00
14 SURR	Fluoranthene-d10	1.170	1.192	-1.9	91	0.00
15 C	Fluoranthene	1.550	1.562	-0.8	90	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	103	0.00
17	Pyrene	1.435	1.393	2.9	91	0.00
18	Benzo(a)anthracene	1.263	1.307	-3.5	109	0.00
19	Chrysene	1.446	1.431	1.0	102	0.00
20 I	Perylene-d12	1.000	1.000	0.0	104	0.00
21	Benzo(b)fluoranthene	1.655	1.666	-0.7	105	0.00
22	Benzo(k)fluoranthene	1.612	1.576	2.2	102	0.00
23 C	Benzo(a)pyrene	1.433	1.443	-0.7	105	0.00
24	Indeno(1,2,3-cd)pyrene	1.643	1.573	4.3	100	0.00
25	Dibenzo(a,h)anthracene	1.322	1.266	4.2	101	-0.01
26	Benzo(g,h,i)perylene	1.428	1.323	7.4	95	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0