

Data Path : Z:\SVOASRV\HPCHEM1\BNA_N\DATA\BN112018\
 Data File : BN003645.D
 Acq On : 21 Nov 2018 08:55
 Operator : MJ/SJ
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTD0.474

Quant Time: Nov 21 15:52:52 2018
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SOM-EPA-SIM-BN111918.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Wed Nov 21 00:09:42 2018
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 20% Max. R.T. Dev 0.50min
 Max. RRF Dev : 20% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	79	0.00
2 I	Naphthalene-d8	1.000	1.000	0.0	90	0.00
3	Naphthalene	1.124	1.138	-1.2	94	0.00
4 SURR	2-Methylnaphthalene-d10	0.615	0.616	-0.2	93	0.00
5	2-Methylnaphthalene	0.785	0.797	-1.5	95	0.00
6 I	Acenaphthene-d10	1.000	1.000	0.0	81	0.00
7	Acenaphthylene	1.801	1.838	-2.1	85	0.00
8 C	Acenaphthene	1.578	1.601	-1.5	83	0.00
9	Fluorene	1.840	1.852	-0.7	82	0.00
10 I	Phenanthrene-d10	1.000	1.000	0.0	80	0.00
11	Pentachlorophenol	0.072	0.070	2.8	92	0.00
12	Phenanthrene	1.313	1.315	-0.2	81	0.00
13	Anthracene	1.079	1.090	-1.0	83	0.00
14 SURR	Fluoranthene-d10	1.170	1.227	-4.9	86	0.00
15 C	Fluoranthene	1.550	1.624	-4.8	86	0.00
16 I	Chrysene-d12	1.000	1.000	0.0	93	0.00
17	Pyrene	1.435	1.436	-0.1	85	0.00
18	Benzo(a)anthracene	1.263	1.284	-1.7	97	0.00
19	Chrysene	1.446	1.443	0.2	93	0.00
20 I	Perylene-d12	1.000	1.000	0.0	94	0.00
21	Benzo(b)fluoranthene	1.655	1.628	1.6	93	0.00
22	Benzo(k)fluoranthene	1.612	1.572	2.5	93	0.00
23 C	Benzo(a)pyrene	1.433	1.425	0.6	94	0.00
24	Indeno(1,2,3-cd)pyrene	1.643	1.577	4.0	91	0.00
25	Dibenzo(a,h)anthracene	1.322	1.244	5.9	90	0.00
26	Benzo(g,h,i)perylene	1.428	1.365	4.4	89	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0