

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN112219\
 Data File : BN008721.D
 Acq On : 23 Nov 2019 00:35
 Operator : JU
 Sample : SSTDICV0.4
 Misc :
 ALS Vial : 9 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 ICVBN112219

Quant Time: Nov 25 10:58:43 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN112219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 25 10:49:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	104	0.00
2	1,4-Dioxane	0.423	0.417	1.4	103	0.00
3	n-Nitrosodimethylamine	0.524	0.510	2.7	109	0.00
4 S	2-Fluorophenol	1.048	1.038	1.0	103	0.00
5 S	Phenol-d6	1.324	1.300	1.8	102	0.00
6	bis(2-Chloroethyl)ether	1.241	1.248	-0.6	103	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	104	0.00
8 S	Nitrobenzene-d5	0.359	0.355	1.1	103	0.00
9	Naphthalene	1.068	1.083	-1.4	103	0.00
10	Hexachlorobutadiene	0.221	0.223	-0.9	102	0.00
11 SURR	2-Methylnaphthalene-d10	0.631	0.637	-1.0	105	0.00
12	2-Methylnaphthalene	0.743	0.747	-0.5	103	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	103	0.00
14 S	2,4,6-Tribromophenol	0.186	0.172	7.5	100	0.00
15 S	2-Fluorobiphenyl	1.563	1.572	-0.6	103	0.00
16	Acenaphthylene	1.841	1.767	4.0	101	0.00
17	Acenaphthene	1.212	1.191	1.7	104	0.00
18	Fluorene	1.552	1.542	0.6	103	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	103	0.00
20	4-Bromophenyl-phenylether	0.250	0.248	0.8	103	0.00
21	Hexachlorobenzene	0.275	0.276	-0.4	104	0.00
22	Pentachlorophenol	0.098	0.079	19.4	97	0.00
23	Phenanthrene	1.233	1.220	1.1	103	-0.01
24	Anthracene	1.103	1.067	3.3	102	0.00
25 SURR	Fluoranthene-d10	1.176	1.167	0.8	103	0.00
26	Fluoranthene	1.448	1.429	1.3	103	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	108	0.00
28	Pvrene	1.477	1.414	4.3	102	0.00
29 S	Terphenyl-d14	0.962	0.932	3.1	103	0.00
30	Benzo(a)anthracene	1.464	1.396	4.6	104	-0.01
31	Chrysene	1.418	1.403	1.1	106	0.00
32	Bis(2-ethylhexyl)phthalate	0.951	0.838	11.9	103	-0.01
33	Indeno(1,2,3-cd)pyrene	1.580	1.558	1.4	105	0.00
34 I	Perylene-d12	1.000	1.000	0.0	105	-0.01
35	Benzo(b)fluoranthene	1.384	1.366	1.3	104	0.00
36	Benzo(k)fluoranthene	1.402	1.415	-0.9	104	0.00
37 C	Benzo(a)pyrene	1.281	1.265	1.2	104	0.00
38	Dibenzo(a,h)anthracene	1.264	1.280	-1.3	106	-0.01
39	Benzo(a,h,i)perylene	1.260	1.268	-0.6	105	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN112219\
Data File : BN008721.D
Acq On : 23 Nov 2019 00:35
Operator : JU
Sample : SSTDICV0.4
Misc :
ALS Vial : 9 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
ICVBN112219

Quant Time: Nov 25 10:58:43 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN112219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 25 10:49:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			