

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN112319\
 Data File : BN008723.D
 Acq On : 23 Nov 2019 11:03
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 25 11:45:27 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN112219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 25 10:49:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	129	0.00
2	1,4-Dioxane	0.423	0.425	-0.5	130	0.00
3	n-Nitrosodimethylamine	0.524	0.480	8.4	126	0.00
4 S	2-Fluorophenol	1.048	1.033	1.4	127	0.00
5 S	Phenol-d6	1.324	1.282	3.2	124	0.00
6	bis(2-Chloroethyl)ether	1.241	1.236	0.4	126	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	128	0.00
8 S	Nitrobenzene-d5	0.359	0.354	1.4	126	0.00
9	Naphthalene	1.068	1.090	-2.1	127	0.00
10	Hexachlorobutadiene	0.221	0.225	-1.8	127	0.00
11 SURR	2-Methylnaphthalene-d10	0.631	0.634	-0.5	128	0.00
12	2-Methylnaphthalene	0.743	0.747	-0.5	127	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	127	0.00
14 S	2,4,6-Tribromophenol	0.186	0.169	9.1	121	0.00
15 S	2-Fluorobiphenyl	1.563	1.577	-0.9	127	0.00
16	Acenaphthylene	1.841	1.791	2.7	126	0.00
17	Acenaphthene	1.212	1.182	2.5	127	0.00
18	Fluorene	1.552	1.526	1.7	126	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	127	0.00
20	4-Bromophenyl-phenylether	0.250	0.247	1.2	128	0.00
21	Hexachlorobenzene	0.275	0.276	-0.4	128	-0.01
22	Pentachlorophenol	0.098	0.081	17.3	122	0.00
23	Phenanthrene	1.233	1.228	0.4	128	-0.01
24	Anthracene	1.103	1.060	3.9	125	0.00
25 SURR	Fluoranthene-d10	1.176	1.157	1.6	126	0.00
26	Fluoranthene	1.448	1.421	1.9	127	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	132	0.00
28	Pvrene	1.477	1.419	3.9	126	0.00
29 S	Terphenyl-d14	0.962	0.935	2.8	127	0.00
30	Benzo(a)anthracene	1.464	1.434	2.0	130	-0.01
31	Chrysene	1.418	1.383	2.5	127	0.00
32	Bis(2-ethylhexyl)phthalate	0.951	0.739	22.3	111	0.00
33	Indeno(1,2,3-cd)pyrene	1.580	1.546	2.2	128	0.00
34 I	Perylene-d12	1.000	1.000	0.0	128	-0.01
35	Benzo(b)fluoranthene	1.384	1.405	-1.5	130	0.00
36	Benzo(k)fluoranthene	1.402	1.390	0.9	125	0.00
37 C	Benzo(a)pyrene	1.281	1.262	1.5	126	0.00
38	Dibenzo(a,h)anthracene	1.264	1.269	-0.4	127	0.00
39	Benzo(a,h,i)perylene	1.260	1.277	-1.3	129	0.00

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN112319\
Data File : BN008723.D
Acq On : 23 Nov 2019 11:03
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Nov 25 11:45:27 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN112219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 25 10:49:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0	CCC's out = 0		