

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN112319\
 Data File : BN008739.D
 Acq On : 23 Nov 2019 21:10
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 25 11:48:06 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN112219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 25 10:49:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	106	0.00
2	1,4-Dioxane	0.423	0.410	3.1	104	0.00
3	n-Nitrosodimethylamine	0.524	0.494	5.7	108	0.00
4 S	2-Fluorophenol	1.048	1.063	-1.4	108	0.00
5 S	Phenol-d6	1.324	1.397	-5.5	112	0.00
6	bis(2-Chloroethyl)ether	1.241	1.293	-4.2	109	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	112	0.00
8 S	Nitrobenzene-d5	0.359	0.360	-0.3	113	-0.01
9	Naphthalene	1.068	1.074	-0.6	110	-0.01
10	Hexachlorobutadiene	0.221	0.221	0.0	109	-0.01
11 SURR	2-Methylnaphthalene-d10	0.631	0.640	-1.4	113	0.00
12	2-Methylnaphthalene	0.743	0.758	-2.0	113	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	116	0.00
14 S	2,4,6-Tribromophenol	0.186	0.185	0.5	121	0.00
15 S	2-Fluorobiphenyl	1.563	1.563	0.0	115	0.00
16	Acenaphthylene	1.841	1.856	-0.8	119	0.00
17	Acenaphthene	1.212	1.197	1.2	118	0.00
18	Fluorene	1.552	1.560	-0.5	118	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	116	0.00
20	4-Bromophenyl-phenylether	0.250	0.249	0.4	118	0.00
21	Hexachlorobenzene	0.275	0.278	-1.1	118	0.00
22	Pentachlorophenol	0.098	0.082	16.3	114	0.00
23	Phenanthrene	1.233	1.224	0.7	117	0.00
24	Anthracene	1.103	1.089	1.3	118	0.00
25 SURR	Fluoranthene-d10	1.176	1.161	1.3	116	0.00
26	Fluoranthene	1.448	1.413	2.4	116	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	118	0.00
28	Pvrene	1.477	1.461	1.1	116	0.00
29 S	Terphenyl-d14	0.962	0.965	-0.3	117	0.00
30	Benzo(a)anthracene	1.464	1.414	3.4	115	0.00
31	Chrysene	1.418	1.435	-1.2	118	0.00
32	Bis(2-ethylhexyl)phthalate	0.951	0.787	17.2	106	0.00
33	Indeno(1,2,3-cd)pyrene	1.580	1.586	-0.4	117	0.00
34 I	Perylene-d12	1.000	1.000	0.0	117	-0.01
35	Benzo(b)fluoranthene	1.384	1.390	-0.4	117	0.00
36	Benzo(k)fluoranthene	1.402	1.405	-0.2	116	0.00
37 C	Benzo(a)pyrene	1.281	1.261	1.6	115	0.00
38	Dibenzo(a,h)anthracene	1.264	1.272	-0.6	117	-0.01
39	Benzo(a,h,i)perylene	1.260	1.261	-0.1	116	-0.01

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN112319\
Data File : BN008739.D
Acq On : 23 Nov 2019 21:10
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Nov 25 11:48:06 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN112219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 25 10:49:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			