

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112320\
 Data File : BN012703.D
 Acq On : 23 Nov 2020 18:54
 Operator : CG/JU
 Sample : L4788-02
 Misc :
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-TT-BPOW1-8-GW-20201113

Quant Time: Nov 24 01:54:27 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN110420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 23 12:44:50 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.441	152	613	0.40	ng	# 0.00
7) Naphthalene-d8	10.212	136	2417	0.40	ng	# 0.01
13) Acenaphthene-d10	14.090	164	1351	0.40	ng	0.00
19) Phenanthrene-d10	16.846	188	2686	0.40	ng	# 0.00
29) Chrysene-d12	21.067	240	3057	0.40	ng	# 0.00
36) Perylene-d12	23.183	264	3481	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.057	112	211	0.10	ng	0.00
5) Phenol-d6	6.636	99	170	0.07	ng	0.00
8) Nitrobenzene-d5	8.590	82	1231	0.43	ng	0.00
11) 2-Methylnaphthalene-d10	11.811	152	1086	0.28	ng	0.00
14) 2,4,6-Tribromophenol	15.600	330	298	0.30	ng	0.00
15) 2-Fluorobiphenyl	12.707	172	2443	0.46	ng	0.00
27) Fluoranthene-d10	18.893	212	3085	0.37	ng	0.00
31) Terphenyl-d14	19.509	244	4353	0.60	ng	0.00
Target Compounds						Qvalue
25) Phenanthrene	16.895	178	188	0.02	ng	# 93
34) Bis(2-ethylhexyl)phtha...	20.993	149	938	0.14	ng	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112320\
Data File : BN012703.D
Acq On : 23 Nov 2020 18:54
Operator : CG/JU
Sample : L4788-02
Misc :
ALS Vial : 13 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-TT-BPOW1-8-GW-20201113

Quant Time: Nov 24 01:54:27 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN110420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 23 12:44:50 2020
Response via : Initial Calibration

