

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112320\
 Data File : BN012705.D
 Acq On : 23 Nov 2020 20:06
 Operator : CG/JU
 Sample : L4787-03
 Misc :
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 BP-VPB178-GW-678-680

Quant Time: Nov 24 01:54:42 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN110420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 23 12:44:50 2020
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.441	152	468	0.40	ng	# 0.00
7) Naphthalene-d8	10.212	136	1989	0.40	ng	# 0.01
13) Acenaphthene-d10	14.090	164	1178	0.40	ng	0.00
19) Phenanthrene-d10	16.846	188	2483	0.40	ng	# 0.00
29) Chrysene-d12	21.067	240	2643	0.40	ng	# 0.00
36) Perylene-d12	23.182	264	3208	0.40	ng	0.00
System Monitoring Compounds						
4) 2-Fluorophenol	5.049	112	249	0.15	ng	0.00
5) Phenol-d6	6.628	99	177	0.09	ng	0.00
8) Nitrobenzene-d5	8.590	82	1071	0.45	ng	0.00
11) 2-Methylnaphthalene-d10	11.806	152	966	0.30	ng	0.00
14) 2,4,6-Tribromophenol	15.600	330	275	0.32	ng	0.00
15) 2-Fluorobiphenyl	12.707	172	2067	0.45	ng	0.00
27) Fluoranthene-d10	18.893	212	2764	0.36	ng	0.00
31) Terphenyl-d14	19.504	244	4088	0.65	ng	0.00
Target Compounds						Qvalue
34) Bis(2-ethylhexyl)phtha...	20.993	149	4062	0.71	ng	# 90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112320\
Data File : BN012705.D
Acq On : 23 Nov 2020 20:06
Operator : CG/JU
Sample : L4787-03
Misc :
ALS Vial : 15 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
BP-VPB178-GW-678-680

Quant Time: Nov 24 01:54:42 2020
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN110420.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 23 12:44:50 2020
Response via : Initial Calibration

