

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112320\
 Data File : BN012692.D
 Acq On : 23 Nov 2020 11:48
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 23 13:36:33 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN110420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 23 12:44:50 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	85	0.00
2	1,4-Dioxane	0.626	0.593	5.3	86	0.00
3	n-Nitrosodimethylamine	1.709	1.495	12.5	83	0.00
4 S	2-Fluorophenol	1.398	1.244	11.0	81	0.00
5 S	Phenol-d6	1.636	1.403	14.2	77	0.00
6	bis(2-Chloroethyl)ether	1.115	0.972	12.8	85	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	85	0.00
8 S	Nitrobenzene-d5	0.473	0.432	8.7	87	0.00
9	Naphthalene	1.122	1.060	5.5	90	0.00
10	Hexachlorobutadiene	0.245	0.228	6.9	87	0.00
11 SURR	2-Methylnaphthalene-d10	0.649	0.614	5.4	92	0.00
12	2-Methylnaphthalene	0.724	0.683	5.7	91	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	87	0.00
14 S	2,4,6-Tribromophenol	0.290	0.275	5.2	96	0.00
15 S	2-Fluorobiphenyl	1.560	1.384	11.3	86	0.00
16	Acenaphthylene	2.005	1.832	8.6	92	0.00
17	Acenaphthene	1.276	1.159	9.2	91	0.00
18	Fluorene	1.627	1.441	11.4	89	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	85	0.00
20	4,6-Dinitro-2-methylphenol	0.097	0.086	11.3	91	0.00
21	4-Bromophenyl-phenylether	0.231	0.203	12.1	85	0.00
22	Hexachlorobenzene	0.301	0.290	3.7	90	0.00
23	Atrazine	0.240	0.218	9.2	88	0.00
24	Pentachlorophenol	0.141	0.133	5.7	96	0.00
25	Phenanthrene	1.184	1.126	4.9	91	0.00
26	Anthracene	1.096	1.024	6.6	91	0.00
27 SURR	Fluoranthene-d10	1.240	1.177	5.1	92	0.00
28	Fluoranthene	1.413	1.327	6.1	91	0.00
29 I	Chrysene-d12	1.000	1.000	0.0	88	0.00
30	Pyrene	1.344	1.253	6.8	92	0.00
31 S	Terphenyl-d14	0.954	0.883	7.4	90	0.00
32	Benzo(a)anthracene	1.273	1.216	4.5	93	0.00
33	Chrysene	1.333	1.204	9.7	89	0.00
34	Bis(2-ethylhexyl)phthalate	0.864	0.857	0.8	101	0.00
35	Indeno(1,2,3-cd)pyrene	2.006	1.890	5.8	86	0.00
36 I	Perylene-d12	1.000	1.000	0.0	88	0.00
37	Benzo(b)fluoranthene	1.276	1.162	8.9	92	0.00
38	Benzo(k)fluoranthene	1.360	1.233	9.3	89	0.00
39 C	Benzo(a)pyrene	1.249	1.124	10.0	89	0.00
40	Dibenzo(a,h)anthracene	1.416	1.310	7.5	86	0.00
41	Benzo(g,h,i)perylene	1.458	1.313	9.9	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4