

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112320\
 Data File : BN012692.D
 Acq On : 23 Nov 2020 11:48
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 23 13:36:33 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN110420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 23 12:44:50 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	85	0.00
2	1,4-Dioxane	0.400	0.379	5.3	86	0.00
3	n-Nitrosodimethylamine	0.400	0.350	12.5	83	0.00
4 S	2-Fluorophenol	0.400	0.356	11.0	81	0.00
5 S	Phenol-d6	0.400	0.343	14.2	77	0.00
6	bis(2-Chloroethyl)ether	0.400	0.349	12.8	85	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	85	0.00
8 S	Nitrobenzene-d5	0.400	0.365	8.8	87	0.00
9	Naphthalene	0.400	0.378	5.5	90	0.00
10	Hexachlorobutadiene	0.400	0.372	7.0	87	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.379	5.3	92	0.00
12	2-Methylnaphthalene	0.400	0.378	5.5	91	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	87	0.00
14 S	2,4,6-Tribromophenol	0.400	0.378	5.5	96	0.00
15 S	2-Fluorobiphenyl	0.400	0.355	11.3	86	0.00
16	Acenaphthylene	0.400	0.365	8.8	92	0.00
17	Acenaphthene	0.400	0.363	9.3	91	0.00
18	Fluorene	0.400	0.354	11.5	89	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	85	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.353	11.8	91	0.00
21	4-Bromophenyl-phenylether	0.400	0.351	12.3	85	0.00
22	Hexachlorobenzene	0.400	0.385	3.8	90	0.00
23	Atrazine	0.400	0.363	9.3	88	0.00
24	Pentachlorophenol	0.400	0.378	5.5	96	0.00
25	Phenanthrene	0.400	0.381	4.8	91	0.00
26	Anthracene	0.400	0.374	6.5	91	0.00
27 SURR	Fluoranthene-d10	0.400	0.380	5.0	92	0.00
28	Fluoranthene	0.400	0.376	6.0	91	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	88	0.00
30	Pyrene	0.400	0.373	6.8	92	0.00
31 S	Terphenyl-d14	0.400	0.370	7.5	90	0.00
32	Benzo(a)anthracene	0.400	0.382	4.5	93	0.00
33	Chrysene	0.400	0.361	9.8	89	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.397	0.8	101	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.377	5.8	86	0.00
36 I	Perylene-d12	0.400	0.400	0.0	88	0.00
37	Benzo(b)fluoranthene	0.400	0.364	9.0	92	0.00
38	Benzo(k)fluoranthene	0.400	0.363	9.3	89	0.00
39 C	Benzo(a)pyrene	0.400	0.360	10.0	89	0.00
40	Dibenzo(a,h)anthracene	0.400	0.370	7.5	86	0.00
41	Benzo(g,h,i)perylene	0.400	0.360	10.0	84	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4