

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112320\
 Data File : BN012707.D
 Acq On : 23 Nov 2020 21:19
 Operator : CG/JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 24 01:54:58 2020
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\8270-SIM-BN110420.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 23 12:44:50 2020
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	78	0.00
2	1,4-Dioxane	0.400	0.380	5.0	79	0.00
3	n-Nitrosodimethylamine	0.400	0.377	5.8	82	0.00
4 S	2-Fluorophenol	0.400	0.383	4.3	80	0.00
5 S	Phenol-d6	0.400	0.377	5.8	77	0.00
6	bis(2-Chloroethyl)ether	0.400	0.370	7.5	83	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	83	0.00
8 S	Nitrobenzene-d5	0.400	0.364	9.0	85	0.00
9	Naphthalene	0.400	0.383	4.3	89	0.00
10	Hexachlorobutadiene	0.400	0.385	3.8	88	0.00
11 SURR	2-Methylnaphthalene-d10	0.400	0.377	5.8	89	0.00
12	2-Methylnaphthalene	0.400	0.383	4.3	89	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	86	0.00
14 S	2,4,6-Tribromophenol	0.400	0.353	11.8	88	0.00
15 S	2-Fluorobiphenyl	0.400	0.361	9.8	86	0.00
16	Acenaphthylene	0.400	0.378	5.5	94	0.00
17	Acenaphthene	0.400	0.363	9.3	90	0.00
18	Fluorene	0.400	0.362	9.5	89	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	89	0.00
20	4,6-Dinitro-2-methylphenol	0.400	0.266	33.5#	72	0.00
21	4-Bromophenyl-phenylether	0.400	0.343	14.2	87	0.00
22	Hexachlorobenzene	0.400	0.366	8.5	90	0.00
23	Atrazine	0.400	0.354	11.5	90	0.00
24	Pentachlorophenol	0.400	0.366	8.5	98	0.00
25	Phenanthrene	0.400	0.368	8.0	92	0.00
26	Anthracene	0.400	0.361	9.8	92	0.00
27 SURR	Fluoranthene-d10	0.400	0.361	9.8	91	0.00
28	Fluoranthene	0.400	0.355	11.3	90	0.00
29 I	Chrysene-d12	0.400	0.400	0.0	88	0.00
30	Pyrene	0.400	0.370	7.5	91	0.00
31 S	Terphenyl-d14	0.400	0.365	8.8	89	0.00
32	Benzo(a)anthracene	0.400	0.363	9.3	89	0.00
33	Chrysene	0.400	0.362	9.5	89	0.00
34	Bis(2-ethylhexyl)phthalate	0.400	0.412	-3.0	105	0.00
35	Indeno(1,2,3-cd)pyrene	0.400	0.380	5.0	87	0.00
36 I	Perylene-d12	0.400	0.400	0.0	87	0.00
37	Benzo(b)fluoranthene	0.400	0.362	9.5	90	0.00
38	Benzo(k)fluoranthene	0.400	0.370	7.5	90	0.00
39 C	Benzo(a)pyrene	0.400	0.368	8.0	91	0.00
40	Dibenzo(a,h)anthracene	0.400	0.372	7.0	87	0.00
41	Benzo(g,h,i)perylene	0.400	0.371	7.3	86	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 0

Instrument :

BNA_N

LabSampleId :

SSTDCCC0.4