

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN112519\
 Data File : BN008741.D
 Acq On : 25 Nov 2019 17:40
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 26 06:07:31 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN112219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 25 10:49:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	146	0.00
2	1,4-Dioxane	0.423	0.469	-10.9	162#	0.00
3	n-Nitrosodimethylamine	0.524	0.594	-13.4	177#	0.00
4 S	2-Fluorophenol	1.048	1.200	-14.5	166#	0.00
5 S	Phenol-d6	1.324	1.502	-13.4	164#	0.00
6	bis(2-Chloroethyl)ether	1.241	1.427	-15.0	164#	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	147	0.00
8 S	Nitrobenzene-d5	0.359	0.410	-14.2	169#	-0.01
9	Naphthalene	1.068	1.218	-14.0	164#	-0.01
10	Hexachlorobutadiene	0.221	0.249	-12.7	162#	-0.01
11 SURR	2-Methylnaphthalene-d10	0.631	0.717	-13.6	167#	0.00
12	2-Methylnaphthalene	0.743	0.840	-13.1	165#	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	144	0.00
14 S	2,4,6-Tribromophenol	0.186	0.191	-2.7	155#	0.00
15 S	2-Fluorobiphenyl	1.563	1.813	-16.0	165#	0.00
16	Acenaphthylene	1.841	2.095	-13.8	167#	0.00
17	Acenaphthene	1.212	1.343	-10.8	164#	0.00
18	Fluorene	1.552	1.766	-13.8	165#	0.00
19 I	Phenanthrene-d10	1.000	1.000	0.0	142	0.00
20	4-Bromophenyl-phenylether	0.250	0.285	-14.0	165#	0.00
21	Hexachlorobenzene	0.275	0.311	-13.1	162#	0.00
22	Pentachlorophenol	0.098	0.089	9.2	151#	0.00
23	Phenanthrene	1.233	1.405	-13.9	164#	0.00
24	Anthracene	1.103	1.242	-12.6	165#	0.00
25 SURR	Fluoranthene-d10	1.176	1.307	-11.1	159#	0.00
26	Fluoranthene	1.448	1.605	-10.8	160#	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	140	0.00
28	Pvrene	1.477	1.696	-14.8	159#	0.00
29 S	Terphenyl-d14	0.962	1.121	-16.5	161#	0.00
30	Benzo(a)anthracene	1.464	1.620	-10.7	156#	0.00
31	Chrysene	1.418	1.635	-15.3	159#	0.00
32	Bis(2-ethylhexyl)phthalate	0.951	0.782	17.8	125	0.00
33	Indeno(1,2,3-cd)pyrene	1.580	1.787	-13.1	157#	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	136	-0.02
35	Benzo(b)fluoranthene	1.384	1.623	-17.3	160#	-0.01
36	Benzo(k)fluoranthene	1.402	1.610	-14.8	154#	-0.01
37 C	Benzo(a)pyrene	1.281	1.452	-13.3	155#	-0.01
38	Dibenzo(a,h)anthracene	1.264	1.464	-15.8	157#	-0.01
39	Benzo(a,h,i)perylene	1.260	1.485	-17.9	160#	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN112519\
Data File : BN008741.D
Acq On : 25 Nov 2019 17:40
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Nov 26 06:07:31 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN112219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 25 10:49:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			