

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN112519\
 Data File : BN008741.D
 Acq On : 25 Nov 2019 17:40
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 26 06:07:31 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN112219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 25 10:49:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	Amount	Calc.	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	0.400	0.400	0.0	146	0.00
2	1,4-Dioxane	0.400	0.443	-10.7	162	0.00
3	n-Nitrosodimethylamine	0.400	0.453	-13.2	177	0.00
4 S	2-Fluorophenol	0.400	0.458	-14.5	166	0.00
5 S	Phenol-d6	0.400	0.454	-13.5	164	0.00
6	bis(2-Chloroethyl)ether	0.400	0.460	-15.0	164	0.00
7 I	Naphthalene-d8	0.400	0.400	0.0	147	0.00
8 S	Nitrobenzene-d5	0.400	0.457	-14.2	169	-0.01
9	Naphthalene	0.400	0.456	-14.0	164	-0.01
10	Hexachlorobutadiene	0.400	0.450	-12.5	162	-0.01
11 SURR	2-Methylnaphthalene-d10	0.400	0.454	-13.5	167	0.00
12	2-Methylnaphthalene	0.400	0.452	-13.0	165	0.00
13 I	Acenaphthene-d10	0.400	0.400	0.0	144	0.00
14 S	2,4,6-Tribromophenol	0.400	0.411	-2.7	155	0.00
15 S	2-Fluorobiphenyl	0.400	0.464	-16.0	165	0.00
16	Acenaphthylene	0.400	0.455	-13.7	167	0.00
17	Acenaphthene	0.400	0.443	-10.7	164	0.00
18	Fluorene	0.400	0.455	-13.7	165	0.00
19 I	Phenanthrene-d10	0.400	0.400	0.0	142	0.00
20	4-Bromophenyl-phenylether	0.400	0.456	-14.0	165	0.00
21	Hexachlorobenzene	0.400	0.453	-13.2	162	0.00
22	Pentachlorophenol	0.400	0.361	9.8	151	0.00
23	Phenanthrene	0.400	0.456	-14.0	164	0.00
24	Anthracene	0.400	0.450	-12.5	165	0.00
25 SURR	Fluoranthene-d10	0.400	0.445	-11.2	159	0.00
26	Fluoranthene	0.400	0.443	-10.7	160	0.00
27 I	Chrysene-d12	0.400	0.400	0.0	140	0.00
28	Pvrene	0.400	0.460	-15.0	159	0.00
29 S	Terphenyl-d14	0.400	0.466	-16.5	161	0.00
30	Benzo(a)anthracene	0.400	0.443	-10.7	156	0.00
31	Chrysene	0.400	0.461	-15.3	159	0.00
32	Bis(2-ethylhexyl)phthalate	0.400	0.329	17.8	125	0.00
33	Indeno(1,2,3-cd)pyrene	0.400	0.452	-13.0	157	-0.02
34 I	Perylene-d12	0.400	0.400	0.0	136	-0.02
35	Benzo(b)fluoranthene	0.400	0.469	-17.2	160	-0.01
36	Benzo(k)fluoranthene	0.400	0.459	-14.8	154	-0.01
37 C	Benzo(a)pyrene	0.400	0.454	-13.5	155	-0.01
38	Dibenzo(a,h)anthracene	0.400	0.463	-15.8	157	-0.01
39	Benzo(a,h,i)perylene	0.400	0.471	-17.7	160	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN112519\
Data File : BN008741.D
Acq On : 25 Nov 2019 17:40
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Nov 26 06:07:31 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN112219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 25 10:49:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	Amount	Calc.	%Dev	Area%	Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0				