

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN112519\
 Data File : BN008750.D
 Acq On : 25 Nov 2019 23:03
 Operator : JU
 Sample : SSTDCCC0.4
 Misc :
 ALS Vial : 2 Sample Multiplier: 1

Instrument :
 BNA_N
 LabSampleId :
 SSTDCCC0.4

Quant Time: Nov 26 02:55:11 2019
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN112219.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 25 10:49:40 2019
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvqRF	CCRF	%Dev	Area%	Dev(min)
1 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	109	0.00
2	1,4-Dioxane	0.423	0.476	-12.5	123	0.00
3	n-Nitrosodimethylamine	0.524	0.523	0.2	117	0.00
4 S	2-Fluorophenol	1.048	1.212	-15.6	126	0.00
5 S	Phenol-d6	1.324	1.521	-14.9	124	0.00
6	bis(2-Chloroethyl)ether	1.241	1.404	-13.1	121	0.00
7 I	Naphthalene-d8	1.000	1.000	0.0	110	0.00
8 S	Nitrobenzene-d5	0.359	0.407	-13.4	125	-0.01
9	Naphthalene	1.068	1.215	-13.8	123	-0.01
10	Hexachlorobutadiene	0.221	0.248	-12.2	121	-0.01
11 SURR	2-Methylnaphthalene-d10	0.631	0.708	-12.2	123	0.00
12	2-Methylnaphthalene	0.743	0.835	-12.4	123	0.00
13 I	Acenaphthene-d10	1.000	1.000	0.0	109	0.00
14 S	2,4,6-Tribromophenol	0.186	0.204	-9.7	125	0.00
15 S	2-Fluorobiphenyl	1.563	1.803	-15.4	124	-0.01
16	Acenaphthylene	1.841	2.104	-14.3	127	-0.01
17	Acenaphthene	1.212	1.330	-9.7	122	-0.01
18	Fluorene	1.552	1.773	-14.2	125	-0.01
19 I	Phenanthrene-d10	1.000	1.000	0.0	109	-0.01
20	4-Bromophenyl-phenylether	0.250	0.286	-14.4	127	-0.01
21	Hexachlorobenzene	0.275	0.314	-14.2	125	-0.01
22	Pentachlorophenol	0.098	0.091	7.1	118	-0.01
23	Phenanthrene	1.233	1.399	-13.5	125	-0.01
24	Anthracene	1.103	1.261	-14.3	128	0.00
25 SURR	Fluoranthene-d10	1.176	1.336	-13.6	125	0.00
26	Fluoranthene	1.448	1.636	-13.0	125	0.00
27 I	Chrysene-d12	1.000	1.000	0.0	109	-0.01
28	Pvrene	1.477	1.703	-15.3	124	0.00
29 S	Terphenyl-d14	0.962	1.129	-17.4	126	0.00
30	Benzo(a)anthracene	1.464	1.611	-10.0	120	-0.01
31	Chrysene	1.418	1.641	-15.7	124	-0.01
32	Bis(2-ethylhexyl)phthalate	0.951	0.822	13.6	102	-0.01
33	Indeno(1,2,3-cd)pyrene	1.580	1.782	-12.8	121	-0.02
34 I	Perylene-d12	1.000	1.000	0.0	106	-0.02
35	Benzo(b)fluoranthene	1.384	1.623	-17.3	124	-0.01
36	Benzo(k)fluoranthene	1.402	1.621	-15.6	121	-0.01
37 C	Benzo(a)pyrene	1.281	1.467	-14.5	122	-0.01
38	Dibenzo(a,h)anthracene	1.264	1.452	-14.9	121	-0.02
39	Benzo(a,h,i)perylene	1.260	1.466	-16.3	123	-0.02

Data Path : Z:\SVOASRV\HPCHEM1\BNA N\DATA\BN112519\
Data File : BN008750.D
Acq On : 25 Nov 2019 23:03
Operator : JU
Sample : SSTDCCC0.4
Misc :
ALS Vial : 2 Sample Multiplier: 1

Instrument :
BNA_N
LabSampleId :
SSTDCCC0.4

Quant Time: Nov 26 02:55:11 2019
Quant Method : Z:\SVOASRV\HPCHEM1\BNA N\METHODS\8270-SIM-BN112219.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 25 10:49:40 2019
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

Compound	AvgRF	CCRF	%Dev Area	% Dev(min)

(#) = Out of Range	SPCC's out = 0 CCC's out = 0			