

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112622\
 Data File : BN022907.D
 Acq On : 27 Nov 2022 07:34
 Operator : CG/JU
 Sample : N5591-07
 Misc :
 ALS Vial : 29 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C9808

Quant Time: Nov 28 03:51:40 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN112622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 28 03:16:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.066	152	8747	0.400	ng/u1	0.00
4) Naphthalene-d8	10.887	136	24901	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.705	164	13205	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.454	188	26659	0.400	ng/u1	0.00
17) Chrysene-d12	21.633	240	15943	0.400	ng/u1	0.00
23) Perylene-d12	24.124	264	14018	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.412	96	43050	4.101	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.471	152	10136	0.267	ng/u1	0.00
18) Fluoranthene-d10	19.476	212	20014	0.340	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112622\
Data File : BN022907.D
Acq On : 27 Nov 2022 07:34
Operator : CG/JU
Sample : N5591-07
Misc :
ALS Vial : 29 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C9808

Quant Time: Nov 28 03:51:40 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN112622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 28 03:16:56 2022
Response via : Initial Calibration

