

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112622\
 Data File : BN022879.D
 Acq On : 26 Nov 2022 13:00
 Operator : CG/JU
 Sample : SSTD3.257
 Misc :
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 SSTD3.2214

Quant Time: Nov 28 03:01:46 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN112622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 28 02:35:52 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.070	152	7580	0.400	ng/ul	0.00
4) Naphthalene-d8	10.892	136	20947	0.400	ng/ul	0.00
9) Acenaphthene-d10	14.710	164	11391	0.400	ng/ul	0.00
13) Phenanthrene-d10	17.458	188	23096	0.400	ng/ul	0.00
17) Chrysene-d12	21.639	240	16264	0.400	ng/ul	0.00
23) Perylene-d12	24.127	264	9845	0.400	ng/ul	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.412	96	28596	3.144	ng/ul	0.00
6) 2-Methylnaphthalene-d10	0.000	152	0d	0.000	ng/ul	
18) Fluoranthene-d10	0.000	212	0d	0.000	ng/ul	
Target Compounds						
					Qvalue	
2) 1,4-Dioxane	3.446	88	29909	3.153	ng/ul	90
14) Pentachlorophenol	17.099	266	18194	3.767	ng/ul#	100

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112622\
Data File : BN022879.D
Acq On : 26 Nov 2022 13:00
Operator : CG/JU
Sample : SSTD3.257
Misc :
ALS Vial : 7 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
SSTD3.2214

Quant Time: Nov 28 03:01:46 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN112622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 28 02:35:52 2022
Response via : Initial Calibration

