

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112622\
 Data File : BN022899.D
 Acq On : 27 Nov 2022 02:03
 Operator : CG/JU
 Sample : N5591-06
 Misc :
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 C9807

Quant Time: Nov 28 03:50:35 2022
 Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN112622.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Mon Nov 28 03:16:56 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) 1,4-Dichlorobenzene-d4	8.070	152	8141	0.400	ng/u1	0.00
4) Naphthalene-d8	10.887	136	23454	0.400	ng/u1	0.00
9) Acenaphthene-d10	14.710	164	12818	0.400	ng/u1	0.00
13) Phenanthrene-d10	17.453	188	25498	0.400	ng/u1	0.00
17) Chrysene-d12	21.633	240	14062	0.400	ng/u1	0.00
23) Perylene-d12	24.121	264	12374	0.400	ng/u1	0.00
System Monitoring Compounds						
3) 1,4-Dioxane-d8	3.412	96	46221	4.731	ng/u1	0.00
6) 2-Methylnaphthalene-d10	12.470	152	11567	0.324	ng/u1	0.00
18) Fluoranthene-d10	19.476	212	21589	0.416	ng/u1	0.00

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112622\
Data File : BN022899.D
Acq On : 27 Nov 2022 02:03
Operator : CG/JU
Sample : N5591-06
Misc :
ALS Vial : 20 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
C9807

Quant Time: Nov 28 03:50:35 2022
Quant Method : Z:\SVOASRV\HPCHEM1\BNA_N\METHODS\SFAM-EPA-SIM-BN112622.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Mon Nov 28 03:16:56 2022
Response via : Initial Calibration

