

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN112624\
 Data File : BN035308.D
 Acq On : 26 Nov 2024 09:59
 Operator : RC/JU
 Sample : P4875-22
 Misc :
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 BNA_N
 ClientSampleId :
 E2A24

Quant Time: Nov 26 22:12:35 2024
 Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN111624.M
 Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
 QLast Update : Tue Nov 26 06:09:15 2024
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


Internal Standards						
1) 1,4-Dichlorobenzene-d4	7.309	152	2861	0.400	ng/ul	0.00
4) Naphthalene-d8	10.058	136	6920	0.400	ng/ul	0.00
9) Acenaphthene-d10	13.972	164	3919	0.400	ng/ul	0.00
13) Phenanthrene-d10	0.000	188	0	0.000	ng/ul	-16.74
17) Chrysene-d12	20.982	240	6097	0.400	ng/ul	# 0.00
23) Perylene-d12	22.951	264	674243	0.400	ng/ul	#-0.13
System Monitoring Compounds						
3) 1,4-Dioxane-d8	2.972	96	14095	5.390	ng/ul	0.00
6) 2-Methylnaphthalene-d10	11.664	152	2982	0.294	ng/ul	0.00
18) Fluoranthene-d10	18.791	212	7563	0.450	ng/ul	0.00
Target Compounds						
					Qvalue	
5) Naphthalene	10.108	128	357	0.020	ng/ul#	61
12) Fluorene	14.980	166	1119	0.080	ng/ul#	17
20) Pyrene	19.158	202	1164	0.053	ng/ul#	1


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\svoasrv\HPCHEM1\BNA_N\Data\BN11624\
Data File : BN035308.D
Acq On : 26 Nov 2024 09:59
Operator : RC/JU
Sample : P4875-22
Misc :
ALS Vial : 8 Sample Multiplier: 1

Instrument :
BNA_N
ClientSampleId :
E2A24

Quant Time: Nov 26 22:12:35 2024
Quant Method : Z:\svoasrv\HPCHEM1\BNA_N\Methods\SFAM-EPA-SIM-BN11624.M
Quant Title : ASP BNA STANDARDS FOR 5 POINT CALIBRATION
QLast Update : Tue Nov 26 06:09:15 2024
Response via : Initial Calibration









#5
Naphthalene
Concen: 0.020 ng/ul
RT: 10.108 min Scan# 11
Delta R.T. -0.000 min
Lab File: BN035308.D
Acq: 26 Nov 2024 09:59




Instrument :
BNA_N
ClientSampleId :
E2A24


Tgt Ion:128 Resp: 357
Ion Ratio Lower Upper
128 100
129 29.7 9.9 14.9#
127 28.7 11.4 17.2#



#12
Fluorene
Concen: 0.080 ng/ul
RT: 14.980 min Scan# 2650
Delta R.T. -0.056 min
Lab File: BN035308.D
Acq: 26 Nov 2024 09:59


Tgt Ion:166 Resp: 1119
Ion Ratio Lower Upper
166 100
165 11.2 78.6 118.0#
167 34.7 11.8 17.6#





#20
 Pyrene
 Concen: 0.053 ng/ul
 RT: 19.158 min Scan# 30
 Delta R.T. -0.028 min
 Lab File: BN035308.D
 Acq: 26 Nov 2024 09:59

Instrument :
 BNA_N
 ClientSampleId :
 E2A24



Tgt Ion:	202	Resp:	1164
Ion Ratio	Lower	Upper	
202	100		
101	145.8	6.2	9.4#
100	1020.5	5.1	7.7#

